



# NuMED

## Summary of Safety and Clinical Performance


### SSCP – Stents – CoA & RVOT

*This Summary of Safety and Clinical Performance (SSCP) is intended to provide public access to an updated summary of the main aspects of the safety and clinical performance of the device.*

*The SSCP is not intended to replace the Instructions for Use as the main document to ensure the safe use of the device, nor is it intended to provide diagnostic or therapeutic suggestions to intended users or patients.*

*The following information is intended for users/healthcare professionals.*


<b>1. Device identification and general information</b>	
Device trade name(s)	<u>NuMED Right Ventricular Outflow Tract (RVOT) / Coarctation of the Aorta (CoA) Stent Family</u> Covered CP Stent Covered Mounted CP Stent
Model Number	<u>RVOT/CoA Stent Family – Model 1650</u> Covered CP Stent – Model 427.1 <u>Mounted RVOT/CoA Stent Family – Model 1660</u> Covered Mounted CP Stent – Model 428.1
Manufacturer's name and address	NuMED, Inc. 2880 Main Street Hopkinton, NY 12965 USA
Manufacturer's single registration number (SRN)	US-MF-000010948
Basic UDI-DI	Covered CP Stent – 08877141650TH Covered Mounted CP Stent – 08877141660TL
Medical device nomenclature description / text	EMDN – P070401020199 – PTFE VASCULAR ENDOPROSTHESES, STRAIGHT - OTHER
Class of device	III
Year when first certificate (CE) was issued	2004 (Covered CP Stent) 2009 (Covered Mounted CP Stent)
Authorised Representative (AR)	G. van Wageningen B.V. Hallenweg 40, 5683 CT Best, The Netherlands
AR SRN	NL-AR-000010437
Notified Body	SGS Belgium NV
Notified Body ID number	1639
<b>2. Intended use of the device</b>	
Indications for use	<b><u>INTENDED PURPOSE</u></b> Intended to dilate aortic coarctations, and/or contain conduit disruptions in the right ventricular outflow tract (RVOT), using percutaneous implantation techniques.



**NuMED**  
**Summary of Safety and Clinical Performance**  
**SSCP – Stents – CoA & RVOT**


	<p><b><u>INDICATIONS</u></b></p> <p><b><u>Coarctation of the Aorta (CoA)</u></b></p> <p>Indicated for treatment of native and/or recurrent coarctation of the aorta on patients with the following clinical conditions:</p> <ul style="list-style-type: none"> <li>• Stenosis of the aorta resulting in significant anatomic narrowing as determined by angiography or non-invasive imaging, i.e. echocardiography, magnetic resonance imaging (MRI), CT scan;</li> <li>• Stenosis of the aorta resulting in hemodynamic alterations, resulting in systolic pressure gradient, systemic hypertension or altered left ventricular function;</li> <li>• Stenosis of the aorta where balloon angioplasty is ineffective or contraindicated;</li> <li>• Stenosis diameter &lt;20% of adjacent vessel diameter. Stenosis that would present increased risk of vascular damage or disruption; or aneurysm associated with coarctation of the aorta.</li> </ul> <p><b><u>Right Ventricular Outflow Tract (RVOT)</u></b></p> <p>Indicated for treatment of right ventricle to pulmonary artery (right ventricular outflow tract) conduit disruptions that are identified during conduit pre-dilatation procedures performed in preparation for transcatheter pulmonary valve replacement.</p>
<p>Contraindications and/or limitations</p>	<p>Contraindications include:</p> <ul style="list-style-type: none"> <li>• Patients too small to allow safe delivery of the stent without compromise to the systemic artery used for delivery;</li> <li>• Unfavorable aortic anatomy that does not dilate with high pressure balloon angioplasty (CoA only);</li> <li>• Occlusion or obstruction of systemic artery precluding delivery of the stent (CoA only);</li> <li>• Clinical or biological signs of infection;</li> <li>• Active endocarditis;</li> <li>• Known allergy to aspirin, other antiplatelet agents, or heparin (CoA only);</li> <li>• Pregnancy.</li> </ul>

<p><b>3. Device description</b></p>	
<p>Description of the device</p>	<p>The Stents are balloon expandable and intended for permanent implant. The Stents are composed of heat treated 90% platinum / 10% iridium wire that is arranged in a "zig" pattern, laser welded at each joint, and over brazed with 24K gold. The number of zigs in a row can be varied and will impact the strength of the stent as well as the eventual expanded diameter and percent stent shortening, while the number of rows will determine the unexpanded length of the stent. The Stents have an ePTFE covering attached to the stent framework. This covering acts as a fluid barrier creating a fluid tight conduit through the stent length.</p> <p>The Balloon In Balloon (BIB) Stent Placement Catheter is triaxial in construction with two lumens being used to inflate the balloon while one lumen is being used for tracking over a guidewire. The inner balloon is ½ of the outer balloon diameter and 1 cm shorter. The purpose of the double balloon catheter is to apply an incremental inflation for the purpose of dilating a stent. The inner balloon provides initial expansion of the stent and also acts as a tool to hold the stent on the catheter prior to the outer balloon being inflated. The outer balloon is then inflated, providing the remainder of the expansion. There are radiopaque platinum marker bands under the balloon shoulders, to aid during placement. The balloons are designed to inflate to the diameter and length listed on the label at a specific pressure. Thus, it is recommended that the device be used in conjunction with a mechanism to monitor pressure, an inflation device with pressure gauge.</p> <p>The Stents are supplied sterile, by ethylene oxide gas, and are intended for single use only.</p>



## NuMED Summary of Safety and Clinical Performance SSCP – Stents – CoA & RVOT

Reference to previous generation(s) or variants	<b>REF</b>	<b>Description</b>	<b>REF</b>	<b>Description</b>
	CVRDCP8Z16	Covered 8 Zig 1.6 cm	CMCP001	CVRD 8 ZIG 1.6 CM 6X1.5////12X2.5X8X8X110X0.035
	CVRDCP8Z22	Covered 8 Zig 2.2 cm	CMCP002	CVRD 8 ZIG 1.6 CM 7X1.5////14X2.5X8X8X110X0.035
	CVRDCP8Z28	Covered 8 Zig 2.8 cm	CMCP003	CVRD 8 ZIG 1.6 CM 8X1.5////16X2.5X9X9X110X0.035
	CVRDCP8Z34	Covered 8 Zig 3.4 cm	CMCP004	CVRD 8 ZIG 2.2 CM 6X1.5////12X2.5X8X8X110X0.035
	CVRDCP8Z39	Covered 8 Zig 3.9 cm	CMCP005	CVRD 8 ZIG 2.2 CM 7X1.5////14X2.5X8X8X110X0.035
	CVRDCP8Z45	Covered 8 Zig 4.5 cm	CMCP006	CVRD 8 ZIG 2.2 CM 8X1.5////16X2.5X9X9X110X0.035
	CVRDCP8Z50	Covered 8 Zig 5.0 cm	CMCP007	CVRD 8 ZIG 2.2 CM 9X1.5////18X2.5X10X9X110X0.035
	CVRDCP8Z55	Covered 8 Zig 5.5 cm	CMCP008	CVRD 8 ZIG 2.8 CM 7X2.5////14X3X8X8X110X0.035
	CVRDCP8Z60	Covered 8 Zig 6.0 cm	CMCP009	CVRD 8 ZIG 2.8 CM 8X2.5////16X3X9X9X110X0.035
	CVRDCP10Z39	Covered 10 Zig 3.9 cm	CMCP010	CVRD 8 ZIG 2.8 CM 9X2.5////18X3X10X9X110X0.035
	CVRDCP10Z45	Covered 10 Zig 4.5 cm	CMCP011	CVRD 8 ZIG 2.8 CM 10X2.5////20X3X10X9X110X0.035
	CVRDCP10Z50	Covered 10 Zig 5.0 cm	CMCP012	CVRD 8 ZIG 3.4 CM 7X2.5////14X3.5X8X8X110X0.035
	CVRDCP10Z55	Covered 10 Zig 5.5 cm	CMCP013	CVRD 8 ZIG 3.4 CM 8X2.5////16X3.5X9X9X110X0.035
	CVRDCP10Z60	Covered 10 Zig 6.0 cm	CMCP014	CVRD 8 ZIG 3.4 CM 9X2.5////18X3.5X10X9X110X0.035
			CMCP015	CVRD 8 ZIG 3.4 CM 10X2.5////20X3.5X10X9X110X0.035
			CMCP016	CVRD 8 ZIG 3.4 CM 11X2.5////22X3.5X11X9X110X0.035
			CMCP017	CVRD 8 ZIG 3.9 CM 7X3.5////14X4X8X8X110X0.035
			CMCP018	CVRD 8 ZIG 3.9 CM 8X3.5////16X4X9X9X110X0.035
			CMCP019	CVRD 8 ZIG 3.9 CM 9X3.5////18X4X10X9X110X0.035
			CMCP020	CVRD 8 ZIG 3.9 CM 10X3.5////20X4X10X9X110X0.035
			CMCP021	CVRD 8 ZIG 3.9 CM 11X3.5////22X4X11X9X110X0.035
			CMCP022	CVRD 8 ZIG 3.9 CM 12X3.5////24X4X11X9X110X0.035
			CMCP023	CVRD 8 ZIG 4.5 CM 7X3.5////14X4.5X8X8X110X0.035
			CMCP024	CVRD 8 ZIG 4.5 CM 8X3.5////16X4.5X9X9X110X0.035
			CMCP025	CVRD 8 ZIG 4.5 CM 9X3.5////18X4.5X10X9X110X0.035
			CMCP026	CVRD 8 ZIG 4.5 CM 10X3.5////20X4.5X10X9X110X0.035
			CMCP027	CVRD 8 ZIG 4.5 CM 11X3.5////22X4.5X11X9X110X0.035
			CMCP028	CVRD 8 ZIG 4.5 CM 12X3.5////24X4.5X11X9X110X0.035
			CMCP029	CVRD 8 ZIG 4.5 CM 7X4.5////14X5X8X8X110X0.035
			CMCP030	CVRD 8 ZIG 4.5 CM 8X4.5////16X5X9X9X110X0.035
			CMCP031	CVRD 8 ZIG 4.5 CM 9X4.5////18X5X10X9X110X0.035
			CMCP032	CVRD 8 ZIG 4.5 CM 10X4.5////20X5X10X9X110X0.035
			CMCP033	CVRD 8 ZIG 4.5 CM 11X4.5////22X5X11X9X110X0.035
			CMCP034	CVRD 8 ZIG 4.5 CM 12X4.5////24X5X11X9X110X0.035
			CMCP035	CVRD 8 ZIG 3.4 CM 6X2.5////12X3.5X8X8X110X0.035
			CMCP036	CVRD 8 ZIG 3.9 CM 6X3.5////12X4X8X8X110X0.035
			CMCP037	CVRD 8 ZIG 4.5 CM 6X4.5////12X5X8X8X110X0.035
			CMCP038	CVRD 8 ZIG 2.8 CM 6X2.5////12X3X8X8X110X0.035
			CMCP040	CVRD 10 ZIG 3.9 CM 13X3.5////26X4X16X11X110X0.035



**NuMED**  
**Summary of Safety and Clinical Performance**  
**SSCP – Stents – CoA & RVOT**

Reference to previous generation(s) or variants		<table border="1"> <thead> <tr> <th>REF</th> <th>Description</th> </tr> </thead> <tbody> <tr><td>CMCP041</td><td>CVRD 10 ZIG 3.9 CM 14X3////28X4X16X11X110X0.035</td></tr> <tr><td>CMCP042</td><td>CVRD 10 ZIG 3.9 CM 15X3////30X4X16X11X110X0.035</td></tr> <tr><td>CMCP044</td><td>CVRD 10 ZIG 4.5 CM 13X4////26X5X16X11X110X0.035</td></tr> <tr><td>CMCP045</td><td>CVRD 10 ZIG 4.5 CM 14X4////28X5X16X11X110X0.035</td></tr> <tr><td>CMCP046</td><td>CVRD 10 ZIG 4.5 CM 15X4////30X5X16X11X110X0.035</td></tr> <tr><td>CMCP048</td><td>CVRD 10 ZIG 5.0 CM 13X4.5////26X5.5X16X11X110X0.035</td></tr> <tr><td>CMCP049</td><td>CVRD 10 ZIG 5.0 CM 14X4.5////28X5.5X16X11X110X0.035</td></tr> <tr><td>CMCP050</td><td>CVRD 10 ZIG 5.0 CM 15X4.5////30X5.5X16X11X110X0.035</td></tr> <tr><td>CMCP052</td><td>CVRD 10 ZIG 5.5 CM 13X5////26X6X16X11X110X0.035</td></tr> <tr><td>CMCP053</td><td>CVRD 10 ZIG 5.5 CM 14X5////28X6X16X11X110X0.035</td></tr> <tr><td>CMCP054</td><td>CVRD 10 ZIG 5.5 CM 15X5////30X6X16X11X110X0.035</td></tr> <tr><td>CMCP056</td><td>CVRD 10 ZIG 6.0 CM 13X5////26X6X16X11X110X0.035</td></tr> <tr><td>CMCP057</td><td>CVRD 10 ZIG 6.0 CM 14X5////28X6X16X11X110X0.035</td></tr> <tr><td>CMCP058</td><td>CVRD 10 ZIG 6.0 CM 15X5////30X6X16X11X110X0.035</td></tr> <tr><td>CMCP059</td><td>CVRD 8 ZIG 5.0 CM 6X4.5////12X5.5X8X8X110X0.035</td></tr> <tr><td>CMCP060</td><td>CVRD 8 ZIG 5.0 CM 7X4.5////14X5.5X8X8X110X0.035</td></tr> <tr><td>CMCP061</td><td>CVRD 8 ZIG 5.0 CM 8X4.5////16X5.5X9X9X110X0.035</td></tr> <tr><td>CMCP062</td><td>CVRD 8 ZIG 5.0 CM 9X4.5////18X5.5X10X9X110X0.035</td></tr> <tr><td>CMCP063</td><td>CVRD 8 ZIG 5.0 CM 10X4.5////20X5.5X10X9X110X0.035</td></tr> <tr><td>CMCP064</td><td>CVRD 8 ZIG 5.0 CM 11X4.5////22X5.5X11X9X110X0.035</td></tr> <tr><td>CMCP065</td><td>CVRD 8 ZIG 5.0 CM 12X4.5////24X5.5X11X9X110X0.035</td></tr> <tr><td>CMCP066</td><td>CVRD 8 ZIG 5.5 CM 6X5////12X6X8X8X110X0.035</td></tr> <tr><td>CMCP067</td><td>CVRD 8 ZIG 5.5 CM 7X5////14X6X8X8X110X0.035</td></tr> <tr><td>CMCP068</td><td>CVRD 8 ZIG 5.5 CM 8X5////16X6X9X9X110X0.035</td></tr> <tr><td>CMCP069</td><td>CVRD 8 ZIG 5.5 CM 9X5////18X6X10X9X110X0.035</td></tr> <tr><td>CMCP070</td><td>CVRD 8 ZIG 5.5 CM 10X5////20X6X10X9X110X0.035</td></tr> <tr><td>CMCP071</td><td>CVRD 8 ZIG 5.5 CM 11X5////22X6X11X9X110X0.035</td></tr> <tr><td>CMCP072</td><td>CVRD 8 ZIG 5.5 CM 12X5////24X6X11X9X110X0.035</td></tr> <tr><td>CMCP073</td><td>CVRD 8 ZIG 6.0 CM 6X5////12X6X8X8X110X0.035</td></tr> <tr><td>CMCP074</td><td>CVRD 8 ZIG 6.0 CM 7X5////14X6X8X8X110X0.035</td></tr> <tr><td>CMCP075</td><td>CVRD 8 ZIG 6.0 CM 8X5////16X6X9X9X110X0.035</td></tr> <tr><td>CMCP076</td><td>CVRD 8 ZIG 6.0 CM 9X5////18X6X10X9X110X0.035</td></tr> <tr><td>CMCP077</td><td>CVRD 8 ZIG 6.0 CM 10X5////20X6X10X9X110X0.035</td></tr> <tr><td>CMCP078</td><td>CVRD 8 ZIG 6.0 CM 11X5////22X6X11X9X110X0.035</td></tr> <tr><td>CMCP079</td><td>CVRD 8 ZIG 6.0 CM 12X5////24X6X11X9X110X0.035</td></tr> </tbody> </table>	REF	Description	CMCP041	CVRD 10 ZIG 3.9 CM 14X3////28X4X16X11X110X0.035	CMCP042	CVRD 10 ZIG 3.9 CM 15X3////30X4X16X11X110X0.035	CMCP044	CVRD 10 ZIG 4.5 CM 13X4////26X5X16X11X110X0.035	CMCP045	CVRD 10 ZIG 4.5 CM 14X4////28X5X16X11X110X0.035	CMCP046	CVRD 10 ZIG 4.5 CM 15X4////30X5X16X11X110X0.035	CMCP048	CVRD 10 ZIG 5.0 CM 13X4.5////26X5.5X16X11X110X0.035	CMCP049	CVRD 10 ZIG 5.0 CM 14X4.5////28X5.5X16X11X110X0.035	CMCP050	CVRD 10 ZIG 5.0 CM 15X4.5////30X5.5X16X11X110X0.035	CMCP052	CVRD 10 ZIG 5.5 CM 13X5////26X6X16X11X110X0.035	CMCP053	CVRD 10 ZIG 5.5 CM 14X5////28X6X16X11X110X0.035	CMCP054	CVRD 10 ZIG 5.5 CM 15X5////30X6X16X11X110X0.035	CMCP056	CVRD 10 ZIG 6.0 CM 13X5////26X6X16X11X110X0.035	CMCP057	CVRD 10 ZIG 6.0 CM 14X5////28X6X16X11X110X0.035	CMCP058	CVRD 10 ZIG 6.0 CM 15X5////30X6X16X11X110X0.035	CMCP059	CVRD 8 ZIG 5.0 CM 6X4.5////12X5.5X8X8X110X0.035	CMCP060	CVRD 8 ZIG 5.0 CM 7X4.5////14X5.5X8X8X110X0.035	CMCP061	CVRD 8 ZIG 5.0 CM 8X4.5////16X5.5X9X9X110X0.035	CMCP062	CVRD 8 ZIG 5.0 CM 9X4.5////18X5.5X10X9X110X0.035	CMCP063	CVRD 8 ZIG 5.0 CM 10X4.5////20X5.5X10X9X110X0.035	CMCP064	CVRD 8 ZIG 5.0 CM 11X4.5////22X5.5X11X9X110X0.035	CMCP065	CVRD 8 ZIG 5.0 CM 12X4.5////24X5.5X11X9X110X0.035	CMCP066	CVRD 8 ZIG 5.5 CM 6X5////12X6X8X8X110X0.035	CMCP067	CVRD 8 ZIG 5.5 CM 7X5////14X6X8X8X110X0.035	CMCP068	CVRD 8 ZIG 5.5 CM 8X5////16X6X9X9X110X0.035	CMCP069	CVRD 8 ZIG 5.5 CM 9X5////18X6X10X9X110X0.035	CMCP070	CVRD 8 ZIG 5.5 CM 10X5////20X6X10X9X110X0.035	CMCP071	CVRD 8 ZIG 5.5 CM 11X5////22X6X11X9X110X0.035	CMCP072	CVRD 8 ZIG 5.5 CM 12X5////24X6X11X9X110X0.035	CMCP073	CVRD 8 ZIG 6.0 CM 6X5////12X6X8X8X110X0.035	CMCP074	CVRD 8 ZIG 6.0 CM 7X5////14X6X8X8X110X0.035	CMCP075	CVRD 8 ZIG 6.0 CM 8X5////16X6X9X9X110X0.035	CMCP076	CVRD 8 ZIG 6.0 CM 9X5////18X6X10X9X110X0.035	CMCP077	CVRD 8 ZIG 6.0 CM 10X5////20X6X10X9X110X0.035	CMCP078	CVRD 8 ZIG 6.0 CM 11X5////22X6X11X9X110X0.035	CMCP079	CVRD 8 ZIG 6.0 CM 12X5////24X6X11X9X110X0.035
	REF	Description																																																																								
	CMCP041	CVRD 10 ZIG 3.9 CM 14X3////28X4X16X11X110X0.035																																																																								
	CMCP042	CVRD 10 ZIG 3.9 CM 15X3////30X4X16X11X110X0.035																																																																								
	CMCP044	CVRD 10 ZIG 4.5 CM 13X4////26X5X16X11X110X0.035																																																																								
	CMCP045	CVRD 10 ZIG 4.5 CM 14X4////28X5X16X11X110X0.035																																																																								
	CMCP046	CVRD 10 ZIG 4.5 CM 15X4////30X5X16X11X110X0.035																																																																								
	CMCP048	CVRD 10 ZIG 5.0 CM 13X4.5////26X5.5X16X11X110X0.035																																																																								
	CMCP049	CVRD 10 ZIG 5.0 CM 14X4.5////28X5.5X16X11X110X0.035																																																																								
	CMCP050	CVRD 10 ZIG 5.0 CM 15X4.5////30X5.5X16X11X110X0.035																																																																								
	CMCP052	CVRD 10 ZIG 5.5 CM 13X5////26X6X16X11X110X0.035																																																																								
	CMCP053	CVRD 10 ZIG 5.5 CM 14X5////28X6X16X11X110X0.035																																																																								
	CMCP054	CVRD 10 ZIG 5.5 CM 15X5////30X6X16X11X110X0.035																																																																								
	CMCP056	CVRD 10 ZIG 6.0 CM 13X5////26X6X16X11X110X0.035																																																																								
	CMCP057	CVRD 10 ZIG 6.0 CM 14X5////28X6X16X11X110X0.035																																																																								
	CMCP058	CVRD 10 ZIG 6.0 CM 15X5////30X6X16X11X110X0.035																																																																								
	CMCP059	CVRD 8 ZIG 5.0 CM 6X4.5////12X5.5X8X8X110X0.035																																																																								
	CMCP060	CVRD 8 ZIG 5.0 CM 7X4.5////14X5.5X8X8X110X0.035																																																																								
	CMCP061	CVRD 8 ZIG 5.0 CM 8X4.5////16X5.5X9X9X110X0.035																																																																								
	CMCP062	CVRD 8 ZIG 5.0 CM 9X4.5////18X5.5X10X9X110X0.035																																																																								
	CMCP063	CVRD 8 ZIG 5.0 CM 10X4.5////20X5.5X10X9X110X0.035																																																																								
	CMCP064	CVRD 8 ZIG 5.0 CM 11X4.5////22X5.5X11X9X110X0.035																																																																								
	CMCP065	CVRD 8 ZIG 5.0 CM 12X4.5////24X5.5X11X9X110X0.035																																																																								
	CMCP066	CVRD 8 ZIG 5.5 CM 6X5////12X6X8X8X110X0.035																																																																								
	CMCP067	CVRD 8 ZIG 5.5 CM 7X5////14X6X8X8X110X0.035																																																																								
	CMCP068	CVRD 8 ZIG 5.5 CM 8X5////16X6X9X9X110X0.035																																																																								
	CMCP069	CVRD 8 ZIG 5.5 CM 9X5////18X6X10X9X110X0.035																																																																								
	CMCP070	CVRD 8 ZIG 5.5 CM 10X5////20X6X10X9X110X0.035																																																																								
	CMCP071	CVRD 8 ZIG 5.5 CM 11X5////22X6X11X9X110X0.035																																																																								
	CMCP072	CVRD 8 ZIG 5.5 CM 12X5////24X6X11X9X110X0.035																																																																								
CMCP073	CVRD 8 ZIG 6.0 CM 6X5////12X6X8X8X110X0.035																																																																									
CMCP074	CVRD 8 ZIG 6.0 CM 7X5////14X6X8X8X110X0.035																																																																									
CMCP075	CVRD 8 ZIG 6.0 CM 8X5////16X6X9X9X110X0.035																																																																									
CMCP076	CVRD 8 ZIG 6.0 CM 9X5////18X6X10X9X110X0.035																																																																									
CMCP077	CVRD 8 ZIG 6.0 CM 10X5////20X6X10X9X110X0.035																																																																									
CMCP078	CVRD 8 ZIG 6.0 CM 11X5////22X6X11X9X110X0.035																																																																									
CMCP079	CVRD 8 ZIG 6.0 CM 12X5////24X6X11X9X110X0.035																																																																									



**NuMED**  
**Summary of Safety and Clinical Performance**  
**SSCP – Stents – CoA & RVOT**

Accessories which are intended to be used in combination with the device	All Stents are designed to be used with the hemostasis valve tools that are provided with the stents.
Description of any other devices and products which are intended to be used in combination with the device	All Stents are designed to be used with delivery catheter, guidewire, introducer, balloon inflation medium, inflation device with pressure gauge, and a stopcock.

**4. Risks and Warning**

Residual risks and undesirable effects	<p>All significant risks were considered, mitigated as far as possible (AFAP), and are acceptable in regard to the clinical benefit of the device.</p> <p>The clinical investigations performed on the CP Stent family of devices reported the following side effects:          Coarctation of the Aorta Stent Trial (COAST): aortic aneurysm, stent fracture          Coarctation of the Aorta Stent Trial II (COAST II): iliac artery dissection          Pulmonary Artery Repair with Covered Stent Trial (PARCS): stent malposition, stent embolization</p> <p>The literature reported the following side effects: acute wall rupture / dissection, aortic aneurysm / pseudoaneurysm, balloon rupture, death, femoral artery dissection / pseudoaneurysm, bleeding, endoleak, homograft rupture, femoral / groin hematoma, thrombosis, late lumen loss, transitory arrhythmia, cerebrovascular incident, stent embolization, stent displacement, stent fracture, stent migration, and cardiogenic / septic shock.</p> <p>Known and foreseeable clinical risks have been considered in accordance with risk management (RM) procedure AP-346 and through the RM files and mitigated as far as possible (AFAP).</p> <p><b>POTENTIAL COMPLICATIONS/ADVERSE EFFECTS</b></p> <p>NOTE: Circumferential tear of the delivery balloon catheter prior to complete expansion of the stent may cause the balloon to become tethered to the stent, requiring surgical removal. In case of rupture of an adequately sized balloon after stent expansion, it can be withdrawn and a new balloon catheter exchanged over a guidewire to complete expansion of the stent.</p> <p>Cardiac catheterization carries certain risks. Potential complications &amp; adverse effects associated with device use and indication include:</p> <ul style="list-style-type: none"> <li>• Femoral artery injury</li> <li>• Stent Migration</li> <li>• Stent Fracture</li> <li>• Aortic Rupture/Tear</li> <li>• Hematoma</li> <li>• Thrombosis</li> <li>• Embolization</li> <li>• Death</li> <li>• Endocarditis</li> <li>• Stent Stenosis</li> <li>• Aneurysm / Pseudoaneurysm</li> <li>• Stent Malposition</li> <li>• Sepsis/Infection</li> <li>• Transitory arrhythmia</li> <li>• Bleeding</li> <li>• Cerebrovascular Incident</li> </ul>
Warning and Precautions	<p>The following Warnings and Precautions have been identified and are called out in the Instruction for Use:</p> <p><b>COVERED STENT WARNINGS</b></p> <ul style="list-style-type: none"> <li>• Radiofrequency heating during MRI scans on overlapped, 10 zig Stents has not been evaluated, and is not recommended.</li> <li>• During the RVOT clinical study, the Medtronic Melody Valve was used for valve restoration.</li> <li>• As with any type of implant, infection secondary to contamination of the stent may lead to aortitis, or abscess.</li> <li>• The platinum/iridium stent may migrate from the site of implant.</li> <li>• Over-stretching of the artery may result in rupture or aneurysm formation.</li> <li>• When the stent is crimped onto a balloon delivery catheter, the maximum balloon inflation pressure must not exceed the recommended inflation pressure specified in the manufacturer’s instructions.</li> <li>• The inflated diameter of the stent should at least equal the diameter of the intended implant site.</li> <li>• Excessive force while crimping may weaken welds of the stent.</li> </ul>



# NuMED

## Summary of Safety and Clinical Performance

### SSCP – Stents – CoA & RVOT

- Crimping the 8 zig stent on a balloon catheter smaller than 12mm, and the 10 zig on a balloon catheter smaller than 26mm, may cause damage to the stent.
- Excessive handling and manipulation of the covering while crimping the stent may cause the covering to tear off of the stent.
- Crimping the device in the opposite direction of the folds in the covering may cause the covering to catch while inserting into the hemostasis valve tool and introducer. This could cause the covering to tear off of the stent.
- Pulling the Covered Stent back through the introducer and/or hemostasis valve may cause the covering to catch and tear off of the stent.
- This device is intended for single use only. Do not resterilize and/or reuse it, as this can potentially result in compromised device performance and increased risk of cross contamination.

#### **COVERED MOUNTED STENT WARNINGS**

- Radiofrequency heating during MRI scans on overlapped, 10 zig Stents has not been evaluated, and is not recommended.
- During the RVOT clinical study, the Medtronic Melody Valve was used for valve restoration.
- As with any type of implant, infection secondary to contamination of the stent may lead to aortitis, or abscess.
- The platinum/iridium stent may migrate from the site of implant.
- Over-stretching of the artery may result in rupture or aneurysm formation.
- The inflated diameter of the stent should at least equal the diameter of the intended implant site.
- Excessive force while crimping may weaken welds of the stent.
- Crimping the 8 zig stent on a balloon catheter smaller than 12mm, and the 10 zig on a balloon catheter smaller than 26mm, may cause damage to the stent.
- Excessive handling and manipulation of the covering while crimping the stent may cause the covering to tear off of the stent.
- Crimping the device in the opposite direction of the folds in the covering may cause the covering to catch while inserting into the hemostasis valve tool and introducer. This could cause the covering to tear off of the stent.
- Pulling the Covered Stent back through the introducer and/or hemostasis valve may cause the covering to catch and tear off of the stent.

#### **BIB STENT PLACEMENT WARNINGS**

- Do not exceed the RBP. An inflation device with pressure gauge is recommended to monitor pressure. Pressure in excess of the RBP can cause balloon rupture and potential inability to withdraw the catheter through the introducer sheath.
- Confirm that the distal end of the introducing sheath is at least 2.5cm back from the most proximal image markers before inflating the outer balloon. Failure to do so may stretch the outer tubing and severely hinder balloon deflation.
- Use two appropriate size inflation devices with pressure gauges for inflation.
- Do not advance the guidewire, balloon dilatation catheter, or any other component if resistance is met, without first determining the cause and taking remedial action.
- This catheter is not recommended for pressure measurement or fluid injection.
- Do not remove the guidewire from the catheter at any time during the procedure except when the procedure has been completed.
- This device is intended for single use only. Do not resterilize and/or reuse it, as this can potentially result in compromised device performance and increased risk of cross contamination.

#### **COVERED STENT PRECAUTIONS**

- Use of an inflation device with pressure gauge is highly recommended during this procedure.
- Stents are delicate devices. Care should be exercised while handling to help prevent the possibility of breakage.
- The stent is rigid and may make negotiation through vessels difficult.
- Dilatation procedures should be conducted under fluoroscopic guidance with appropriate x-ray equipment.
- Guidewires are delicate instruments. Care should be exercised while handling to help prevent the possibility of breakage.



## NuMED Summary of Safety and Clinical Performance SSCP – Stents – CoA & RVOT

	<ul style="list-style-type: none"> <li>Careful attention must be paid to the maintenance of tight catheter connections and by aspiration before proceeding to avoid air introduction into the system.</li> <li>Under no circumstances should any portion of the catheter system be advanced against resistance. The cause of the resistance should be identified with fluoroscopy and action taken to remedy the problem.</li> </ul> <p><b>COVERED MOUNTED STENT PRECAUTIONS</b></p> <ul style="list-style-type: none"> <li>Use of an inflation device with pressure gauge is highly recommended during this procedure.</li> <li>Stents are delicate devices. Care should be exercised while handling to help prevent the possibility of breakage.</li> <li>The stent is rigid and may make negotiation through vessels difficult.</li> <li>Dilatation procedures should be conducted under fluoroscopic guidance with appropriate x-ray equipment.</li> <li>Guidewires are delicate instruments. Care should be exercised while handling to help prevent the possibility of breakage.</li> <li>Careful attention must be paid to the maintenance of tight catheter connections and by aspiration before proceeding to avoid air introduction into the system.</li> <li>The inflation diameter of the balloon used during stent delivery should approximate the diameter of the obstructive vessel and the intended implant site.</li> <li>Under no circumstances should any portion of the catheter system be advanced against resistance. The cause of the resistance should be identified with fluoroscopy and action taken to remedy the problem.</li> <li>If resistance is felt upon removal, then the balloon, guidewire and the sheath should be removed together as a unit, particularly if balloon rupture or leakage is known or suspected. This may be accomplished by firmly grasping the balloon catheter and sheath as a unit and withdrawing both together, using a gentle twisting motion combined with traction.</li> <li>The balloons must be completely deflated before retracting into the sheath.</li> <li>Proper functioning of the catheter depends on its integrity. Care should be used when handling the catheter. Damage may result from kinking, stretching, or forceful wiping of the catheter.</li> </ul>
Other relevant aspects of safety, including a summary of any field safety corrective actions (FSCA including FSN) if applicable	There have not been any Field Safety Corrective Actions or Field Safety Notices on any versions of the Stents listed in this SSCP.

### 5. Summary of clinical evaluation and post-market clinical follow-up (PMCF)

#### Summary of clinical data related to equivalent device:

An equivalent device was not used for the clinical evaluation.

#### Summary of clinical data from conducted investigations of the device:

1.	<b>Study name: COAST</b>							
<b>Appraisal</b>								
	Level of Evidence	Study Method/Design	Question Applied	Oxford LOE 2011				
		Prospective, multicenter, single arm interventional, open label study.	To evaluate the intermediate results of CP Stent (Bare and Covered) to treat native and recurrent CoA in selected children, adolescents, and adult.	1	2	3	4	5
	Suitability	Relevant Data		Grading				
	Device	<ul style="list-style-type: none"> <li>CP Stents (Bare and Covered)</li> <li>CP Stent delivered on a BIB (not pre-mounted)</li> <li>Because of the known risk of aortic wall complications during CoA intervention, NuMED covered CP Stents (CCPs) were made available to study centers for use in the event of aortic wall injury (AWI). Data on patients receiving a CCPs are included in this report for safety</li> </ul>		D1	D2	D3		



## NuMED Summary of Safety and Clinical Performance SSCP – Stents – CoA & RVOT

	outcomes. These patients were then enrolled in the COAST II trial of aortic covered stents (NCT01278303) for subsequent evaluation of efficacy and long-term outcomes.			
Application	- CoA (native and recurrent)	<b>A1</b>	A2	A3
Patient	- Patients with native or recurrent CoA - Sampling: n=105 - Mean age: 16 (range: 8 to 52) years old - Sex: 73M; 32F	<b>P1</b>	P2	P3
Report	- High quality	<b>R1</b>	R2	R3
Suitability Grade (Range 4-12)		4		

Data Contribution	Relevant Data	Grading	
Outcomes/Endpoints	- Blood pressure gradient - Coarctation minimum diameter: cardiac catheterization before and after CP Stent placement - Safety	<b>Yes 1</b>	No 2
Follow-up	- 12 - 24 months with certain assessments extending to 5 years.	<b>Yes 1</b>	No 2
Statistical analysis	- Descriptive statistics are presented as mean ± SD or median (minimum–maximum). Bivariate comparisons of preimplantation and post implantation catheterization data and subsequent blood pressures were performed with the paired t test. Comparison of means or proportions between populations were performed by unpaired t test or Wilcoxon rank-sum test based on distribution and the Fisher exact test, respectively. Multivariable analysis of dichotomous outcome variables was performed with logistic regression. Analysis of time-dependent occurrences was presented graphically with Kaplan-Meier plots and analyzed statistically by the log-rank test. Predictors of time-dependent outcomes such as reintervention were obtained from Cox proportional hazards modeling.	<b>Yes 1</b>	No 2
Clinical significance	- The CP stent is safe and associated with persistent relief of aortic obstruction. Stent fracture and progression of fracture occur but have not resulted in clinically important sequelae. Reintervention is common and related to early and late aortic wall injury and need for re-expansion of small-diameter stents.	<b>Yes 1</b>	No 2
Data Contribution Grade (Range 4-8)		4	

### Overall S&P Appraisal, Disposition and Weighting

S&P Grade (Range 9-25)	LOE (2) + Suitability (4) + Data Contribution (4) = 10	Disposition and Weighting (select)	<b>Accepted and Pivotal 9-12</b> Accepted but not Pivotal, 13-21 Excluded, 22-25
---------------------------	---	---------------------------------------	--

**Purpose:** to provide information that will support labeling of both the CP Stent (Bare and Covered) to treat native and recurrent CoA in selected children, adolescents and adult

**Clinical Study Methodology:** Single arm interventional study (open label). The COAST is a prospective, multicenter, single-arm clinical study involving 19 pediatric cardiology centers in the United States. The study includes patients with native or recurrent CoA treated by physicians at the participating institutions. A total of 105 patients underwent attempted implantation, with 104 successes.

**Reference to the clinical study plan (and amendment) n°:** NCT00552812

Investigation Site: 19 pediatric cardiology centers in United States  <b>Reference to Approved Consent Forms:</b> N/A	<b>Ethics Committee Approvals:</b> Institutional Review Board approvals from all participating institutions  <b>Reference to Document n°:</b> N/A	<b>Regulatory Authority Approvals:</b> Investigational Device Exemption from US FDA (August 3, 2007)  <b>Reference to Documents n°:</b> G060057
---	--	--

**Patient Population:** Patients with native or recurrent CoA. A total of 105 patients underwent attempted implantation, median age 16 years (range from 8 to 52 years) and with 69.5% male.

**Clinical Study Results:** Results held by Sponsor.

<b>Purpose</b>	<b>Criteria</b>	<b>Results</b>
----------------	-----------------	----------------





## NuMED Summary of Safety and Clinical Performance SSCP – Stents – CoA & RVOT

Safety	Adverse events	No serious adverse events reported, 7% of the patients experienced somewhat serious events. Aortic aneurysms (n=6): 5 were successfully treated with covered stent placement, and 1 resolved without intervention. Stent fractures were seen in 2 patients after one year, 11 patients at two years and 12 additional fractures above 2 years.
--------	----------------	--

**Reference to the Clinical Study Report n°:** NCT00552812

**Device Used:** Bare CP Stent and BIB catheter; covered stents were available in case of aortic wall injury

**Conclusion:** The CP stent is safe and associated with persistent relief of aortic obstruction. Stent fracture and progression of fracture occur but have not resulted in clinically important sequelae. Reintervention is common and related to early and late aortic wall injury and need for re-expansion of small-diameter stents.

**Clinical Publication:** Meadows J, Minahan M, McElhinney DB, McEnaney K, Ringel R. Intermediate Outcomes in the Prospective, Multicenter Coarctation of the Aorta Stent Trial (COAST). Circulation 131, 1656-1664 (2015)

### 2. Study name: COAST II

#### Appraisal

Level of Evidence	Study Method/Design	Question Applied	Oxford LOE 2011				
			1	2	3	4	5
	Multicenter, single arm interventional study.	To evaluate safety and short-term efficacy of the CP Stent (Covered) in treating or preventing aortic wall injury in patients with CoA.					

Suitability	Relevant Data	Grading		
Device	- CP Stents (Covered, pre-mounted on BIB)	<b>D1</b>	D2	D3
Application	- CoA at risk of aortic wall injury or with existing aortic wall injury	<b>A1</b>	A2	A3
Patient	- CoA Patients at risk of aortic wall injury or with existing aortic wall injury - Sampling: n=158 (83 treatment cohort; 75 prevention cohort) - Mean age: 19 (range: 5 to 70) years old - Sex: 103M;55F	<b>P1</b>	P2	P3
Report	- High quality	<b>R1</b>	R2	R3
Suitability Grade (Range 4-12)		4		

Data Contribution	Relevant Data	Grading	
Outcomes/Endpoints	- Blood pressure gradient (at 1 month) - Safety	<b>Yes 1</b>	No 2
Follow-up	- 12 - 24 months with certain assessments extending to 5 years.	<b>Yes 1</b>	No 2
Statistical analysis	- Categorical data are summarized as frequency (percent); continuous data are summarized as mean SD for normally distributed variables and median (range) otherwise. Comparisons between treatment groups were performed using Fisher exact test or the Wilcoxon rank sum test. Post-intervention measurements were compared to baseline using the paired Student t test for continuous data and McNemar test for categorical variables.	<b>Yes 1</b>	No 2
Clinical significance	- The CP Stent can effectively treat and potentially prevent aortic wall injury associated with aortic coarctation. Access site arterial injury is the most common important complication. Longer-term follow up is necessary to define mid- and late-term outcomes.	<b>Yes 1</b>	No 2
Data Contribution Grade (Range 4-8)		4	



## NuMED Summary of Safety and Clinical Performance SSCP – Stents – CoA & RVOT

### Overall S&P Appraisal, Disposition and Weighting

S&P Grade (Range 9-25)	LOE (2) + Suitability (4) + Data Contribution (4) = 10	Disposition and Weighting (select)	<b>Accepted and Pivotal 9-12</b> Accepted but not Pivotal, 13-21 Excluded, 22-25
---------------------------	---	------------------------------------	--

**Purpose:** To evaluate safety and short-term efficacy of the CP Stent in treating or preventing aortic wall injury in patients with CoA.

**Clinical Study Methodology:** Single arm interventional study. Patients were enrolled if they had a history of CoA with pre-existing aortic wall injury (Treatment group) or with increased risk of aortic wall injury (Prevention group). Pre/post-implant hemodynamics and angiography were reported. A core laboratory performed standardized review of all angiograms. One-month follow-up was reported.

**Reference to the clinical study plan (and amendment) n°:** NCT01278303

<b>Investigation Site:</b> 19 pediatric cardiac centers in United States	<b>Ethics Committee Approvals:</b> Johns Hopkins Institutional Review Board and Institutional review boards of all participating centers.	<b>Regulatory Authority Approvals:</b> Investigational Device Exemption
<b>Reference to Approved Consent Forms:</b> N/A	<b>Reference to Document n°:</b> N/A	<b>Reference to Documents n°:</b> G060057a

**Patient Population:** Patients with aortic coarctation at risk of aortic wall injury or with existing aortic wall injury. A total of 158 patients (83 treatment cohort and 75 prevention cohort, median age 19 years (range from 5 to 70 years) and with 103 males and 55 females.

**Clinical Study Results:** Results held by Sponsor.

Purpose	Criteria	Results
Short term efficacy	Blood pressure gradient (at 1 month)	All: from 24 ± 26 mmHg to -1 ± 15 mmHg Treatment group: from 14 ± 24 to -2 ± 14 mmHg Prevention group: from 35 ± 23 to 1 ± 15 mmHg
Safety	Adverse events	17 adverse events; 2 serious (dissection of the iliac artery) and 15 somewhat serious. No deaths. Device related AEs included local stent migration (n=1) and stent malposition (n=1).

**Reference to the Clinical Study Report n°:** NCT01278303

**Device Used:** Covered CP Stent by NuMED, pre-mounted on BIB stent placement catheter

**Conclusion:** The CP Stent can effectively treat and potentially prevent aortic wall injury associated with aortic coarctation. Access site arterial injury is the most common important complication. Longer-term follow up is necessary to define mid- and late-term outcomes.

**Clinical Publication:** Taggart NW, Minahan M, Cabalka AK, Cetta F, Usmani K, Ringel RE, Immediate Outcomes of Covered Stent Placement for Treatment or Prevention of Aortic Wall Injury Associated With Coarctation of the Aorta (COAST II). JACC Cardiovasc Interv 9, 484-493 (2016).

### 3. Study name: PARCS

#### Appraisal

Level of Evidence	Study Method/Design	Question Applied	Oxford LOE 2011				
			1	2	3	4	5
	Prospective, multicenter, single-arm pivotal clinical trial.	To evaluate the safety and efficacy of the CP Stent (Covered) for treatment of conduit injury occurring during TPVR. The study also evaluated immediate and short-term TPVR function in patients receiving CP covered stents.					

Suitability	Relevant Data	Grading		
Device	- CP Stent (Covered, pre-mounted on BIB)	D1	D2	D3
Application	- Treatment of RVOT conduit injury occurring during TPVR	A1	A2	A3
Patient	- Patients: patients requiring treatment of RVOT conduit injury occurring during TPVR - Sampling: 120 (patients had a wall injury identified and treated with CP Stent (Covered). - Pivotal trial: 50 - Continued access: 70	P1	P2	P3



## NuMED Summary of Safety and Clinical Performance SSCP – Stents – CoA & RVOT

	<ul style="list-style-type: none"> <li>- Mean age:               <ul style="list-style-type: none"> <li>- Pivotal trial: 17 (6-44) year</li> <li>- Continued access: 16 (7-49) year</li> </ul> </li> <li>- Sex:               <ul style="list-style-type: none"> <li>- Pivotal trial: 28M; 22F</li> <li>- Continued access: 40M; 30F</li> </ul> </li> </ul>			
Report	- High quality	<b>R1</b>	R2	R3
Suitability Grade (Range 4-12)		4		

Data Contribution	Relevant Data	Grading	
Outcomes/Endpoints	<ul style="list-style-type: none"> <li>- CP Stent (Covered) implant success</li> <li>- Stent-related AEs</li> </ul>	<b>Yes 1</b>	No 2
Follow-up	- 6 months (short term follow-up)	Yes 1	<b>No 2</b>
Statistical analysis	- All results were considered statistically significant at $P \leq 0.05$ . SAS v9.4 (SAS Institute, Cary, NC) was used for all calculations.	<b>Yes 1</b>	No 2
Clinical significance	- The study results demonstrate the safety and efficacy of use of the covered CP Stent when used for pre-stenting in the RVOT prior to Melody TPV implantation.	<b>Yes 1</b>	No 2
Data Contribution Grade (Range 4-8)		5	

### Overall S&P Appraisal, Disposition and Weighting

S&P Grade (Range 9-25)	LOE (2) + Suitability (4) + Data Contribution (5) = 11	Disposition and Weighting (select)	<b>Accepted and Pivotal 9-12</b> Accepted but not Pivotal, 13-21 Excluded, 22-25
---------------------------	---	---------------------------------------	--

**Purpose:** Evaluation of the Covered CP Stent for repair of tears that occur in the pulmonary artery during dilation (enlargement) of a conduit (passageway) connecting the right ventricle of the heart to the pulmonary arteries.

**Clinical Study Methodology:** Single arm prospective study. The PARCS trial was a prospective, multicenter, single-arm pivotal clinical trial. Forty US centers participated in either the pivotal trial (22 centers) or the continued access protocol, which immediately followed the pivotal trial during Food and Drug Administration submission. If no conduit wall injury occurred during the procedure, the patient was considered a screen failure. If at any point during the procedure, including before intervention, the implanting physician identified an area of wall injury, a CCPS could then be selected and implanted.

**Reference to the clinical study plan (and amendment) n°:** NCT01824160

<p><b>Investigation Site:</b> 38 cardiac centers in United States</p> <p><b>Reference to Approved Consent Forms:</b> N/A</p>	<p><b>Ethics Committee Approvals:</b> The Johns Hopkins Institutional Review Board approved the trial, and all participating institutions received local Institutional Review Board approval before participation.</p> <p><b>Reference to Document n°:</b> N/A</p>	<p><b>Regulatory Authority Approvals:</b> IDE</p> <p><b>Reference to Documents n°:</b> G120188</p>
--	--	--

**Patient Population:** Participants receiving a Covered CP stent for repair of pulmonary artery injury. In the pivotal trial, fifty patients met the inclusion criteria, mean age 17 years (range from 6 to 44 years) and 56% of male patients. In the continued access, seventy patients with mean age of 16 years (range from 7 to 49 years) and 57% of male patients.

### Clinical Study Results: Results held by Sponsor.

Purpose	Criteria	Results
Pivotal (n=50): Severity of illness	Median improvement by at least 1 level from baseline to post-procedure	Median improvement of 1 level in Severity of Illness Score
Pivotal (n=50): Procedure success	≥ 75% patients, based on both device success and lesion success	Procedure success achieved in 68% of patients
Pivotal (n=50): Successful implantation of the Melody TPV	Coverage of conduit disruption defined as either no residual disruption or contained disruption, followed by successful implantation	Successful implantation achieved in 83% of patients



**NuMED**  
**Summary of Safety and Clinical Performance**  
**SSCP – Stents – CoA & RVOT**

	of Melody valve in $\geq 80\%$ patients	
Pivotal (n=50): Adverse events attributed to covered CP Stent within 30 days	$\geq 80\%$ patients free of adverse events attributed to the covered CP Stent within 30 days	At least 80% were free of an adverse event attributed to the covered CP Stent. There was 1 report of stent malposition where the stent became dislodged and migrated into the pulmonary arteries
All patients (n=120): Performance	Covered CP Stent Implant success	CCPS implants successfully treated 95% of conduit injuries with either no or minimal residual conduit wall injury.  Melody TPVR was successfully performed in 94% of the enrolled cohort, and TPV function was not adversely affected by placement within the CCPS substrate, with 6-month follow-up data comparing favorably with other previously published cohorts.
All patients (n=120): Safety	Stent-related AEs	AEs that specifically related to the CCPS and its implantation were uncommon. One serious (stent malposition) and one somewhat serious (stent embolization) AE occurred (both in the same patient who is described above). A device usage issue was identified whereby the expanded poly tetrafluoroethylene covering separated from the stent during attempts to load the CCPS device into the delivery sheath. This was identified before deployment; the stent was removed and replaced with a new CCPS without consequence to the patient.

**Reference to the Clinical Study Report n°:** NCT01824160

**Device Used:** Covered Mounted CP Stent

**Conclusion:** The study results demonstrate the safety and efficacy of use of the covered CP Stent when used for pre-stenting in the RVOT prior to Melody TPV implantation.

**Clinical Publication:** Delaney JW, Goldstein BH, Bishnoi RN, Bisselou KSM, McEnaney K, Minahan M, Ringel RE; PARCS Investigators. Covered CP stent for treatment of right ventricular conduit injury during melody transcatheter pulmonary valve replacement, *Circ Cardiovasc Interv* **11(10)**, 1-10 (2018).



# NuMED

## Summary of Safety and Clinical Performance

### SSCP – Stents – CoA & RVOT

**Summary of clinical data from other sources:**

First Author (Year)	Appraisal/Results																																							
1. Delaney et al. (2018) <table border="1" style="margin-top: 10px; width: 100%; border-collapse: collapse;"> <thead> <tr style="background-color: #e0e0e0;"> <th colspan="2">Contribution</th> </tr> </thead> <tbody> <tr> <td style="width: 50%;">S&amp;P</td> <td style="text-align: center;">x</td> </tr> <tr> <td>SOA</td> <td style="text-align: center;">x</td> </tr> </tbody> </table>	Contribution		S&P	x	SOA	x	<p><b>Safety &amp; Performance</b></p> <p>This publication presents the results from the PARCS trial – Covered CP Stent for Treatment of Right Ventricular Conduit Injury During Melody Transcatheter Pulmonary Valve Replacement (NCT01824160). Please refer information presented in <b>Table G-1</b> for safety and performance of the subject devices, Study no. 3. The state of the art information is presented below.</p> <p><b>State of the Art</b></p> <p><b>Appraisal</b></p> <table border="1" style="width: 100%; border-collapse: collapse; text-align: center;"> <thead> <tr style="background-color: #e0e0e0;"> <th colspan="2">Medical condition</th> <th colspan="2">Alternatives</th> <th colspan="2">Risk/benefit</th> <th colspan="2">Side-effects</th> <th colspan="2">Equivalence</th> <th colspan="2">Surrogate endpoints</th> </tr> </thead> <tbody> <tr> <td><b>Yes 1</b></td> <td>No 2</td> <td><b>Yes 1</b></td> <td>No 2</td> <td><b>Yes 1</b></td> <td>No 2</td> <td><b>Yes 1</b></td> <td>No 2</td> <td>Yes 1</td> <td><b>No 2</b></td> <td>Yes 1</td> <td><b>No 2</b></td> </tr> </tbody> </table> <p><b>Overall SOA Appraisal and Disposition</b></p> <table border="1" style="width: 100%; border-collapse: collapse; text-align: center;"> <thead> <tr style="background-color: #e0e0e0;"> <th>SOA Grade (Range 6-12)</th> <th>Disposition (select)</th> </tr> </thead> <tbody> <tr> <td>8</td> <td><b>Accepted, &lt; 12</b> Excluded, 12</td> </tr> </tbody> </table> <p><b>Relevant SOA Results</b></p> <table border="1" style="width: 100%; border-collapse: collapse;"> <thead> <tr style="background-color: #e0e0e0;"> <th style="width: 30%;">SOA data</th> <th></th> </tr> </thead> <tbody> <tr> <td></td> <td style="vertical-align: top; padding: 5px;"> <ul style="list-style-type: none"> <li>- Current knowledge:               <ul style="list-style-type: none"> <li>o Important conduit injury can occur during ultrahigh pressure angioplasty. Ultra-high pressure angioplasty is often required to dilate conduits effectively for TPVR. Conduit injury, once identified, could preclude further dilation of the conduit out of concern for extension of the area of injury.</li> <li>- Stenting of the conduit before valve implantation improves the durability of the implanted valve.</li> <li>- Covered stents have been used in the vascular space to isolate areas of injury.</li> </ul> </li> <li>- RVOT reconstruction:               <ul style="list-style-type: none"> <li>o RVOT reconstruction with a valved conduit or bioprosthetic pulmonary valve placement is necessary during surgical repair of a substantial subset of patients with congenital heart disease.</li> <li>o All valved RVOT substrates, regardless of type, have been associated with functional deterioration, with between 50% and 80% requiring replacement by 10 years.</li> <li>o RVOT dysfunction may be associated with substantial patient morbidity and even mortality.</li> <li>o Transcatheter RVOT conduit rehabilitation using high-pressure angioplasty with or without stent placement has been utilized to delay or defer the need for surgical pulmonary valve replacement. An injury within the wall of the conduit is likely to occur with any successful conduit dilation, although minor injuries may not be clinically relevant or recognized with angiography.</li> <li>o Successful RVOT conduit angioplasty often requires the use of ultrahigh pressure noncompliant balloons to effectively relieve the stenosis but with a higher rate of recognized conduit injury (≤33%). The vast majority of these injuries was not associated with hemodynamic compromise.</li> </ul> </li> <li>- Melody transcatheter pulmonary valve:               <ul style="list-style-type: none"> <li>o Introduction of the Melody transcatheter pulmonary valve (TPV; Medtronic) led to more frequent percutaneous conduit rehabilitation because cardiologists could effectively treat both stenosis and insufficiency, without the need for open-heart surgery.</li> </ul> </li> </ul> </td> </tr> </tbody> </table>		Medical condition		Alternatives		Risk/benefit		Side-effects		Equivalence		Surrogate endpoints		<b>Yes 1</b>	No 2	<b>Yes 1</b>	No 2	<b>Yes 1</b>	No 2	<b>Yes 1</b>	No 2	Yes 1	<b>No 2</b>	Yes 1	<b>No 2</b>	SOA Grade (Range 6-12)	Disposition (select)	8	<b>Accepted, &lt; 12</b> Excluded, 12	SOA data			<ul style="list-style-type: none"> <li>- Current knowledge:               <ul style="list-style-type: none"> <li>o Important conduit injury can occur during ultrahigh pressure angioplasty. Ultra-high pressure angioplasty is often required to dilate conduits effectively for TPVR. Conduit injury, once identified, could preclude further dilation of the conduit out of concern for extension of the area of injury.</li> <li>- Stenting of the conduit before valve implantation improves the durability of the implanted valve.</li> <li>- Covered stents have been used in the vascular space to isolate areas of injury.</li> </ul> </li> <li>- RVOT reconstruction:               <ul style="list-style-type: none"> <li>o RVOT reconstruction with a valved conduit or bioprosthetic pulmonary valve placement is necessary during surgical repair of a substantial subset of patients with congenital heart disease.</li> <li>o All valved RVOT substrates, regardless of type, have been associated with functional deterioration, with between 50% and 80% requiring replacement by 10 years.</li> <li>o RVOT dysfunction may be associated with substantial patient morbidity and even mortality.</li> <li>o Transcatheter RVOT conduit rehabilitation using high-pressure angioplasty with or without stent placement has been utilized to delay or defer the need for surgical pulmonary valve replacement. An injury within the wall of the conduit is likely to occur with any successful conduit dilation, although minor injuries may not be clinically relevant or recognized with angiography.</li> <li>o Successful RVOT conduit angioplasty often requires the use of ultrahigh pressure noncompliant balloons to effectively relieve the stenosis but with a higher rate of recognized conduit injury (≤33%). The vast majority of these injuries was not associated with hemodynamic compromise.</li> </ul> </li> <li>- Melody transcatheter pulmonary valve:               <ul style="list-style-type: none"> <li>o Introduction of the Melody transcatheter pulmonary valve (TPV; Medtronic) led to more frequent percutaneous conduit rehabilitation because cardiologists could effectively treat both stenosis and insufficiency, without the need for open-heart surgery.</li> </ul> </li> </ul>
	Contribution																																							
	S&P	x																																						
	SOA	x																																						
	Medical condition		Alternatives		Risk/benefit		Side-effects		Equivalence		Surrogate endpoints																													
	<b>Yes 1</b>	No 2	<b>Yes 1</b>	No 2	<b>Yes 1</b>	No 2	<b>Yes 1</b>	No 2	Yes 1	<b>No 2</b>	Yes 1	<b>No 2</b>																												
	SOA Grade (Range 6-12)	Disposition (select)																																						
	8	<b>Accepted, &lt; 12</b> Excluded, 12																																						
	SOA data																																							
		<ul style="list-style-type: none"> <li>- Current knowledge:               <ul style="list-style-type: none"> <li>o Important conduit injury can occur during ultrahigh pressure angioplasty. Ultra-high pressure angioplasty is often required to dilate conduits effectively for TPVR. Conduit injury, once identified, could preclude further dilation of the conduit out of concern for extension of the area of injury.</li> <li>- Stenting of the conduit before valve implantation improves the durability of the implanted valve.</li> <li>- Covered stents have been used in the vascular space to isolate areas of injury.</li> </ul> </li> <li>- RVOT reconstruction:               <ul style="list-style-type: none"> <li>o RVOT reconstruction with a valved conduit or bioprosthetic pulmonary valve placement is necessary during surgical repair of a substantial subset of patients with congenital heart disease.</li> <li>o All valved RVOT substrates, regardless of type, have been associated with functional deterioration, with between 50% and 80% requiring replacement by 10 years.</li> <li>o RVOT dysfunction may be associated with substantial patient morbidity and even mortality.</li> <li>o Transcatheter RVOT conduit rehabilitation using high-pressure angioplasty with or without stent placement has been utilized to delay or defer the need for surgical pulmonary valve replacement. An injury within the wall of the conduit is likely to occur with any successful conduit dilation, although minor injuries may not be clinically relevant or recognized with angiography.</li> <li>o Successful RVOT conduit angioplasty often requires the use of ultrahigh pressure noncompliant balloons to effectively relieve the stenosis but with a higher rate of recognized conduit injury (≤33%). The vast majority of these injuries was not associated with hemodynamic compromise.</li> </ul> </li> <li>- Melody transcatheter pulmonary valve:               <ul style="list-style-type: none"> <li>o Introduction of the Melody transcatheter pulmonary valve (TPV; Medtronic) led to more frequent percutaneous conduit rehabilitation because cardiologists could effectively treat both stenosis and insufficiency, without the need for open-heart surgery.</li> </ul> </li> </ul>																																						



## NuMED

### Summary of Safety and Clinical Performance

#### SSCP – Stents – CoA & RVOT

		<ul style="list-style-type: none"> <li>○ Melody TPV implants, without stent reinforcement of the conduit before valve implant, have been associated with a high rate of progressive valve deformity and stent fracture leading to valvular dysfunction.</li> <li>○ Conduit wall injury is a known complication of isolated or serial balloon angioplasty of the RVOT conduit. Although bare metal stents may provide some reinforcement of a damaged conduit wall, they are not likely to allow for safe, continued dilation of an injured RVOT conduit that has not been fully prepared (e.g., left with hemodynamically important residual stenosis) for TPVR, and they are not anticipated to be effective in treating catastrophic conduit injuries.</li> <li>○ Covered CP Stent (NuMED) is a balloon-expandable, large-diameter, covered stent whose construction and applications for vascular wall injury, tears, or leak have been reported previously. Experience with the Covered CP Stent outside of the United States is extensive and has included its routine use in the pre-stenting process for valve implantation. The European experience has suggested that this practice may reduce the clinical impact of conduit injury.</li> <li>○ Some US centers did have access to the Covered CP Stent as participants in the COAST (Coarctation of the Aorta Stent Trial) and could apply for emergency use if an unexpected RVOT wall injury occurred. Non-COAST centers could apply for a single-patient compassionate use exemption if they felt a patient was at high risk for conduit injury.</li> </ul>												
	Comments	- High-pressure balloon and stent angioplasty are frequently necessary to prepare the dysfunctional RVOT conduit before transcatheter pulmonary valve replacement (TPVR). Conduit injury can result, which may be catastrophic to the patient or prevent successful TPVR. Severe conduit injury was found to be rare but unpredictable. The covered stent was effective in either treating or mitigating this problem. The vast majority of patients, even with identified conduit injury, was able to complete the valve replacement procedure. The covered stent did not interfere with Melody valve function at short-term, 6-month follow-up.												
<b>Safety &amp; Performance (for safety only)</b>														
<b>Appraisal</b>														
2. Baykan et al. (2018) <table border="1" style="margin-top: 10px; width: 100%;"> <tr> <th colspan="2">Contribution</th> </tr> <tr> <td>S&amp;P</td> <td>X (S only)</td> </tr> <tr> <td>SOA</td> <td>-</td> </tr> </table>	Contribution		S&P	X (S only)	SOA	-	Level of Evidence	Study Method/Design	Question Applied	Oxford LOE 2011				
	Contribution													
S&P	X (S only)													
SOA	-													
		Control study. Study group was composed of 20 CoA patients who were treated with CP Stent between the dates October 2008 and February 2015, and control group was composed of 20 healthy children with age and sex matched.	To address the presence of hypertension and risk for cardiovascular diseases in patients with CoA who were treated with endovascular stent placement.	1	2	3	4	5						
	Suitability	Relevant Data	Grading											
	Device	- CP Stents (Bare and Covered) - Unknown whether pre-mounted on BIB	<b>D1</b>	D2	D3									
	Application	- CoA	<b>A1</b>	A2	A3									
	Patient	- Patients who had undergone stent placement for CoA compared with control group (healthy children with age and sex matched).	<b>P1</b>	P2	P3									



## NuMED Summary of Safety and Clinical Performance SSCP – Stents – CoA & RVOT

	<ul style="list-style-type: none"> <li>- Sampling: n=20 CoA and n=20 healthy children</li> <li>- Mean age:               <ul style="list-style-type: none"> <li>- CoA group: 14.2 (SD: 3.9) years</li> <li>- Control group: 13.7 (SD: 2.7) years</li> </ul> </li> <li>- Sex:               <ul style="list-style-type: none"> <li>- CoA group: 16M; 4F</li> <li>- Control group: 15M; 5F</li> </ul> </li> </ul>			
Report	- High quality	<b>R1</b>	R2	R3
Suitability Grade (Range 4-12)		4		
Data Contribution	Relevant Data	Grading		
Outcomes/Endpoints	- Ambulatory blood pressure	<b>Yes 1</b>	No 2	
Follow-up	- 6 months and 6 years	<b>Yes 1</b>	No 2	
Statistical analysis	- Student t-test was used if the two independent group comparisons were normal and the Mann-Whitney U test was used if the normal distribution was not present. Pearson chi-square analysis was performed to determine whether there was a difference in categorical variables between the case and control groups.	<b>Yes 1</b>	No 2	
Clinical significance	<ul style="list-style-type: none"> <li>- It was shown that hypertension incidence as demonstrated by ambulatory blood pressure monitorization and risk for cardiovascular diseases as indicated by carotid intima media thickness and pulse wave velocity were higher than those in healthy population even after CoA is corrected.</li> <li>- CoA should be carefully monitored for hypertension, even if it has been completely corrected by any method. This study suggests that CoA is a part of generalized vasculopathy rather than being a localized narrowing.</li> </ul>	<b>Yes 1</b>	No 2	
Data Contribution Grade (Range 4-8)		4		
<b>Overall S&amp;P Appraisal, Disposition and Weighting</b>				
S&P Grade (Range 9-25)	LOE (3) + Suitability (4) + Data Contribution (4) = 11	Disposition and Weighting (select)	<b>Accepted and Pivotal 9-12</b> Accepted but not Pivotal, 13-21 Excluded, 22-25	
<b>Relevant S&amp;P Results</b>				
Safety data	<ul style="list-style-type: none"> <li>- Hypertensive:               <ul style="list-style-type: none"> <li>- Daytime: 5% were hypertensive and 20% were pre-hypertensive in the study group compared to 0% in the control group.</li> <li>- Night: 15% were hypertensive and 15% were pre-hypertensive in the study group compared to 0% in the control group.</li> </ul> </li> </ul>			
Benefits/claims data	- N/A			
Strengths	- N/A			
Weaknesses/ Potential bias	- Patients were treated only with “NuMED brand Bare and Covered Stent” types. In the future the authors can do more extensive studies with more cases and different types of stents.			



## NuMED Summary of Safety and Clinical Performance SSCP – Stents – CoA & RVOT

		<ul style="list-style-type: none"> <li>- Same methodology (blood pressure monitoring with Holter device) in pre-operative period could not be used because at that time they did not have a blood pressure Holter device.</li> </ul>																																																																																					
<p>3. Morgan et al. (2017)</p> <table border="1" style="width: 100%; margin-top: 10px;"> <thead> <tr> <th colspan="2">Contribution</th> </tr> </thead> <tbody> <tr> <td>S&amp;P</td> <td style="text-align: center;">x</td> </tr> <tr> <td>SOA</td> <td style="text-align: center;">x</td> </tr> </tbody> </table>	Contribution		S&P	x	SOA	x	<p><b>Safety &amp; Performance Appraisal</b></p> <table border="1" style="width: 100%; border-collapse: collapse;"> <thead> <tr> <th style="width: 25%;">Level of Evidence</th> <th style="width: 30%;">Study Method/Design</th> <th style="width: 30%;">Question Applied</th> <th colspan="5">Oxford LOE 2011</th> </tr> </thead> <tbody> <tr> <td></td> <td>Retrospective data collected of the first NuDEL delivery systems used in patients from three centers (UK and Ireland).</td> <td>To evaluate the first-in-man use of a new system (NuDEL) for implantation of CP Stent (Covered) in patients with complex structural and CHD.</td> <td style="text-align: center;">1</td> <td style="text-align: center;">2</td> <td style="text-align: center;">3</td> <td style="text-align: center;">4</td> <td style="text-align: center;">5</td> </tr> </tbody> </table> <table border="1" style="width: 100%; border-collapse: collapse;"> <thead> <tr> <th style="width: 25%;">Suitability</th> <th style="width: 50%;">Relevant Data</th> <th colspan="3">Grading</th> </tr> </thead> <tbody> <tr> <td>Device</td> <td>- NuDEL Delivery System</td> <td style="text-align: center;"><b>D1</b></td> <td style="text-align: center;">D2</td> <td style="text-align: center;">D3</td> </tr> <tr> <td>Application</td> <td>- CoA and RVOT</td> <td style="text-align: center;"><b>A1</b></td> <td style="text-align: center;">A2</td> <td style="text-align: center;">A3</td> </tr> <tr> <td>Patient</td> <td> <ul style="list-style-type: none"> <li>- Patients with COA and RVOT</li> <li>- Sampling: n=12 (13 CP Stents, Covered, delivered via 12 NuDELs); with 6 CoA, 5 RVOT, and 1 with severe stenosis of a Mustard systemic venous baffle. Note: "P2" due to one with severe stenosis of a Mustard systemic venous baffle.</li> <li>- Age: 10-43 years</li> <li>- Sex: Not reported</li> </ul> </td> <td style="text-align: center;">P1</td> <td style="text-align: center;"><b>P2</b></td> <td style="text-align: center;">P3</td> </tr> <tr> <td>Report</td> <td>- High quality with minor deficiency as device performance is based on descriptive information.</td> <td style="text-align: center;">R1</td> <td style="text-align: center;"><b>R2</b></td> <td style="text-align: center;">R3</td> </tr> <tr> <td colspan="2" style="text-align: right;">Suitability Grade (Range 4-12)</td> <td colspan="3" style="text-align: center;">6</td> </tr> </tbody> </table> <table border="1" style="width: 100%; border-collapse: collapse;"> <thead> <tr> <th style="width: 25%;">Data Contribution</th> <th style="width: 50%;">Relevant Data</th> <th colspan="2">Grading</th> </tr> </thead> <tbody> <tr> <td>Outcomes/Endpoints</td> <td> <ul style="list-style-type: none"> <li>- Procedure complications.</li> <li>- Ease of use.</li> </ul> </td> <td style="text-align: center;"><b>Yes 1</b></td> <td style="text-align: center;">No 2</td> </tr> <tr> <td>Follow-up</td> <td>- Not reported.</td> <td style="text-align: center;">Yes 1</td> <td style="text-align: center;"><b>No 2</b></td> </tr> <tr> <td>Statistical analysis</td> <td>- Not reported.</td> <td style="text-align: center;">Yes 1</td> <td style="text-align: center;"><b>No 2</b></td> </tr> <tr> <td>Clinical significance</td> <td>- NuDEL system is a safe and effective means of covered stent deployment in challenging anatomy.</td> <td style="text-align: center;"><b>Yes 1</b></td> <td style="text-align: center;">No 2</td> </tr> <tr> <td colspan="2" style="text-align: right;">Data Contribution Grade (Range 4-8)</td> <td colspan="2" style="text-align: center;">6</td> </tr> </tbody> </table> <p><b>Overall S&amp;P Appraisal, Disposition and Weighting</b></p> <table border="1" style="width: 100%; border-collapse: collapse;"> <tr> <td style="width: 25%;">S&amp;P Grade (Range 9-25)</td> <td style="width: 25%;">LOE (4) + Suitability (6) + Data Contribution (6) = 16</td> <td style="width: 25%;">Disposition and Weighting (select)</td> <td style="width: 25%;">Accepted and Pivotal 9-12 <b>Accepted but not Pivotal, 13-21</b> Excluded, 22-25</td> </tr> </table> <p><b>Relevant S&amp;P Results</b></p> <table border="1" style="width: 100%; border-collapse: collapse;"> <tr> <td style="width: 25%;">Safety data</td> <td>- No procedural complications and no reports of equipment failure or dysfunction.</td> </tr> <tr> <td style="width: 25%;">Performance data</td> <td> <ul style="list-style-type: none"> <li>- The system required minimal preparation – flushing only; therefore, despite a lack of familiarity, it was ready for deployment in each case in under two minutes after removing the packaging.</li> <li>- Key positive feedback and observations were that the assembly tracked well through the access site and through tortuous and narrowed anatomy in either the outflow tract or the descending aorta.</li> </ul> </td> </tr> </table>			Level of Evidence	Study Method/Design	Question Applied	Oxford LOE 2011						Retrospective data collected of the first NuDEL delivery systems used in patients from three centers (UK and Ireland).	To evaluate the first-in-man use of a new system (NuDEL) for implantation of CP Stent (Covered) in patients with complex structural and CHD.	1	2	3	4	5	Suitability	Relevant Data	Grading			Device	- NuDEL Delivery System	<b>D1</b>	D2	D3	Application	- CoA and RVOT	<b>A1</b>	A2	A3	Patient	<ul style="list-style-type: none"> <li>- Patients with COA and RVOT</li> <li>- Sampling: n=12 (13 CP Stents, Covered, delivered via 12 NuDELs); with 6 CoA, 5 RVOT, and 1 with severe stenosis of a Mustard systemic venous baffle. Note: "P2" due to one with severe stenosis of a Mustard systemic venous baffle.</li> <li>- Age: 10-43 years</li> <li>- Sex: Not reported</li> </ul>	P1	<b>P2</b>	P3	Report	- High quality with minor deficiency as device performance is based on descriptive information.	R1	<b>R2</b>	R3	Suitability Grade (Range 4-12)		6			Data Contribution	Relevant Data	Grading		Outcomes/Endpoints	<ul style="list-style-type: none"> <li>- Procedure complications.</li> <li>- Ease of use.</li> </ul>	<b>Yes 1</b>	No 2	Follow-up	- Not reported.	Yes 1	<b>No 2</b>	Statistical analysis	- Not reported.	Yes 1	<b>No 2</b>	Clinical significance	- NuDEL system is a safe and effective means of covered stent deployment in challenging anatomy.	<b>Yes 1</b>	No 2	Data Contribution Grade (Range 4-8)		6		S&P Grade (Range 9-25)	LOE (4) + Suitability (6) + Data Contribution (6) = 16	Disposition and Weighting (select)	Accepted and Pivotal 9-12 <b>Accepted but not Pivotal, 13-21</b> Excluded, 22-25	Safety data	- No procedural complications and no reports of equipment failure or dysfunction.	Performance data	<ul style="list-style-type: none"> <li>- The system required minimal preparation – flushing only; therefore, despite a lack of familiarity, it was ready for deployment in each case in under two minutes after removing the packaging.</li> <li>- Key positive feedback and observations were that the assembly tracked well through the access site and through tortuous and narrowed anatomy in either the outflow tract or the descending aorta.</li> </ul>
	Contribution																																																																																						
	S&P	x																																																																																					
	SOA	x																																																																																					
	Level of Evidence	Study Method/Design	Question Applied	Oxford LOE 2011																																																																																			
		Retrospective data collected of the first NuDEL delivery systems used in patients from three centers (UK and Ireland).	To evaluate the first-in-man use of a new system (NuDEL) for implantation of CP Stent (Covered) in patients with complex structural and CHD.	1	2	3	4	5																																																																															
	Suitability	Relevant Data	Grading																																																																																				
	Device	- NuDEL Delivery System	<b>D1</b>	D2	D3																																																																																		
	Application	- CoA and RVOT	<b>A1</b>	A2	A3																																																																																		
	Patient	<ul style="list-style-type: none"> <li>- Patients with COA and RVOT</li> <li>- Sampling: n=12 (13 CP Stents, Covered, delivered via 12 NuDELs); with 6 CoA, 5 RVOT, and 1 with severe stenosis of a Mustard systemic venous baffle. Note: "P2" due to one with severe stenosis of a Mustard systemic venous baffle.</li> <li>- Age: 10-43 years</li> <li>- Sex: Not reported</li> </ul>	P1	<b>P2</b>	P3																																																																																		
Report	- High quality with minor deficiency as device performance is based on descriptive information.	R1	<b>R2</b>	R3																																																																																			
Suitability Grade (Range 4-12)		6																																																																																					
Data Contribution	Relevant Data	Grading																																																																																					
Outcomes/Endpoints	<ul style="list-style-type: none"> <li>- Procedure complications.</li> <li>- Ease of use.</li> </ul>	<b>Yes 1</b>	No 2																																																																																				
Follow-up	- Not reported.	Yes 1	<b>No 2</b>																																																																																				
Statistical analysis	- Not reported.	Yes 1	<b>No 2</b>																																																																																				
Clinical significance	- NuDEL system is a safe and effective means of covered stent deployment in challenging anatomy.	<b>Yes 1</b>	No 2																																																																																				
Data Contribution Grade (Range 4-8)		6																																																																																					
S&P Grade (Range 9-25)	LOE (4) + Suitability (6) + Data Contribution (6) = 16	Disposition and Weighting (select)	Accepted and Pivotal 9-12 <b>Accepted but not Pivotal, 13-21</b> Excluded, 22-25																																																																																				
Safety data	- No procedural complications and no reports of equipment failure or dysfunction.																																																																																						
Performance data	<ul style="list-style-type: none"> <li>- The system required minimal preparation – flushing only; therefore, despite a lack of familiarity, it was ready for deployment in each case in under two minutes after removing the packaging.</li> <li>- Key positive feedback and observations were that the assembly tracked well through the access site and through tortuous and narrowed anatomy in either the outflow tract or the descending aorta.</li> </ul>																																																																																						





## NuMED Summary of Safety and Clinical Performance SSCP – Stents – CoA & RVOT

Benefits/claims data	<ul style="list-style-type: none"> <li>- Stent was easy to uncover, and the markers on the system provided added re-assurance to this process.</li> <li>- Most difficult part of the catheterization procedure is getting the stent into the required anatomical position before deploying it. To this end, the NuDEL system has been developed.</li> <li>- NuDEL reported to require minimal preparation and tracked well through the access site and tortuous and narrowed anatomy.</li> </ul>
Strengths	<ul style="list-style-type: none"> <li>- Our initial series suggests that the NuDEL system provides a safe, efficient method of deploying a covered stent in patients with complex outflow tract stenosis and those with CoA. Using this system avoids some of the pitfalls associated with stent mounting and management of the stent–balloon–delivery system complex.</li> <li>- Stent was easy to uncover and markers on the system provided added re-assurance of the process.</li> </ul>
Weaknesses/ Potential bias	<ul style="list-style-type: none"> <li>- Conflict of interest: None</li> <li>- Financial support: Research received no specific grant from any funding agency or from commercial or not-for-profit sectors</li> </ul>

### State of the Art

#### Appraisal

Medical condition		Alternatives		Risk/benefit		Side-effects		Equivalence		Surrogate endpoints	
Yes 1	No 2	Yes 1	No 2	Yes 1	No 2	Yes 1	No 2	Yes 1	No 2	Yes 1	No 2

#### Overall SOA Appraisal and Disposition

SOA Grade (Range 6-12)	10	Disposition (select)	<b>Accepted, &lt; 12</b> Excluded, 12
---------------------------	----	----------------------	--

#### Relevant SOA Results

SOA data	<ul style="list-style-type: none"> <li>- Conduit rupture is an anxiety-provoking potential complication; the availability of a “ready-to-go” covered stent system may provide an attractive emergency backup. This may be of benefit to operators who perform a low volume of large-caliber stent procedures and are not conversant with the techniques involved, even in the elective setting.</li> <li>- The range of stents available for these therapies has developed well over the last 10–15 years, allowing authors to make semi-quantitative decisions about stent choice for each individual case.</li> <li>- Safety and accuracy of deployment are at least partially dependent on the precise mounting of the stent on its delivery balloon and passing it into and along the delivery sheath to its required position. The most difficult part of the catheterization procedure is getting the stent into the required anatomical position before deploying it. A lot of time is taken in getting these essential steps right, and there is potential for safety, efficiency, and efficacy problems at every step. Slipping of the stents off balloons leading to migration of the stent before or during deployment and damage to the balloon or stent during mounting are some of the problems encountered, which may require re-crossing of the target areas, lead to vascular risk of removing and re-inserting a long, large-caliber sheath, and at worst can have major safety consequences.</li> </ul>
Comments	<ul style="list-style-type: none"> <li>- Stent implantation in the RVOT and for the treatment of CoA has become standard practice for congenital interventional cardiologists.</li> </ul>



## NuMED Summary of Safety and Clinical Performance SSCP – Stents – CoA & RVOT

<p>4. Bishnoi et al. (2015)</p> <table border="1" style="margin-top: 10px; width: 100%; border-collapse: collapse;"> <thead> <tr> <th colspan="2" style="text-align: left;">Contribution</th> </tr> </thead> <tbody> <tr> <td style="width: 30%;">S&amp;P</td> <td style="text-align: center;">x</td> </tr> <tr> <td>SOA</td> <td style="text-align: center;">x</td> </tr> </tbody> </table>	Contribution		S&P	x	SOA	x	<b>Safety &amp; Performance Appraisal</b>				
	Contribution										
	S&P	x									
	SOA	x									
	Level of Evidence	Study Method/Design	Question Applied	Oxford LOE 2011							
		Retrospective review of incidence and potential predictors of conduit disruption.	To assesses the frequency of RVOT conduit disruption during transcatheter pulmonary valve replacement (TPVR) and the effectiveness and safety of NuMED Covered Mounted CP Stents for its prevention or treatment.	1	2	3	4	5			
	Suitability	Relevant Data		Grading							
	Device	- Covered Mounted CP Stents (12 to 22mm)		D1	D2	D3					
	Application	- NuMED CP Stents (Covered) for prevention or treatment of RVOT conduit disruption during TPVR		A1	A2	A3					
	Patient	<ul style="list-style-type: none"> <li>- Population: Patients undergoing TPVR requiring treatment of RVOT conduit disruption (patients with pre-existing tears, patients developed tears after performing conduit dilation, and patients developed tears after transcatheter pulmonary valve implantation, or prophylactically placed in patients of perceived high risk related to degree of calcification and/or severity of homograft stenosis).</li> <li>- Sampling: 50 patients receiving 69 Covered CP Stents during TPVR/PPVI procedures (comparative cohort: 251 implants US Melody transcatheter pulmonary valve IDE Trial, planned for bare metal stenting of supported conduit)</li> </ul> <p style="margin: 0;"><u>Note:</u> overall incidence of conduit disruption requiring intervention was 6% in the study.</p> <ul style="list-style-type: none"> <li>- Mean Age: 21.4 (SD 3.7) years (5.5 to 56 years)</li> <li>- Sex: not reported</li> </ul>		P1	P2	P3					
Report	- High quality		R1	R2	R3						
Suitability Grade (Range 4-12)			4								
Data Contribution	Relevant Data		Grading								
Outcomes/Endpoints	<ul style="list-style-type: none"> <li>- Peak-to-peak RVOT gradient</li> <li>- Mean Doppler RVOT Gradient at 6 months</li> <li>- Valve competence with no or trivial pulmonary regurgitation</li> <li>- Safety</li> </ul>		Yes 1		No 2						
Follow-up	- 6 months		Yes 1	No 2							
Statistical analysis	- Not provided		Yes 1	No 2							
Clinical significance	- Covered Mounted CP Stent implantation can successfully treat RVOT conduit disruption without negative impact on the transcatheter pulmonary valve function		Yes 1		No 2						
Data Contribution Grade (Range 4-8)			6								
<b>Overall S&amp;P Appraisal, Disposition and Weighting</b>											
S&P Grade	LOE (3) + Suitability (4) +		Disposition and Weighting (select)		Accepted and Pivotal 9-12						



## NuMED Summary of Safety and Clinical Performance SSCP – Stents – CoA & RVOT

(Range 9-25)	Data Contribution (6) = 13		<b>Accepted but not Pivotal, 13-21</b> Excluded, 22-25
<b>Relevant S&amp;P Results</b>			
Safety data	<ul style="list-style-type: none"> <li>- Stent fracture: 1/50 patients with 2 CP Stents (Covered) developed multiple stent fracture 3 years following implantation. Patient was successfully treated with implantation of a stainless steel stent to support remaining portions of the CP Stent (Covered).</li> <li>- No Covered CP Stent-related acute complications were reported.</li> </ul>		
Performance data	<ul style="list-style-type: none"> <li>- Peak-to-peak RVOT gradient: Decreased from 45.5 ± 17.5 mm Hg to 10.6 ± 6.3 mm Hg</li> <li>- Mean Doppler RVOT Gradient at 6 months: 12.86 ± 5.0 mmHg compared to 20.0 ± 8.6 mmHg from the Melody TPV IDE trial.</li> <li>- Valve competence with no or trivial pulmonary regurgitation: At follow-up 94% in study group and 93% in comparator group.</li> <li>- Conduit tears: prevented or repaired in 49/50 patients.</li> </ul>		
Benefits/claims data	- N/A		
Strengths	<ul style="list-style-type: none"> <li>- CCPS implantation can successfully treat RVOT conduit disruption without negative impact on the TPV function.</li> <li>- This retrospective analysis suggests high RVOT conduit systolic pressure gradient is a risk factor for conduit tears during PPVI.</li> </ul>		
Weaknesses/ Potential bias	<ul style="list-style-type: none"> <li>- Retrospective analysis of prospectively collected data for other purposes and thus, suffers the biases of such investigations.</li> <li>- Sample size is small and in most cases the follow-up period is short. Long term results are unknown.</li> <li>- Conflict of interest reported:               <ul style="list-style-type: none"> <li>- Bishnoi RN: none</li> <li>- Jones T: research grant and consultant for Medtronic; research grant support from NuMED.</li> <li>- Kreutzer J: research grant support from Medtronic and St. Jude Medical; consultant for Medtronic, Inc.</li> <li>- Ringel RE: research grant support from Medtronic, Inc. and NuMED.</li> </ul> </li> </ul>		
<b>State of the Art</b>			
<b>Appraisal</b>			
Medical condition	Alternatives	Risk/benefit	Side-effects
Yes 1    No 2	Yes 1    No 2	Yes 1    No 2	Yes 1    No 2
Equivalence	Surrogate endpoints		
Yes 1    No 2	Yes 1    No 2		Yes 1    No 2
<b>Overall SOA Appraisal and Disposition</b>			
SOA Grade (Range 6-12)	9	Disposition (select)	<b>Accepted, &lt; 12</b> Excluded, 12
<b>Relevant SOA Results</b>			
SOA data	<ul style="list-style-type: none"> <li>- Surgical management of patients with CHD such as tetralogy of Fallot, pulmonary atresia, transposition of the great arteries, truncus arteriosus, and those undergoing Ross procedure for treatment of aortic valve disease, often includes implantation of a bioprosthetic valve or RVOT conduit.</li> <li>- The lifespan of bioprosthetic valves or RVOT conduits is limited by progressive obstruction and/or regurgitation due to</li> </ul>		



## NuMED Summary of Safety and Clinical Performance SSCP – Stents – CoA & RVOT

		<p>variety of factors including mechanical fatigue, immunologic reaction to the surgical implant, extrinsic conduit compression and somatic outgrowth on growing children.</p> <ul style="list-style-type: none"> <li>- Endovascular treatment using balloon dilatation and bare stent implantation has been shown to extend conduit lifespan and reduce a patient’s need for repeated open heart surgeries.</li> <li>- While resolving the problem of conduit obstruction, bare stent placement leads to creation or exacerbation of pulmonary regurgitation.</li> </ul>			
	Comments	- No further comment			
	<b>Safety &amp; Performance Appraisal</b>				
	Level of Evidence	Study Method/Design	Question Applied	Oxford LOE 2011	
		Prospective randomized controlled trial.	To evaluate outcomes of treatment with Covered versus Bare NuMED CP Stents.	1 2 3 4 5	
	Suitability	Relevant Data	Grading		
	Device	- NuMED CP Stent (Bare and Covered) - Stent was hand-crimped down onto BIB	D1	D2	D3
	Application	- Severe native CoA	A1	A2	A3
	Patient	- Patients with severe native CoA - Sampling: n=120 (60 CP Stents versus 60 CP Stents, Covered) - Mean age: 23.6±10.99 (range 12 to 58) years - Sex: 79 M; 41 F	P1	P2	P3
	Report	- High quality.	R1	R2	R3
	Suitability Grade (Range 4-12)			4	
	Data Contribution	Relevant Data	Grading		
	Outcomes/Endpoints	- Procedural success - Reduction in systolic blood pressure gradient - Reduction in mean diameter of coarctation segment - Adverse effects	Yes 1		No 2
	Follow-up	- 31.1 ± 19.2 months	Yes 1		No 2
	Statistical analysis	- A p-value <0.05 was considered significant.	Yes 1		No 2
	Clinical significance	- Implanting CP Stent (Bare) and CP Stent (Covered) have very high success rates with remarkable hemodynamic effects in severe native CoA patients, with no significant complication during the procedure and hospitalization. - Patients undergoing CP Stent (Covered) implantation experienced a non-significantly lower re-coarctation rate and a higher occurrence of pseudoaneurysm formation with respect to CP Stent (Bare) stenting during follow-up. - In both groups, blood pressure was significantly reduced after intervention.	Yes 1		No 2
5. Sohrabi et al. (2014)	Contribution				
	S&P	x			
	SOA	-			



## NuMED Summary of Safety and Clinical Performance SSCP – Stents – CoA & RVOT

	- These findings indicate that CoA stenting is a safe procedure.		
	Data Contribution Grade (Range 4-8)		4
<b>Overall S&amp;P Appraisal, Disposition and Weighting</b>			
S&P Grade (Range 9-25)	LOE (2) + Suitability (4) + Data Contribution (4) = 10	Disposition and Weighting (select)	<b>Accepted and Pivotal 9-12</b> Accepted but not Pivotal, 13-21 Excluded, 22-25
<b>Relevant S&amp;P Results</b>			
Safety data	<ul style="list-style-type: none"> <li>- Pseudoaneurysms: 0 (CP Stent, Bare) versus 2 (CP Stent, Covered)</li> <li>- Mortality: 1 (CP Stent, Bare) versus 0 (CP Stent, Covered)</li> </ul>		
Performance data	<ul style="list-style-type: none"> <li>- Successful placement: successful in all patients</li> <li>- Mean systolic blood pressure gradient reduction: from 54.61 (CP Stent, Bare) and 54.42 (CP Stent, Covered) to 3.47 and 3.36 mmHg respectively; no significant difference between the two types of stent, P&lt;0.001</li> <li>- Mean diameter of coarctation segment reduction: From 3.34 (CP Stent, Bare) and 3.30 (CP Stent, Covered) to 16.07 and 15.82 mm respectively; no significant difference between the two types of stent, P&lt;0.001</li> <li>- Recurring coarctation: 4 (CP Stent, Bare) versus 0 (CP Stent, Covered), non-significant</li> </ul>		
Benefits/claims data	<ul style="list-style-type: none"> <li>- Reduction in mean systolic blood pressure gradient</li> <li>- Reduction in diameter of coarctation segment</li> </ul>		
Strengths	<ul style="list-style-type: none"> <li>- The CP Stent was hand-crimped down onto a balloon-in-balloon catheter (NuMED), which allows a precise and safe stent delivery</li> </ul>		
Weaknesses/ Potential bias	<ul style="list-style-type: none"> <li>- Although the first randomized clinical trial in this respect, study was limited in some aspects. First, during follow-up, patients did not undergo 24-hour ambulatory blood pressure monitoring, which could have diagnosed the normotensive state more accurately. Second, evaluation of the blood pressure response during exercise testing could have been more valuable in defining the procedure outcome.</li> </ul>		
<b>Safety &amp; Performance Appraisal</b>			
6. Vanagt et al. (2014)	Level of Evidence	Study Method/Design	Question Applied
		Single-center retrospective study (CHD database of all CP Stent, Covered, during 2003-2012)	To evaluate possibilities and safety of CP Stent (Covered) in CHD.
		Oxford LOE 2011	
		1	2
		3	4
		5	
Contribution			
S&P	x		
SOA	-		
Suitability		Relevant Data	Grading
Device	<ul style="list-style-type: none"> <li>- Covered CP Stent</li> <li>- The stent was hand-crimped on BIB</li> </ul>		<b>D1</b> D2    D3
Application	<ul style="list-style-type: none"> <li>- CoA and RVOT pre-stenting for percutaneous revalvulation</li> </ul>		<b>A1</b> A2    A3
Patient	<ul style="list-style-type: none"> <li>- Patients with CoA and RVOT pre-stenting for percutaneous revalvulation. For the RVOT group, CP Stent (Covered) was chosen for delivery balloon protection after rupture of the pre-dilation balloon in 7/37 patients (19%) and 30 (81%) because tear, rupture, or fracture of the conduit</li> </ul>		<b>P1</b> P2    P3



## NuMED Summary of Safety and Clinical Performance SSCP – Stents – CoA & RVOT

	<p>was expected, or further stent expansion following somatic growth was anticipated.</p> <ul style="list-style-type: none"> <li>- Sampling: n= 51 (CoA group), n=37 (RVOT group)</li> <li>- Mean age:             <ul style="list-style-type: none"> <li>- CoA group: 19 (range from 8 to 69) years</li> <li>- RVOT group: 16 (range from 6 to 43) years</li> </ul> </li> <li>- Sex:             <ul style="list-style-type: none"> <li>- CoA group: 38M; 13F</li> <li>- RVOT group: 26M; 11F</li> </ul> </li> </ul>			
Report	- High quality.	<b>R1</b>	R2	R3
Suitability Grade (Range 4-12)		4		
Data Contribution	Relevant Data	Grading		
Outcomes/Endpoints	<ul style="list-style-type: none"> <li>- Increase in diameter at coarctation (CoA group)</li> <li>- Decrease in peak to peak gradient (CoA group)</li> <li>- Number of procedures for pre-stenting and pulmonary valve delivery (RVOT Group)</li> <li>- increase in graft diameter (RVOT Group)</li> <li>- Adverse effects</li> </ul>	<b>Yes 1</b>	No 2	
Follow-up	- Not specified.	Yes 1	<b>No 2</b>	
Statistical analysis	- Two-sided p<0.05 was considered significant.	<b>Yes 1</b>	No 2	
Clinical significance	- CP Stents (Covered) can safely be applied in CHD patients. The covering allows adequate sealing of existing or expected tears, thereby increasing the safety margin with more complete dilation.	<b>Yes 1</b>	No 2	
Data Contribution Grade (Range 4-8)		5		
<b>Overall S&amp;P Appraisal, Disposition and Weighting</b>				
S&P Grade (Range 9-25)	LOE (4) + Suitability (4) + Data Contribution (5) = 13	Disposition and Weighting (select)	Accepted and Pivotal 9-12 <b>Accepted but not Pivotal, 13-21</b> Excluded, 22-25	
<b>Relevant S&amp;P Results</b>				
Safety data	<ul style="list-style-type: none"> <li>- CoA Group:             <ul style="list-style-type: none"> <li>- No acute bleeding, aneurysm formation or life-threatening complications.</li> <li>- Mild procedure related-complications included groin hematoma (n = 3), transient nodal rhythm (n = 1, no wire present in left ventricle), and transient atrioventricular block with nodal escape rhythm (n = 1, while wire was present in left ventricle).</li> <li>- During follow-up: no stent fractures, nor stent recompression occurred, and none of the patients had limb ischemia or signs of vessel occlusion at the puncture site.</li> </ul> </li> <li>- RVOT group:             <ul style="list-style-type: none"> <li>- No procedure-related complications and no extravasation.</li> <li>- No embolization nor fracture of CP Stent (Covered) found on annual chest X-ray follow-up.</li> </ul> </li> </ul>			
Performance data	- Diameter at coarctation (CoA group):			



## NuMED Summary of Safety and Clinical Performance SSCP – Stents – CoA & RVOT

		<ul style="list-style-type: none"> <li>- Increased from 6 (0-15) to 14 (7-20) mm, P&lt;0.001.</li> <li>- Peak to peak gradient (CoA group):               <ul style="list-style-type: none"> <li>- Reduced from 23 (0-86) to 2 (0-25) mm Hg, P&lt;0.001.</li> </ul> </li> <li>- Number of procedures for pre-stenting and pulmonary valve delivery (RVOT Group):               <ul style="list-style-type: none"> <li>- 22/37 single procedure and 15/37 in a second procedure.</li> </ul> </li> <li>- Graft diameter (RVOT Group)               <ul style="list-style-type: none"> <li>- Increased from graft stenosis diameter of 13 (5-22) mm to 22 (16-26) mm at pre-revalvulation, P&lt;0.001.</li> </ul> </li> </ul>												
	Benefits/claims data	- Increase in luminal diameter in CoA patients.												
	Strengths	<ul style="list-style-type: none"> <li>- CP Stent (Covered) frame is made from 90% platinum and 10% iridium 0.013" wire, welded in a zig pattern with additional gold soldering. The strut thickness is slightly larger than most other stents, but makes the stent edges relatively atraumatic.</li> <li>- CP Stent (Covered) was hand-crimped on a balloon-in-balloon (BIB, Numed). Hand-inflation of the balloon was performed with a 10 ml syringe on the inner balloon and 20 ml syringe on the outer balloon, automatically limiting inflation pressures to 4–6 atmospheres.</li> </ul>												
	Weaknesses/ Potential bias	- In this retrospective study, there are no control groups with bare stents, the lack of which is inherently related to the fact that some of these procedures would have been impossible, or significantly less safe, if bare stents were used.												
	<b>Safety &amp; Performance Appraisal</b>													
	Level of Evidence	Study Method/Design	Question Applied	Oxford LOE 2011										
		Retrospective and observational study.	To investigate reduction in aortic wall rupture and dissection, as well as aneurysms by implanting covered stents.	<table border="1" style="display: inline-table; border-collapse: collapse;"> <tr> <td style="width: 20px;">1</td> <td style="width: 20px;">2</td> <td style="width: 20px;">3</td> <td style="width: 20px;">4</td> <td style="width: 20px;">5</td> </tr> <tr> <td></td> <td></td> <td></td> <td style="text-align: center;">4</td> <td></td> </tr> </table>	1	2	3	4	5				4	
1	2	3	4	5										
			4											
	Suitability	Relevant Data	Grading											
	Device	<ul style="list-style-type: none"> <li>- Covered CP Stent</li> <li>- BIB or Z-Med balloons (NuMED) – 9 of the 17 patients had BIB</li> <li>- Hand crimped</li> </ul>	<b>D1</b>	D2	D3									
	Application	- CoA and re-coarctation	<b>A1</b>	A2	A3									
	Patient	<ul style="list-style-type: none"> <li>- Patients treated for CoA and re-coarctation (2 adolescents and 15 adults treated between November 2005 and January 2012).</li> <li>- Sampling: n=17 (11 native CoA and 6 re-coarctation)</li> <li>- Mean age: 35 (range 14-65) years</li> <li>- Sex: 4 M; 13 F</li> </ul>	<b>P1</b>	P2	P3									
	Report	- High quality.	<b>R1</b>	R2	R3									
	Suitability Grade (Range 4-12)		4											
	Data Contribution	Relevant Data	Grading											
	Outcomes/Endpoints	- Reduction in blood pressure	<b>Yes 1</b>		No 2									

7. Alcibar et al. (2013)

Contribution	
S&P	x
SOA	-



## NuMED Summary of Safety and Clinical Performance SSCP – Stents – CoA & RVOT

	<ul style="list-style-type: none"> <li>- Reduction in lumen diameter</li> <li>- Reduction of hypertensive medications at follow-up</li> <li>- Adverse effects</li> </ul>		
Follow-up	- 2.5 years	<b>Yes 1</b>	No 2
Statistical analysis	- Significance was considered as P<0.05.	<b>Yes 1</b>	No 2
Clinical significance	- CP Stents (Covered) are effective in treating CoA and re-coarctation in adolescents and adults, are the treatment of choice in patients with complex anatomy, and must be available in the operating room as a rescue device when implanting a conventional stent.	<b>Yes 1</b>	No 2
Data Contribution Grade (Range 4-8)		4	

### Overall S&P Appraisal, Disposition and Weighting

S&P Grade (Range 9-25)	LOE (4) + Suitability (4) + Data Contribution (4) = 12	Disposition and Weighting (select)	<b>Accepted and Pivotal 9-12</b> Accepted but not Pivotal, 13-21 Excluded, 22-25
---------------------------	---	------------------------------------	--

### Relevant S&P Results

Safety data	<ul style="list-style-type: none"> <li>- One death: patient died two days post-op due to massive hematemesis as a result of the combination of an extreme increase in blood pressure and an existing aneurysm.</li> <li>- No local complications occurred, except one hematoma that resolved spontaneously.</li> <li>- No patient had any complication at the iliac-femoral level that required stenting.</li> </ul>
Performance data	<ul style="list-style-type: none"> <li>- Blood pressure gradient: Reduced from 40 to 2 mmHg (P&lt;0.001)</li> <li>- Lumen diameter: Increased from 4 to 15 mm (P&lt;0.001)</li> <li>- At follow-up (2.5 years):               <ul style="list-style-type: none"> <li>- All good initial outcome persisted without any signs of re-obstruction.</li> <li>- 13/17 patients underwent imaging study; no aneurysms, dissections, and/or obstructive processes were observed.</li> <li>- Medication for hypertension was reduced in 5 patients and in 2 patients could not be discontinued.</li> </ul> </li> </ul>
Benefits/claims data	<ul style="list-style-type: none"> <li>- Increased in luminal diameter</li> <li>- Decreased in antihypertensive medication use</li> </ul>
Strengths	- Having observed the case of aortic rupture, and with the aim of reducing these complications in patients who have had CoA and re-coarctation since their youth, the authors decided to electively implant a NuMED (Hopkinton, New York, United States) ePTFE CP Stent (Covered). This stent is mounted on a balloon catheter and protects the vascular wall when expanded.
Weaknesses/ Potential bias	- Retrospective and observational study with no control group of patients receiving conventional stents. Although all patients underwent clinical follow-up, this did not include an imaging study in all cases, and so authors cannot determine with certainty the incidence of potential aneurysms.





## NuMED Summary of Safety and Clinical Performance SSCP – Stents – CoA & RVOT

<p>8. Chang et al. (2012)</p> <table border="1" style="margin-top: 10px; width: 100%; border-collapse: collapse;"> <thead> <tr> <th colspan="2">Contribution</th> </tr> </thead> <tbody> <tr> <td style="width: 50%;">S&amp;P</td> <td style="text-align: center;">x</td> </tr> <tr> <td>SOA</td> <td style="text-align: center;">-</td> </tr> </tbody> </table>	Contribution		S&P	x	SOA	-	<b>Safety &amp; Performance Appraisal</b>				
	Contribution										
	S&P	x									
	SOA	-									
	Level of Evidence	Study Method/Design Single arm interventional study.	Question Applied To evaluate the use of CP Stent (Covered) as the primary modality in the treatment for native CoA.	Oxford LOE 2011							
				1	2	3	4	5			
	<b>Suitability</b>	<b>Relevant Data</b>			<b>Grading</b>						
	Device	<ul style="list-style-type: none"> <li>- CP Stent (Covered and Bare) 25 covered stents and 2 bare stents in 25 patients (one patient had 3 stents (2 bare, one covered) for native CoA with aortic arch hypoplasia (combination covered + stents = a new approach)</li> <li>- The covered CP stent was hand-crimped down onto BIB</li> </ul>			<b>D1</b>	D2	D3				
	Application	- Native CoA			<b>A1</b>	A2	A3				
	Patient	<ul style="list-style-type: none"> <li>- Patients with native CoA without previous treatment</li> <li>- Sampling: n=25</li> <li>- Mean age: 22.5 (range 14-46) years</li> <li>- Sex: 16 M; 9 F</li> </ul>			<b>P1</b>	P2	P3				
Report	- High quality.			<b>R1</b>	R2	R3					
Suitability Grade (Range 4-12)				4							
<b>Data Contribution</b>	<b>Relevant Data</b>			<b>Grading</b>							
Outcomes/Endpoints	<ul style="list-style-type: none"> <li>- Decrease in systolic gradient</li> <li>- Increase in stenotic segment diameter</li> </ul>			<b>Yes 1</b>		No 2					
Follow-up	- 32 (7-72) months			<b>Yes 1</b>		No 2					
Statistical analysis	- P<0.05 was set as statistically significant.			<b>Yes 1</b>		No 2					
Clinical significance	<ul style="list-style-type: none"> <li>- Implantation of CP Stent (Covered) as the primary modality is safe and effective in the treatment for native CoA in adolescents and adults.</li> <li>- Treatment modality of native CoA in adolescents and adults acquired excellent results, such as significant reduction in peak systolic gradient across CoA, successful relief of anatomic stenosis, and reduction of systemic hypertension.</li> <li>- Above all, no adverse events were encountered during the procedure or during the follow-up period of up to 72 months.</li> </ul>			<b>Yes 1</b>		No 2					
Data Contribution Grade (Range 4-8)				4							
<b>Overall S&amp;P Appraisal, Disposition and Weighting</b>											
S&P Grade (Range 9-25)	LOE (4) + Suitability (4) + Data Contribution (4) = 12	Disposition and Weighting (select)	<b>Accepted and Pivotal 9-12</b> Accepted but not Pivotal, 13-21 Excluded, 22-25								
<b>Relevant S&amp;P Results</b>											
Safety data	<ul style="list-style-type: none"> <li>- No acute complications were observed.</li> <li>- During a follow-up period of up to 72 months (median, 32 months and quartile range, 51 months), no adverse effects</li> </ul>										



## NuMED Summary of Safety and Clinical Performance SSCP – Stents – CoA & RVOT

		(e.g., dissection, aneurysm formation, stent migration, stent fracture) were encountered.												
	Performance data	<ul style="list-style-type: none"> <li>- In the patient with the implantation of three CP stents, the aneurysm formation related to the bare CP stent was not encountered, the left subclavian artery crossed by the bare CP stent presented patent without thrombosis, and the left arm ischemia was not detected.</li> <li>- Peak systolic gradient across the lesion:             <ul style="list-style-type: none"> <li>- Decreased from median 67.5 mmHg to median 2 mmHg (P&lt;0.0001)</li> </ul> </li> <li>- Stenotic segment diameter             <ul style="list-style-type: none"> <li>- Increased from median 5.0mm to median 17.9mm (P&lt;0.0001)</li> </ul> </li> <li>- At follow-up (up to 72 months):             <ul style="list-style-type: none"> <li>- Most patients (21/25) were normotensive; except from 4/25 patients still required antihypertensive medication during follow-up</li> </ul> </li> </ul>												
	Benefits/claims data	<ul style="list-style-type: none"> <li>- Reduced in peak systolic gradient.</li> <li>- Reduced in luminal diameter.</li> <li>- BIB offered precise and safe control over the stent implantation without any stent migration</li> </ul>												
	Strengths	<ul style="list-style-type: none"> <li>- Use of covered CP stents as the primary treatment modality may reduce the risk of significant complications related to stent implantation.</li> </ul>												
	Weaknesses/ Potential bias	<ul style="list-style-type: none"> <li>- Conflict of interest: not reported.</li> </ul>												
	<b>Safety &amp; Performance (for safety only)</b>													
	<b>Appraisal</b>													
	Level of Evidence	Study Method/Design	Question Applied	Oxford LOE 2011										
		Single arm interventional study.	To present author's institutional experience of endovascular CP Stent implantation in children and adults with native and recurrent CoA.	1	2	3	4	5						
	Suitability	Relevant Data	Grading											
	Device	<ul style="list-style-type: none"> <li>- CP Stent (16 Covered or 31 Bare) – n=47</li> <li>- BIB (n=29) (not subject device) or single balloon catheter (n=18) (not subject device), Z-med (not subject device); manually crimped</li> </ul>	D1	D2	D3									
	Application	<ul style="list-style-type: none"> <li>- Patients with native or recurrent CoA</li> </ul>	A1	A2	A3									
	Patient	<ul style="list-style-type: none"> <li>- Patients with native CoA (Group 1); recurrent CoA and/or aneurysm developed after either surgery or balloon angioplasty (Group 2)</li> <li>- Sampling: n=45 (47 CP Stents, Covered or Bare)</li> <li>- Median age: 11 (range: 5-33) years</li> <li>- Sex: 34M; 11F</li> </ul>	P1	P2	P3									
	Report	<ul style="list-style-type: none"> <li>- High quality.</li> </ul>	R1	R2	R3									
	Suitability Grade (Range 4-12)			5										
9. Erdem et al. (2011)	<table border="1" style="width: 100%; border-collapse: collapse;"> <tr> <th colspan="2" style="text-align: left;">Contribution</th> </tr> <tr> <td style="width: 50%;">S&amp;P</td> <td style="width: 50%; text-align: center;">X (S only)</td> </tr> <tr> <td>SOA</td> <td style="text-align: center;">-</td> </tr> </table>								Contribution		S&P	X (S only)	SOA	-
Contribution														
S&P	X (S only)													
SOA	-													



## NuMED Summary of Safety and Clinical Performance SSCP – Stents – CoA & RVOT

Data Contribution	Relevant Data	Grading	
Outcomes/Endpoints	<ul style="list-style-type: none"> <li>- Decrease in invasive and echocardiographic gradients</li> <li>- Increase in lesion diameter</li> <li>- Adverse effects</li> </ul>	<b>Yes 1</b>	No 2
Follow-up	- 12.1±7.1 months; median 11 month (range 2-29)	<b>Yes 1</b>	No 2
Statistical analysis	- A p value <0.05 was considered statistically significant.	<b>Yes 1</b>	No 2
Clinical significance	<ul style="list-style-type: none"> <li>- Early and short- term follow-up results indicate that stent implantation is safe and very effective in reducing coarctation gradient and increasing lesion diameter both in native and recurrent CoA.</li> <li>- Some serious complications do occur and hypertension remains in some patients.</li> <li>- Aortic disruption and stent displacement are potentially catastrophic complications of stenting but implanting a second covered stent can seal the ruptured wall and parking in a safe area or replacement of displaced stent carried by half-inflated balloon could solve the problem.</li> </ul>	<b>Yes 1</b>	No 2
Data Contribution Grade (Range 4-8)		4	

### Overall S&P Appraisal, Disposition and Weighting

S&P Grade (Range 9-25)	LOE (4) + Suitability (5) + Data Contribution (4) = 13	Disposition and Weighting (select)	Accepted and Pivotal 9-12 <b>Accepted but not Pivotal, 13-21</b> Excluded, 22-25
---------------------------	---	------------------------------------	--

### Relevant S&P Results

Safety data	<ul style="list-style-type: none"> <li>- No procedure related death.</li> <li>- Two immediate complications relating to stenting:               <ul style="list-style-type: none"> <li>- One an acute wall rupture, successfully managed immediately in the same session with implantation of a second covered stent</li> <li>- One stent was displaced before it was completely opened. It was carried with support of partially inflated balloon and long sheath, and repositioned into the correct place.</li> </ul> </li> <li>- No femoral arterial complications</li> <li>- No difficulty in catheter manipulation.</li> <li>- None of the patients required intensive care following the procedure, and all were discharged home the following day except the patient with aortic rupture and after stenting with covered stent this patient was followed two days in intensive care unit.</li> </ul>
Benefits/claims data	- Increase in luminal/lesion diameter.
Strengths	<ul style="list-style-type: none"> <li>- CP stent is the one of the most commonly used stent in pediatric cardiology</li> <li>- This stent has excellent radial strength even at larger diameters and also has brilliant visibility on fluoroscopy.</li> </ul>
Weaknesses/ Potential bias	<ul style="list-style-type: none"> <li>- Some limitations have to be noted about this study:               <ul style="list-style-type: none"> <li>- Firstly, there is a need a greater number of patients have undergone stent implantation and their long-term results.</li> <li>- Secondly, population included both children and adult.</li> <li>- Thirdly, this was a single-center report and patients were not compared with surgery or balloon angioplasty</li> </ul> </li> </ul>



## NuMED Summary of Safety and Clinical Performance SSCP – Stents – CoA & RVOT

		<p>alone.</p> <ul style="list-style-type: none"> <li>- Fourthly, 24-hour ambulatory blood pressure monitoring before stenting was not performed in any patients.</li> <li>- Finally, radiologic imaging for aneurysm was done in limited number of patients after procedure.</li> <li>- Conflict of interest: None declared.</li> </ul>																																																																																
<p>10. Butera et al. (2011)</p> <table border="1" style="margin-top: 10px; width: 100%; border-collapse: collapse;"> <tr> <th colspan="2" style="text-align: left;">Contribution</th> </tr> <tr> <td style="width: 50px;">S&amp;P</td> <td style="text-align: center;">x</td> </tr> <tr> <td>SOA</td> <td style="text-align: center;">-</td> </tr> </table>	Contribution		S&P	x	SOA	-	<p><b>Safety &amp; Performance Appraisal</b></p> <table border="1" style="width: 100%; border-collapse: collapse;"> <tr> <th style="width: 25%;">Level of Evidence</th> <th style="width: 30%;">Study Method/Design</th> <th style="width: 30%;">Question Applied</th> <th colspan="5" style="text-align: center;">Oxford LOE 2011</th> </tr> <tr> <td></td> <td>Prospective single arm interventional study.</td> <td>To evaluate the management of aneurysms associated with CoA by covered stent deployment.</td> <td style="text-align: center;">1</td> <td style="text-align: center;">2</td> <td style="text-align: center;">3</td> <td style="text-align: center;">4</td> <td style="text-align: center;">5</td> </tr> </table> <table border="1" style="width: 100%; border-collapse: collapse;"> <tr> <th style="width: 25%;">Suitability</th> <th style="width: 50%;">Relevant Data</th> <th colspan="3" style="text-align: center;">Grading</th> </tr> <tr> <td>Device</td> <td> <ul style="list-style-type: none"> <li>- Covered CP Stent</li> <li>- BIB or Crystal balloon (not subject device) (manually crimped)</li> </ul> </td> <td style="text-align: center;">D1</td> <td style="text-align: center;"><b>D2</b></td> <td style="text-align: center;">D3</td> </tr> <tr> <td>Application</td> <td> <ul style="list-style-type: none"> <li>- Patients with native CoA associated with aortic wall aneurysm</li> </ul> </td> <td style="text-align: center;"><b>A1</b></td> <td style="text-align: center;">A2</td> <td style="text-align: center;">A3</td> </tr> <tr> <td>Patient</td> <td> <ul style="list-style-type: none"> <li>- Patients with CoA associated with aortic wall aneurysm</li> <li>- Sampling: n=11 (3 native CoA, 3 with previous surgical repair, 3 with previous balloon angioplasty, and 2 with previous bare stent implantation)</li> <li>- Median age: 13 (range: 6-66) years</li> <li>- Sex: Not reported</li> </ul> </td> <td style="text-align: center;">P1</td> <td style="text-align: center;"><b>P2</b></td> <td style="text-align: center;">P3</td> </tr> <tr> <td>Report</td> <td> <ul style="list-style-type: none"> <li>- High quality.</li> </ul> </td> <td style="text-align: center;"><b>R1</b></td> <td style="text-align: center;">R2</td> <td style="text-align: center;">R3</td> </tr> <tr> <td colspan="2" style="text-align: right;">Suitability Grade (Range 4-12)</td> <td colspan="3" style="text-align: center;">6</td> </tr> </table> <table border="1" style="width: 100%; border-collapse: collapse;"> <tr> <th style="width: 25%;">Data Contribution</th> <th style="width: 50%;">Relevant Data</th> <th colspan="2" style="text-align: center;">Grading</th> </tr> <tr> <td>Outcomes/Endpoints</td> <td> <ul style="list-style-type: none"> <li>- Systolic pressure gradient reduction</li> <li>- Increase in aortic diameter</li> <li>- Adverse effects</li> </ul> </td> <td style="text-align: center;"><b>Yes 1</b></td> <td style="text-align: center;">No 2</td> </tr> <tr> <td>Follow-up</td> <td> <ul style="list-style-type: none"> <li>- Median follow-up 50 (16-61) months</li> </ul> </td> <td style="text-align: center;"><b>Yes 1</b></td> <td style="text-align: center;">No 2</td> </tr> <tr> <td>Statistical analysis</td> <td> <ul style="list-style-type: none"> <li>- P-value less than 0.05 was considered to be statistically significant</li> </ul> </td> <td style="text-align: center;"><b>Yes 1</b></td> <td style="text-align: center;">No 2</td> </tr> <tr> <td>Clinical significance</td> <td> <ul style="list-style-type: none"> <li>- CP Stent (Covered) are a safe and effective treatment with low risk of complication for the treatment of CoA associated with aortic wall aneurysm.</li> <li>- CP Stents (Covered, e-PTFE) may be considered the treatment of choice for native CoA associated with aortic wall aneurysm.</li> </ul> </td> <td style="text-align: center;"><b>Yes 1</b></td> <td style="text-align: center;">No 2</td> </tr> <tr> <td colspan="2" style="text-align: right;">Data Contribution Grade (Range 4-8)</td> <td colspan="2" style="text-align: center;">4</td> </tr> </table> <p><b>Overall S&amp;P Appraisal, Disposition and Weighting</b></p> <table border="1" style="width: 100%; border-collapse: collapse;"> <tr> <td style="width: 25%;">S&amp;P Grade (Range 9-25)</td> <td style="width: 25%;">LOE (3) + Suitability (6) + Data Contribution (4) = 13</td> <td style="width: 25%;">Disposition and Weighting (select)</td> <td style="width: 25%;">Accepted and Pivotal 9-12 <b>Accepted but not Pivotal, 13-21</b> Excluded, 22-25</td> </tr> </table>		Level of Evidence	Study Method/Design	Question Applied	Oxford LOE 2011						Prospective single arm interventional study.	To evaluate the management of aneurysms associated with CoA by covered stent deployment.	1	2	3	4	5	Suitability	Relevant Data	Grading			Device	<ul style="list-style-type: none"> <li>- Covered CP Stent</li> <li>- BIB or Crystal balloon (not subject device) (manually crimped)</li> </ul>	D1	<b>D2</b>	D3	Application	<ul style="list-style-type: none"> <li>- Patients with native CoA associated with aortic wall aneurysm</li> </ul>	<b>A1</b>	A2	A3	Patient	<ul style="list-style-type: none"> <li>- Patients with CoA associated with aortic wall aneurysm</li> <li>- Sampling: n=11 (3 native CoA, 3 with previous surgical repair, 3 with previous balloon angioplasty, and 2 with previous bare stent implantation)</li> <li>- Median age: 13 (range: 6-66) years</li> <li>- Sex: Not reported</li> </ul>	P1	<b>P2</b>	P3	Report	<ul style="list-style-type: none"> <li>- High quality.</li> </ul>	<b>R1</b>	R2	R3	Suitability Grade (Range 4-12)		6			Data Contribution	Relevant Data	Grading		Outcomes/Endpoints	<ul style="list-style-type: none"> <li>- Systolic pressure gradient reduction</li> <li>- Increase in aortic diameter</li> <li>- Adverse effects</li> </ul>	<b>Yes 1</b>	No 2	Follow-up	<ul style="list-style-type: none"> <li>- Median follow-up 50 (16-61) months</li> </ul>	<b>Yes 1</b>	No 2	Statistical analysis	<ul style="list-style-type: none"> <li>- P-value less than 0.05 was considered to be statistically significant</li> </ul>	<b>Yes 1</b>	No 2	Clinical significance	<ul style="list-style-type: none"> <li>- CP Stent (Covered) are a safe and effective treatment with low risk of complication for the treatment of CoA associated with aortic wall aneurysm.</li> <li>- CP Stents (Covered, e-PTFE) may be considered the treatment of choice for native CoA associated with aortic wall aneurysm.</li> </ul>	<b>Yes 1</b>	No 2	Data Contribution Grade (Range 4-8)		4		S&P Grade (Range 9-25)	LOE (3) + Suitability (6) + Data Contribution (4) = 13	Disposition and Weighting (select)	Accepted and Pivotal 9-12 <b>Accepted but not Pivotal, 13-21</b> Excluded, 22-25
	Contribution																																																																																	
	S&P	x																																																																																
	SOA	-																																																																																
	Level of Evidence	Study Method/Design	Question Applied	Oxford LOE 2011																																																																														
		Prospective single arm interventional study.	To evaluate the management of aneurysms associated with CoA by covered stent deployment.	1	2	3	4	5																																																																										
	Suitability	Relevant Data	Grading																																																																															
	Device	<ul style="list-style-type: none"> <li>- Covered CP Stent</li> <li>- BIB or Crystal balloon (not subject device) (manually crimped)</li> </ul>	D1	<b>D2</b>	D3																																																																													
	Application	<ul style="list-style-type: none"> <li>- Patients with native CoA associated with aortic wall aneurysm</li> </ul>	<b>A1</b>	A2	A3																																																																													
	Patient	<ul style="list-style-type: none"> <li>- Patients with CoA associated with aortic wall aneurysm</li> <li>- Sampling: n=11 (3 native CoA, 3 with previous surgical repair, 3 with previous balloon angioplasty, and 2 with previous bare stent implantation)</li> <li>- Median age: 13 (range: 6-66) years</li> <li>- Sex: Not reported</li> </ul>	P1	<b>P2</b>	P3																																																																													
Report	<ul style="list-style-type: none"> <li>- High quality.</li> </ul>	<b>R1</b>	R2	R3																																																																														
Suitability Grade (Range 4-12)		6																																																																																
Data Contribution	Relevant Data	Grading																																																																																
Outcomes/Endpoints	<ul style="list-style-type: none"> <li>- Systolic pressure gradient reduction</li> <li>- Increase in aortic diameter</li> <li>- Adverse effects</li> </ul>	<b>Yes 1</b>	No 2																																																																															
Follow-up	<ul style="list-style-type: none"> <li>- Median follow-up 50 (16-61) months</li> </ul>	<b>Yes 1</b>	No 2																																																																															
Statistical analysis	<ul style="list-style-type: none"> <li>- P-value less than 0.05 was considered to be statistically significant</li> </ul>	<b>Yes 1</b>	No 2																																																																															
Clinical significance	<ul style="list-style-type: none"> <li>- CP Stent (Covered) are a safe and effective treatment with low risk of complication for the treatment of CoA associated with aortic wall aneurysm.</li> <li>- CP Stents (Covered, e-PTFE) may be considered the treatment of choice for native CoA associated with aortic wall aneurysm.</li> </ul>	<b>Yes 1</b>	No 2																																																																															
Data Contribution Grade (Range 4-8)		4																																																																																
S&P Grade (Range 9-25)	LOE (3) + Suitability (6) + Data Contribution (4) = 13	Disposition and Weighting (select)	Accepted and Pivotal 9-12 <b>Accepted but not Pivotal, 13-21</b> Excluded, 22-25																																																																															



## NuMED Summary of Safety and Clinical Performance SSCP – Stents – CoA & RVOT

	<p><b>Relevant S&amp;P Results</b></p> <table border="1" style="width: 100%; border-collapse: collapse;"> <tr> <td style="width: 20%;">Safety data</td> <td>- No early complications observed.</td> </tr> <tr> <td>Performance data</td> <td> <ul style="list-style-type: none"> <li>- Successful device deployment: Achieved in all patients.</li> <li>- Successful relief of stenoses and complete sealing of all aneurysms.</li> <li>- Systolic pressure gradient reduction: From median 30 (25-50) to 5 (0-20) mmHg, P&lt;0.01</li> <li>- Increase of aortic diameter: From median 6 (0.5 – 11) to 12 (10-22) mm, P&lt;0.001</li> <li>- Re-dilatation required at follow-up: four patients developed systemic hypertension (one intrastent restenosis secondary to significant endothelial growth, three showed restenosis secondary to somatic growth). Re-dilatation with a larger balloon was performed without complication in all cases.</li> </ul> </td> </tr> <tr> <td>Benefits/claims data</td> <td> <ul style="list-style-type: none"> <li>- Increase in luminal diameter</li> <li>- Reduce systolic pressure gradient</li> <li>- Reduce/prevent aortic wall injury (patients associated with aortic wall aneurysm)</li> </ul> </td> </tr> <tr> <td>Strengths</td> <td>- Covered CP stents are manufactured with an alloy of 90% platinum and 10% iridium. Theoretically, this combination is more malleable and with good radial strength, which is enhanced by being designed in a “zig” pattern. The CP stent has rounded edges, decreasing the risk of balloon rupture or injury to the vessel wall and, in addition, the platinum component makes it more radio-opaque. Furthermore, the e-PTFE protects the stenotic and diseased segment.</td> </tr> <tr> <td>Weaknesses/ Potential bias</td> <td>- No conflict of interest reported.</td> </tr> </table>					Safety data	- No early complications observed.	Performance data	<ul style="list-style-type: none"> <li>- Successful device deployment: Achieved in all patients.</li> <li>- Successful relief of stenoses and complete sealing of all aneurysms.</li> <li>- Systolic pressure gradient reduction: From median 30 (25-50) to 5 (0-20) mmHg, P&lt;0.01</li> <li>- Increase of aortic diameter: From median 6 (0.5 – 11) to 12 (10-22) mm, P&lt;0.001</li> <li>- Re-dilatation required at follow-up: four patients developed systemic hypertension (one intrastent restenosis secondary to significant endothelial growth, three showed restenosis secondary to somatic growth). Re-dilatation with a larger balloon was performed without complication in all cases.</li> </ul>	Benefits/claims data	<ul style="list-style-type: none"> <li>- Increase in luminal diameter</li> <li>- Reduce systolic pressure gradient</li> <li>- Reduce/prevent aortic wall injury (patients associated with aortic wall aneurysm)</li> </ul>	Strengths	- Covered CP stents are manufactured with an alloy of 90% platinum and 10% iridium. Theoretically, this combination is more malleable and with good radial strength, which is enhanced by being designed in a “zig” pattern. The CP stent has rounded edges, decreasing the risk of balloon rupture or injury to the vessel wall and, in addition, the platinum component makes it more radio-opaque. Furthermore, the e-PTFE protects the stenotic and diseased segment.	Weaknesses/ Potential bias	- No conflict of interest reported.																																																		
Safety data	- No early complications observed.																																																																
Performance data	<ul style="list-style-type: none"> <li>- Successful device deployment: Achieved in all patients.</li> <li>- Successful relief of stenoses and complete sealing of all aneurysms.</li> <li>- Systolic pressure gradient reduction: From median 30 (25-50) to 5 (0-20) mmHg, P&lt;0.01</li> <li>- Increase of aortic diameter: From median 6 (0.5 – 11) to 12 (10-22) mm, P&lt;0.001</li> <li>- Re-dilatation required at follow-up: four patients developed systemic hypertension (one intrastent restenosis secondary to significant endothelial growth, three showed restenosis secondary to somatic growth). Re-dilatation with a larger balloon was performed without complication in all cases.</li> </ul>																																																																
Benefits/claims data	<ul style="list-style-type: none"> <li>- Increase in luminal diameter</li> <li>- Reduce systolic pressure gradient</li> <li>- Reduce/prevent aortic wall injury (patients associated with aortic wall aneurysm)</li> </ul>																																																																
Strengths	- Covered CP stents are manufactured with an alloy of 90% platinum and 10% iridium. Theoretically, this combination is more malleable and with good radial strength, which is enhanced by being designed in a “zig” pattern. The CP stent has rounded edges, decreasing the risk of balloon rupture or injury to the vessel wall and, in addition, the platinum component makes it more radio-opaque. Furthermore, the e-PTFE protects the stenotic and diseased segment.																																																																
Weaknesses/ Potential bias	- No conflict of interest reported.																																																																
<p>11. Tanous et al. (2010)</p> <table border="1" style="width: 100%; border-collapse: collapse; margin-top: 10px;"> <tr> <th colspan="2">Contribution</th> </tr> <tr> <td>S&amp;P</td> <td style="text-align: center;">x</td> </tr> <tr> <td>SOA</td> <td style="text-align: center;">-</td> </tr> </table>	Contribution		S&P	x	SOA	-	<p><b>Safety &amp; Performance Appraisal</b></p> <table border="1" style="width: 100%; border-collapse: collapse;"> <tr> <th style="width: 20%;">Level of Evidence</th> <th style="width: 30%;">Study Method/Design</th> <th style="width: 30%;">Question Applied</th> <th colspan="5">Oxford LOE 2011</th> </tr> <tr> <td></td> <td>Single arm interventional study.</td> <td>To determine the safety and efficacy of the CP Stent (Covered) in adults with CoA.</td> <td style="text-align: center;">1</td> <td style="text-align: center;">2</td> <td style="text-align: center;">3</td> <td style="text-align: center;">4</td> <td style="text-align: center;">5</td> </tr> </table> <table border="1" style="width: 100%; border-collapse: collapse;"> <thead> <tr> <th style="width: 20%;">Suitability</th> <th style="width: 50%;">Relevant Data</th> <th colspan="3">Grading</th> </tr> </thead> <tbody> <tr> <td>Device</td> <td>- Covered CP Stent hand-crimped on Z-Med II (NuMED) or Covered Mounted CP Stent</td> <td style="text-align: center;">D1</td> <td style="text-align: center;">D2</td> <td style="text-align: center;">D3</td> </tr> <tr> <td>Application</td> <td>- Native CoA (n=14) and previous treatment (n=8)</td> <td style="text-align: center;">A1</td> <td style="text-align: center;">A2</td> <td style="text-align: center;">A3</td> </tr> <tr> <td>Patient</td> <td> <ul style="list-style-type: none"> <li>- Patients with native CoA and CoA with previous treatment</li> <li>- Sampling: 14 native CoA; 8 CoA (with previous treatment)</li> <li>- Mean age: 39±14 (range 19 to 67) years</li> <li>- Sex: 11 M; 11 F</li> </ul> </td> <td style="text-align: center;">P1</td> <td style="text-align: center;">P2</td> <td style="text-align: center;">P3</td> </tr> <tr> <td>Report</td> <td>- High quality.</td> <td style="text-align: center;">R1</td> <td style="text-align: center;">R2</td> <td style="text-align: center;">R3</td> </tr> <tr> <td colspan="2" style="text-align: right;">Suitability Grade (Range 4-12)</td> <td colspan="3" style="text-align: center;">5</td> </tr> </tbody> </table> <table border="1" style="width: 100%; border-collapse: collapse;"> <thead> <tr> <th style="width: 20%;">Data Contribution</th> <th style="width: 50%;">Relevant Data</th> <th colspan="2">Grading</th> </tr> </thead> <tbody> <tr> <td>Outcomes/Endpoints</td> <td> <ul style="list-style-type: none"> <li>- Reduction in peak systolic gradient across coarctation site</li> <li>- Adverse effects</li> </ul> </td> <td style="text-align: center;">Yes 1</td> <td style="text-align: center;">No 2</td> </tr> </tbody> </table>					Level of Evidence	Study Method/Design	Question Applied	Oxford LOE 2011						Single arm interventional study.	To determine the safety and efficacy of the CP Stent (Covered) in adults with CoA.	1	2	3	4	5	Suitability	Relevant Data	Grading			Device	- Covered CP Stent hand-crimped on Z-Med II (NuMED) or Covered Mounted CP Stent	D1	D2	D3	Application	- Native CoA (n=14) and previous treatment (n=8)	A1	A2	A3	Patient	<ul style="list-style-type: none"> <li>- Patients with native CoA and CoA with previous treatment</li> <li>- Sampling: 14 native CoA; 8 CoA (with previous treatment)</li> <li>- Mean age: 39±14 (range 19 to 67) years</li> <li>- Sex: 11 M; 11 F</li> </ul>	P1	P2	P3	Report	- High quality.	R1	R2	R3	Suitability Grade (Range 4-12)		5			Data Contribution	Relevant Data	Grading		Outcomes/Endpoints	<ul style="list-style-type: none"> <li>- Reduction in peak systolic gradient across coarctation site</li> <li>- Adverse effects</li> </ul>	Yes 1	No 2
Contribution																																																																	
S&P	x																																																																
SOA	-																																																																
Level of Evidence	Study Method/Design	Question Applied	Oxford LOE 2011																																																														
	Single arm interventional study.	To determine the safety and efficacy of the CP Stent (Covered) in adults with CoA.	1	2	3	4	5																																																										
Suitability	Relevant Data	Grading																																																															
Device	- Covered CP Stent hand-crimped on Z-Med II (NuMED) or Covered Mounted CP Stent	D1	D2	D3																																																													
Application	- Native CoA (n=14) and previous treatment (n=8)	A1	A2	A3																																																													
Patient	<ul style="list-style-type: none"> <li>- Patients with native CoA and CoA with previous treatment</li> <li>- Sampling: 14 native CoA; 8 CoA (with previous treatment)</li> <li>- Mean age: 39±14 (range 19 to 67) years</li> <li>- Sex: 11 M; 11 F</li> </ul>	P1	P2	P3																																																													
Report	- High quality.	R1	R2	R3																																																													
Suitability Grade (Range 4-12)		5																																																															
Data Contribution	Relevant Data	Grading																																																															
Outcomes/Endpoints	<ul style="list-style-type: none"> <li>- Reduction in peak systolic gradient across coarctation site</li> <li>- Adverse effects</li> </ul>	Yes 1	No 2																																																														



## NuMED Summary of Safety and Clinical Performance SSCP – Stents – CoA & RVOT

	<table border="1" style="width: 100%; border-collapse: collapse;"> <tr> <td style="width: 30%;">Follow-up</td> <td style="width: 40%;">- 12 (9-15) months</td> <td style="width: 10%; text-align: center;"><b>Yes 1</b></td> <td style="width: 10%; text-align: center;">No 2</td> </tr> <tr> <td>Statistical analysis</td> <td>- A P-value &lt;0.05 was considered significant.</td> <td style="text-align: center;"><b>Yes 1</b></td> <td style="text-align: center;">No 2</td> </tr> <tr> <td>Clinical significance</td> <td>- Covered stents are safe, durable, and efficacious in the management of CoA.</td> <td style="text-align: center;"><b>Yes 1</b></td> <td style="text-align: center;">No 2</td> </tr> <tr> <td colspan="2" style="text-align: right;">Data Contribution Grade (Range 4-8)</td> <td colspan="2" style="text-align: center;">4</td> </tr> </table> <p><b>Overall S&amp;P Appraisal, Disposition and Weighting</b></p> <table border="1" style="width: 100%; border-collapse: collapse;"> <tr> <td style="width: 25%;">S&amp;P Grade (Range 9-25)</td> <td style="width: 30%;">LOE (4) + Suitability (5) + Data Contribution (4) = 13</td> <td style="width: 20%;">Disposition and Weighting (select)</td> <td style="width: 25%;">Accepted and Pivotal 9-12 <b>Accepted but not Pivotal, 13-21</b> Excluded, 22-25</td> </tr> </table> <p><b>Relevant S&amp;P Results</b></p> <table border="1" style="width: 100%; border-collapse: collapse;"> <tr> <td style="width: 25%;">Safety data</td> <td>- One pseudoaneurysm. Patient was treated successfully. Note: this problem may have been caused because the stent was hand crimped. When pre-mounted stents were used the problem did not reoccur.</td> </tr> <tr> <td>Performance data</td> <td>- Reduction in peak systolic gradient across coarctation site: From average 29 ± 17 to 3 ± 5 mmHg immediately post intervention and 6 ± 9 mmHg at follow up, P&lt;0.001</td> </tr> <tr> <td>Benefits/claims data</td> <td>- Reduction in peak systolic gradient</td> </tr> <tr> <td>Strengths</td> <td>- N/A</td> </tr> <tr> <td>Weaknesses/ Potential bias</td> <td>- This review is limited by the small sample size and lack of a randomized comparison group. - This study was not intended to demonstrate the efficacy of percutaneous therapy, or the superiority of covered stents, but rather to document a single-center experience as an alternative and safe treatment option in a broad spectrum of patients with aortic coarctation.</td> </tr> </table>	Follow-up	- 12 (9-15) months	<b>Yes 1</b>	No 2	Statistical analysis	- A P-value <0.05 was considered significant.	<b>Yes 1</b>	No 2	Clinical significance	- Covered stents are safe, durable, and efficacious in the management of CoA.	<b>Yes 1</b>	No 2	Data Contribution Grade (Range 4-8)		4		S&P Grade (Range 9-25)	LOE (4) + Suitability (5) + Data Contribution (4) = 13	Disposition and Weighting (select)	Accepted and Pivotal 9-12 <b>Accepted but not Pivotal, 13-21</b> Excluded, 22-25	Safety data	- One pseudoaneurysm. Patient was treated successfully. Note: this problem may have been caused because the stent was hand crimped. When pre-mounted stents were used the problem did not reoccur.	Performance data	- Reduction in peak systolic gradient across coarctation site: From average 29 ± 17 to 3 ± 5 mmHg immediately post intervention and 6 ± 9 mmHg at follow up, P<0.001	Benefits/claims data	- Reduction in peak systolic gradient	Strengths	- N/A	Weaknesses/ Potential bias	- This review is limited by the small sample size and lack of a randomized comparison group. - This study was not intended to demonstrate the efficacy of percutaneous therapy, or the superiority of covered stents, but rather to document a single-center experience as an alternative and safe treatment option in a broad spectrum of patients with aortic coarctation.																								
Follow-up	- 12 (9-15) months	<b>Yes 1</b>	No 2																																																				
Statistical analysis	- A P-value <0.05 was considered significant.	<b>Yes 1</b>	No 2																																																				
Clinical significance	- Covered stents are safe, durable, and efficacious in the management of CoA.	<b>Yes 1</b>	No 2																																																				
Data Contribution Grade (Range 4-8)		4																																																					
S&P Grade (Range 9-25)	LOE (4) + Suitability (5) + Data Contribution (4) = 13	Disposition and Weighting (select)	Accepted and Pivotal 9-12 <b>Accepted but not Pivotal, 13-21</b> Excluded, 22-25																																																				
Safety data	- One pseudoaneurysm. Patient was treated successfully. Note: this problem may have been caused because the stent was hand crimped. When pre-mounted stents were used the problem did not reoccur.																																																						
Performance data	- Reduction in peak systolic gradient across coarctation site: From average 29 ± 17 to 3 ± 5 mmHg immediately post intervention and 6 ± 9 mmHg at follow up, P<0.001																																																						
Benefits/claims data	- Reduction in peak systolic gradient																																																						
Strengths	- N/A																																																						
Weaknesses/ Potential bias	- This review is limited by the small sample size and lack of a randomized comparison group. - This study was not intended to demonstrate the efficacy of percutaneous therapy, or the superiority of covered stents, but rather to document a single-center experience as an alternative and safe treatment option in a broad spectrum of patients with aortic coarctation.																																																						
<p>12. Moltzer et al. (2010)</p> <table border="1" style="width: 100%; border-collapse: collapse; margin-top: 10px;"> <tr> <th colspan="2">Contribution</th> </tr> <tr> <td style="width: 50%;">S&amp;P</td> <td style="width: 50%;">X (S only)</td> </tr> <tr> <td>SOA</td> <td style="text-align: center;">-</td> </tr> </table>	Contribution		S&P	X (S only)	SOA	-	<p><b>Safety &amp; Performance (for safety only)</b></p> <p><b>Appraisal</b></p> <table border="1" style="width: 100%; border-collapse: collapse;"> <tr> <th style="width: 20%;">Level of Evidence</th> <th style="width: 30%;">Study Method/Design</th> <th style="width: 30%;">Question Applied</th> <th colspan="5" style="text-align: center;">Oxford LOE 2011</th> </tr> <tr> <td></td> <td>Prospective observational study.</td> <td>To evaluate the intermediate-term outcome of stent implantation for CoA in adults.</td> <td style="text-align: center;">1</td> <td style="text-align: center;">2</td> <td style="text-align: center;"><b>3</b></td> <td style="text-align: center;">4</td> <td style="text-align: center;">5</td> </tr> </table> <table border="1" style="width: 100%; border-collapse: collapse; margin-top: 10px;"> <tr> <th style="width: 25%;">Suitability</th> <th style="width: 50%;">Relevant Data</th> <th colspan="3" style="text-align: center;">Grading</th> </tr> <tr> <td>Device</td> <td>- CP Stent (Bare and Covered) – 6 of the 24 patients had covered stents - BIB (manually crimped)</td> <td style="text-align: center;"><b>D1</b></td> <td style="text-align: center;">D2</td> <td style="text-align: center;">D3</td> </tr> <tr> <td>Application</td> <td>- Native CoA and re-coarctation</td> <td style="text-align: center;"><b>A1</b></td> <td style="text-align: center;">A2</td> <td style="text-align: center;">A3</td> </tr> <tr> <td>Patient</td> <td>- Patients with native CoA and re-coarctation - Sampling: n=24 - Mean age: 36 (18-60) years - Sex: 12 M; 12 F</td> <td style="text-align: center;"><b>P1</b></td> <td style="text-align: center;">P2</td> <td style="text-align: center;">P3</td> </tr> <tr> <td>Report</td> <td>- High quality.</td> <td style="text-align: center;"><b>R1</b></td> <td style="text-align: center;">R2</td> <td style="text-align: center;">R3</td> </tr> <tr> <td colspan="2" style="text-align: right;">Suitability Grade (Range 4-12)</td> <td colspan="3" style="text-align: center;">4</td> </tr> </table>	Level of Evidence	Study Method/Design	Question Applied	Oxford LOE 2011						Prospective observational study.	To evaluate the intermediate-term outcome of stent implantation for CoA in adults.	1	2	<b>3</b>	4	5	Suitability	Relevant Data	Grading			Device	- CP Stent (Bare and Covered) – 6 of the 24 patients had covered stents - BIB (manually crimped)	<b>D1</b>	D2	D3	Application	- Native CoA and re-coarctation	<b>A1</b>	A2	A3	Patient	- Patients with native CoA and re-coarctation - Sampling: n=24 - Mean age: 36 (18-60) years - Sex: 12 M; 12 F	<b>P1</b>	P2	P3	Report	- High quality.	<b>R1</b>	R2	R3	Suitability Grade (Range 4-12)		4				
Contribution																																																							
S&P	X (S only)																																																						
SOA	-																																																						
Level of Evidence	Study Method/Design	Question Applied	Oxford LOE 2011																																																				
	Prospective observational study.	To evaluate the intermediate-term outcome of stent implantation for CoA in adults.	1	2	<b>3</b>	4	5																																																
Suitability	Relevant Data	Grading																																																					
Device	- CP Stent (Bare and Covered) – 6 of the 24 patients had covered stents - BIB (manually crimped)	<b>D1</b>	D2	D3																																																			
Application	- Native CoA and re-coarctation	<b>A1</b>	A2	A3																																																			
Patient	- Patients with native CoA and re-coarctation - Sampling: n=24 - Mean age: 36 (18-60) years - Sex: 12 M; 12 F	<b>P1</b>	P2	P3																																																			
Report	- High quality.	<b>R1</b>	R2	R3																																																			
Suitability Grade (Range 4-12)		4																																																					



## NuMED Summary of Safety and Clinical Performance SSCP – Stents – CoA & RVOT

	Data Contribution	Relevant Data	Grading						
	Outcomes/Endpoints	- Decrease in systolic gradient - Increase in minimum aortic diameter - Adverse effects	Yes 1	No 2					
	Follow-up	- 24 hours post intervention and 33 (8-77) months	Yes 1	No 2					
	Statistical analysis	- All statistical tests were two-sided and a p-value <0.05 was considered statistically significant	Yes 1	No 2					
	Clinical significance	- Stenting in adults results in significant blood pressure gradient decrease and increase in vessel diameter. However, serious complications do occur and hypertension remains in the majority of patients.	Yes 1	No 2					
	Data Contribution Grade (Range 4-8)		4						
<b>Overall S&amp;P Appraisal, Disposition and Weighting</b>									
	S&P Grade (Range 9-25)	LOE (3) + Suitability (4) + Data Contribution (4) = 11	Disposition and Weighting (select)	<b>Accepted and Pivotal 9-12</b>			Accepted but not Pivotal, 13-21 Excluded, 22-25		
<b>Relevant S&amp;P Results</b>									
	Safety data	- One death due to aorta ruptured. - Two groin hematoma post-op.							
	Benefits/claims data	- Reduced in systolic gradient - Increased in minimum aortic diameter							
	Strengths	- N/A							
	Weaknesses/ Potential bias	- Only a small number of patients have undergone stent implantation since the authors started this procedure in 2003. This was a single-center report and patients were not compared with surgery or balloon angioplasty alone. Finally, 24-hour blood pressure monitoring before stenting was not performed in the majority of the patients. Post-stent 24-hour ambulatory blood pressure monitoring is therefore difficult to translate in terms of blood pressure reduction.							
<b>Safety &amp; Performance (for safety only)</b>									
<b>Appraisal</b>									
13. Agnoletti et al. (2009)	Level of Evidence	Study Method/Design	Question Applied			Oxford LOE 2011			
		Two arms comparative interventional study.	To compare the CP Stent and the Palmaz stent for treatment of native and postoperative lesions of CHD patients.			1	2	3	4
Contribution									
S&P	X (S only)								
SOA	-								
	Suitability	Relevant Data	Grading						
	Device	- CP Stent (Bare & Covered), crimped on BIB in 77 cases - 96 CP Stents (34 covered), 77 Palma Stents - Palmaz stent, crimped on BIB and simple balloons	D1	D2	D3				
	Application	- Patients with CHD (including CoA/re-coarctation, RVOT)	A1	A2		A3			
	Patient	- Patients with CHD (including CoA/re-coarctation, RVOT and other CHD conditions, such as	P1	P2		P3			



## NuMED Summary of Safety and Clinical Performance SSCP – Stents – CoA & RVOT

	<ul style="list-style-type: none"> <li>transposition of the great arteries, ventricular septal defect, single ventricle, etc.)</li> <li>- Sampling: n= 153               <ul style="list-style-type: none"> <li>- 89 patients with CP Stents (crimped on 77 BIB &amp; 12 other balloons)</li> <li>- 64 patients with Palmaz Stents (crimped on 23 BIB and 41 simple balloons)</li> </ul> </li> <li>- Mean age:               <ul style="list-style-type: none"> <li>- CP Stents: 15.4 (SD: 9.2) years</li> <li>- Palmaz Stents: 11.6 (SD: 8.1) years</li> </ul> </li> <li>- Sex: Not reported</li> </ul>			
Report	- High quality.	<b>R1</b>	R2	R3
Suitability Grade (Range 4-12)		6		
Data Contribution	Relevant Data	Grading		
Outcomes/Endpoints	<ul style="list-style-type: none"> <li>- Blood pressure gradient reduction</li> <li>- Vessel diameter reduction</li> <li>- Adverse effects</li> </ul>	<b>Yes 1</b>	No 2	
Follow-up	- Not reported.	Yes 1	<b>No 2</b>	
Statistical analysis	- A P-value less than 0.05 was considered statistically significant for stent group comparison.	<b>Yes 1</b>	No 2	
Clinical significance	<ul style="list-style-type: none"> <li>- The use of the CP Stents to treat stenotic lesions of CHD is effective and relatively safe. The overall efficacy of CP Stents for the treatment of stenotic lesions is superior to that of the Palmaz stent.</li> <li>- CP Stents' overall safety is higher than that of the Palmaz stent; but Palmaz stents have a lower profile when inserted.</li> </ul>	<b>Yes 1</b>	No 2	
Data Contribution Grade (Range 4-8)		5		
<b>Overall S&amp;P Appraisal, Disposition and Weighting</b>				
S&P Grade (Range 9-25)	LOE (3) + Suitability (6) + Data Contribution (5) = 14	Disposition and Weighting (select)	Accepted and Pivotal 9-12 <b>Accepted but not Pivotal, 13-21</b> Excluded, 22-25	
<b>Relevant S&amp;P Results</b>				
Safety data	<ul style="list-style-type: none"> <li>- Stent-related complications:               <ul style="list-style-type: none"> <li>- CP Stents (n=89 patients): 4 mild, 1 moderate, 1 severe.</li> <li>- Palmaz (n=64 patients): 10 mild, 2 moderate, 2 severe.</li> </ul> </li> <li>- Stent migration:               <ul style="list-style-type: none"> <li>- CP Stents: 7.</li> <li>- Palmaz: 4.</li> </ul> </li> <li>- Non stent related complications:               <ul style="list-style-type: none"> <li>- CP Stents: 1 mild, 2 moderate.</li> <li>- Palmaz: 1 mild, 2 moderate, 5 severe.</li> </ul> </li> <li>- Urgent surgery:               <ul style="list-style-type: none"> <li>- CP Stents: 2 due to homograft rupture and stent migration.</li> </ul> </li> </ul>			





## NuMED Summary of Safety and Clinical Performance SSCP – Stents – CoA & RVOT

		<ul style="list-style-type: none"> <li>- Palmaz: 1 for aortic dissection.</li> <li>- Balloon related complications: Balloon burst</li> <li>- CP Stents: 0.</li> <li>- Palmaz: 7 (3 in BIB, 3 with simple balloons, and 1 on pre-mounted stent).</li> </ul>																																
	Benefits/claims data	<ul style="list-style-type: none"> <li>- Decreased in blood pressure gradient.</li> <li>- Increased in vessel diameter.</li> </ul>																																
	Strengths	<ul style="list-style-type: none"> <li>- Efficacy of CP Stents was similar to that of Palmaz stent for stenting of the right ventricular outflow, and higher than that of Palmaz for the stenting of aorta, but the difference was not statistically.</li> </ul>																																
	Weaknesses/ Potential bias	<ul style="list-style-type: none"> <li>- Study presented retrospective results obtained in 153 consecutive patients.</li> <li>- CP stents were used for patients weighing more than 15 kg; and thus two populations were different concerning age, weight, indication for stenting; however, none of these differences were related to occurrence of complications.</li> <li>- Subgroup analyses were not performed.</li> </ul>																																
	<b>Safety &amp; Performance Appraisal</b>																																	
	Level of Evidence	Study Method/Design	Question Applied	Oxford LOE 2011																														
		Prospective interventional single arm study to evaluate placement of CP Stent (Covered) with smallest possible balloon sizes, then dilation with larger diameter balloons until pressure gradient was <20 mmHg and the stent was opposed to the aortic wall.	To report on the early results of treatment of native CoA by implantation and serial dilations of covered stents.	<table border="1" style="width: 100%; border-collapse: collapse;"> <tr> <td style="width: 12.5%; text-align: center;">1</td> <td style="width: 12.5%; text-align: center;">2</td> <td style="width: 12.5%; text-align: center;">3</td> <td style="width: 12.5%; text-align: center;">4</td> <td style="width: 12.5%; text-align: center;">5</td> </tr> </table>	1	2	3	4	5																									
1	2	3	4	5																														
14. Bruckheimer et al. (2009)	<table border="1" style="width: 100%; border-collapse: collapse;"> <thead> <tr> <th style="width: 15%;">Suitability</th> <th style="width: 60%;">Relevant Data</th> <th colspan="3" style="width: 25%;">Grading</th> </tr> </thead> <tbody> <tr> <td>Device</td> <td> <ul style="list-style-type: none"> <li>- Covered CP Stents</li> <li>- Manually crimped on a balloon</li> </ul> </td> <td style="text-align: center;"><b>D1</b></td> <td style="text-align: center;">D2</td> <td style="text-align: center;">D3</td> </tr> <tr> <td>Application</td> <td> <ul style="list-style-type: none"> <li>- Native CoA</li> </ul> </td> <td style="text-align: center;"><b>A1</b></td> <td style="text-align: center;">A2</td> <td style="text-align: center;">A3</td> </tr> <tr> <td>Patient</td> <td> <ul style="list-style-type: none"> <li>- Patients with native CoA</li> <li>- Sampling: n=22</li> <li>- Mean age: 15.5 (7.8 – 38.6) years</li> <li>- Sex: 14 M; 8 F</li> </ul> </td> <td style="text-align: center;"><b>P1</b></td> <td style="text-align: center;">P2</td> <td style="text-align: center;">P3</td> </tr> <tr> <td>Report</td> <td> <ul style="list-style-type: none"> <li>- High quality.</li> </ul> </td> <td style="text-align: center;"><b>R1</b></td> <td style="text-align: center;">R2</td> <td style="text-align: center;">R3</td> </tr> <tr> <td colspan="2" style="text-align: right;">Suitability Grade (Range 4-12)</td> <td colspan="3" style="text-align: center;">4</td> </tr> </tbody> </table>				Suitability	Relevant Data	Grading			Device	<ul style="list-style-type: none"> <li>- Covered CP Stents</li> <li>- Manually crimped on a balloon</li> </ul>	<b>D1</b>	D2	D3	Application	<ul style="list-style-type: none"> <li>- Native CoA</li> </ul>	<b>A1</b>	A2	A3	Patient	<ul style="list-style-type: none"> <li>- Patients with native CoA</li> <li>- Sampling: n=22</li> <li>- Mean age: 15.5 (7.8 – 38.6) years</li> <li>- Sex: 14 M; 8 F</li> </ul>	<b>P1</b>	P2	P3	Report	<ul style="list-style-type: none"> <li>- High quality.</li> </ul>	<b>R1</b>	R2	R3	Suitability Grade (Range 4-12)		4		
Suitability	Relevant Data	Grading																																
Device	<ul style="list-style-type: none"> <li>- Covered CP Stents</li> <li>- Manually crimped on a balloon</li> </ul>	<b>D1</b>	D2	D3																														
Application	<ul style="list-style-type: none"> <li>- Native CoA</li> </ul>	<b>A1</b>	A2	A3																														
Patient	<ul style="list-style-type: none"> <li>- Patients with native CoA</li> <li>- Sampling: n=22</li> <li>- Mean age: 15.5 (7.8 – 38.6) years</li> <li>- Sex: 14 M; 8 F</li> </ul>	<b>P1</b>	P2	P3																														
Report	<ul style="list-style-type: none"> <li>- High quality.</li> </ul>	<b>R1</b>	R2	R3																														
Suitability Grade (Range 4-12)		4																																
	Data Contribution	Relevant Data	Grading																															
	Outcomes/Endpoints	<ul style="list-style-type: none"> <li>- Increase of coarctation diameter</li> <li>- Reduction of peak pressure gradient</li> <li>- Adverse effects</li> </ul>	<b>Yes 1</b>	No 2																														
	Follow-up	<ul style="list-style-type: none"> <li>- Median 18.5 (1.6-31.4) months</li> </ul>	<b>Yes 1</b>	No 2																														



## NuMED Summary of Safety and Clinical Performance SSCP – Stents – CoA & RVOT

	<table border="1" style="width: 100%; border-collapse: collapse;"> <tr> <td style="width: 30%;">Statistical analysis</td> <td style="width: 35%;">- P-values reported.</td> <td style="width: 10%; text-align: center;"><b>Yes 1</b></td> <td style="width: 10%; text-align: center;">No 2</td> </tr> <tr> <td>Clinical significance</td> <td>- Serial dilation of CP Stents (Covered) is feasible, safe and an effective percutaneous method for the treatment of native CoA.</td> <td style="text-align: center;"><b>Yes 1</b></td> <td style="text-align: center;">No 2</td> </tr> <tr> <td colspan="2" style="text-align: right;">Data Contribution Grade (Range 4-8)</td> <td colspan="2" style="text-align: center;">4</td> </tr> </table>	Statistical analysis	- P-values reported.	<b>Yes 1</b>	No 2	Clinical significance	- Serial dilation of CP Stents (Covered) is feasible, safe and an effective percutaneous method for the treatment of native CoA.	<b>Yes 1</b>	No 2	Data Contribution Grade (Range 4-8)		4			
Statistical analysis	- P-values reported.	<b>Yes 1</b>	No 2												
Clinical significance	- Serial dilation of CP Stents (Covered) is feasible, safe and an effective percutaneous method for the treatment of native CoA.	<b>Yes 1</b>	No 2												
Data Contribution Grade (Range 4-8)		4													
<b>Overall S&amp;P Appraisal, Disposition and Weighting</b>															
S&P Grade (Range 9-25)	LOE (3) + Suitability (4) + Data Contribution (4) = 11	Disposition and Weighting (select)	<b>Accepted and Pivotal 9-12</b> Accepted but not Pivotal, 13-21 Excluded, 22-25												
<b>Relevant S&amp;P Results</b>															
Safety data	<ul style="list-style-type: none"> <li>- One small tear at the distal stent edge</li> <li>- One femoral pseudoaneurysm which spontaneously resolved</li> </ul>														
Performance data	<ul style="list-style-type: none"> <li>- Increase of coarctation diameter:               <ul style="list-style-type: none"> <li>- From 3.6 ± 1.9 mm pre-intervention to 12.6 ± 1.9 mm post-intervention, P=0.001</li> </ul> </li> <li>- Reduction of peak pressure gradient:               <ul style="list-style-type: none"> <li>- From 29.4 ± 8.5 to 6.7 ± 5.7 mmHg, P=0.001</li> </ul> </li> </ul>														
Benefits/claims data	<ul style="list-style-type: none"> <li>- Increase of coarctation diameter</li> <li>- Reduction of peak pressure gradient</li> </ul>														
Strengths	- N/A														
Weaknesses/ Potential bias	- No conflict of interest reported.														

15. Tzifa et al.  
(2006)

Contribution	
S&P	x
SOA	-

<b>Safety &amp; Performance Appraisal</b>							
Level of Evidence	Study Method/Design	Question Applied	Oxford LOE 2011				
	Single arm interventional study.	To evaluate the use of Covered CP Stents in treatment of CoA.	1	2	3	4	5
Suitability	Relevant Data	Grading					
Device	<ul style="list-style-type: none"> <li>- Covered CP Stent</li> <li>- BIB (hand-crimped)</li> </ul>	<b>D1</b>	D2	D3			
Application	- CoA	<b>A1</b>	A2	A3			
Patient	<ul style="list-style-type: none"> <li>- Patients with CoA (fully grown patients)</li> <li>- Sampling: n=30</li> <li>- Mean age: 28±17.5 (range 8 to 65) years</li> <li>- Sex: not reported</li> </ul>	<b>P1</b>	P2	P3			
Report	- High quality.	R1	<b>R2</b>	R3			
Suitability Grade (Range 4-12)			5				



## NuMED Summary of Safety and Clinical Performance SSCP – Stents – CoA & RVOT

Data Contribution	Relevant Data	Grading	
Outcomes/Endpoints	- Reduction in blood pressure gradient - Reduction in coarctation diameter	Yes 1	No 2
Follow-up	- 11 months	Yes 1	No 2
Statistical analysis	- Statistical significance was defined as P<0.05.	Yes 1	No 2
Clinical significance	- CP Stents (Covered) may be used as the therapy of choice in patients with complications after CoA repairs, whereas they provide a safe alternative to conventional stenting in patients with severe and complex CoA lesions or advanced age.	Yes 1	No 2
Data Contribution Grade (Range 4-8)		4	

### Overall S&P Appraisal, Disposition and Weighting

S&P Grade (Range 9-25)	LOE (4) + Suitability (5) + Data Contribution (4) = 13	Disposition and Weighting (select)	Accepted and Pivotal 9-12 <b>Accepted but not Pivotal, 13-21</b> Excluded, 22-25
---------------------------	---	------------------------------------	--

### Relevant S&P Results

Safety data	- Two stent fractures in the “old” design of the stent, no fractures in the “new” stent design Note: Since May 2002, the CP Stents (Covered) have been produced with reinforced golden soldering joints as the “new” stent design
Performance data	- Blood pressure gradient: From 36 + 20 mmHg to 4 + 4 mmHg, P<0.0001 - Diameter at coarctation: From 6.4 +3.8 mm to 17.1 + 3.1 mm, P<0.0001
Benefits/claims data	- Reduction in blood pressure gradient - Reduction in coarctation diameter - BIB allows readjustment of position after inflation of the inner balloon.
Strengths	- Covered stents were chosen: 1) as a rescue treatment in patients with CoA aneurysms or previous stent-related complications; and 2) in patients at risk of complications because of complex CoA anatomy or advanced age (defined as >65 years) - Covered CP stents are made of a framework of platinum iridium wire welded in a zig pattern. The addition of a gold soldering to each weld spot fills any voids caused by the welding and transfers the stresses to a larger area of the stent. The gold also serves to encapsulate the welded area, once again adding to the total strength of the weld. The stent is then fitted with a covering of ePTFE to achieve a solid tubular structure that retains fluid. The ePTFE covering is initially approximately 7 mm in diameter and will stretch over the range of diameters of expansion (usually from 12 to 24 mm diameter), and will always be taut over the stent when expanded. When the covering is mounted, it is folded over the crimped stent and expands uniformly when the balloon is inflated. - The BIB allows for readjustment of position after inflation of the inner balloon.
Weaknesses/ Potential bias	- Not reported.



## NuMED Summary of Safety and Clinical Performance SSCP – Stents – CoA & RVOT

<p>16. Meadows et al. (2015)</p> <table border="1" style="width: 100%; border-collapse: collapse;"> <tr> <th colspan="2">Contribution</th> </tr> <tr> <td style="width: 50%;">S&amp;P</td> <td>X (S only)</td> </tr> <tr> <td>SOA</td> <td>-</td> </tr> </table>	Contribution		S&P	X (S only)	SOA	-	<p><b>Safety &amp; Performance (for safety only)</b> This publication presents the results from the COAST trial for CP Stent (Bare and Covered) to treat native and recurrent CoA in selected children, adolescents and adult (NCT00552812). Please refer information presented in <b>Table G-1</b> for safety and performance of the subject devices, Study no. 1.</p>																																																										
Contribution																																																																	
S&P	X (S only)																																																																
SOA	-																																																																
<p>17. Taggart et al. (2016)</p> <table border="1" style="width: 100%; border-collapse: collapse;"> <tr> <th colspan="2">Contribution</th> </tr> <tr> <td style="width: 50%;">S&amp;P</td> <td>x</td> </tr> <tr> <td>SOA</td> <td>-</td> </tr> </table>	Contribution		S&P	x	SOA	-	<p><b>Safety &amp; Performance</b> This publication presents the results from the COAST II trial to evaluate the safety and short-term efficacy of the CP Stent in treating or preventing aortic wall injury in patients with CoA (NCT01278303). Please refer information presented in <b>Table G-1</b> for safety and performance of the subject devices, Study no. 2.</p>																																																										
Contribution																																																																	
S&P	x																																																																
SOA	-																																																																
<p>18. Sasikumar et al. (2020)</p> <table border="1" style="width: 100%; border-collapse: collapse;"> <tr> <th colspan="2">Contribution</th> </tr> <tr> <td style="width: 50%;">S&amp;P</td> <td>X (S only)</td> </tr> <tr> <td>SOA</td> <td>x</td> </tr> </table>	Contribution		S&P	X (S only)	SOA	x	<p><b>Safety &amp; Performance (for safety only)</b> <b>Appraisal</b></p> <table border="1" style="width: 100%; border-collapse: collapse;"> <thead> <tr> <th style="width: 25%;">Level of Evidence</th> <th style="width: 30%;">Study Method/Design</th> <th style="width: 30%;">Question Applied</th> <th colspan="5">Oxford LOE 2011</th> </tr> </thead> <tbody> <tr> <td></td> <td>Retrospective study.</td> <td>To study the early and late outcomes after stenting of native and recurrent CoA with uncovered and covered stents.</td> <td style="text-align: center;">1</td> <td style="text-align: center;">2</td> <td style="text-align: center;">3</td> <td style="text-align: center;">4</td> <td style="text-align: center;">5</td> </tr> </tbody> </table> <table border="1" style="width: 100%; border-collapse: collapse;"> <thead> <tr> <th style="width: 25%;">Suitability</th> <th style="width: 50%;">Relevant Data</th> <th colspan="3">Grading</th> </tr> </thead> <tbody> <tr> <td>Device</td> <td> <ul style="list-style-type: none"> <li>- CP Stent (Bare and Covered) – “D1” for subject devices</li> <li>- Unknown if pre-mounted on BIB</li> <li>- Other devices, including Advanta V12 stent (covered), Andra XL and XXL stents, Palmaz XL</li> </ul> </td> <td style="text-align: center;"><b>D1</b></td> <td style="text-align: center;">D2</td> <td style="text-align: center;">D3</td> </tr> <tr> <td>Application</td> <td>- CoA (native and recurrent)</td> <td style="text-align: center;"><b>A1</b></td> <td style="text-align: center;">A2</td> <td style="text-align: center;">A3</td> </tr> <tr> <td>Patient</td> <td> <ul style="list-style-type: none"> <li>- Patients with CoA (native and recurrent)</li> <li>- Sampling: n=45 (20 covered stents, 25 non-covered stents)                             <ul style="list-style-type: none"> <li>- Covered stents used were covered 7 CP Stent; 13 Advanta V12 Stent</li> <li>- Non-covered stents used were 17 CP Stent, 6 Andra XL and XXL stents, and 2 Palmaz XL.</li> </ul> </li> <li>- Mean age: 28±17.5 (range 8 to 65) years. Age per device group was not reported.</li> <li>- Sex: 32 M, 13 F. Sex per device group was not reported.</li> </ul> </td> <td style="text-align: center;">P1</td> <td style="text-align: center;"><b>P2</b></td> <td style="text-align: center;">P3</td> </tr> <tr> <td>Report</td> <td>- High quality with deficiencies</td> <td style="text-align: center;">R1</td> <td style="text-align: center;"><b>R2</b></td> <td style="text-align: center;">R3</td> </tr> <tr> <td colspan="2" style="text-align: right;">Suitability Grade (Range 4-12)</td> <td colspan="3" style="text-align: center;">6</td> </tr> </tbody> </table> <table border="1" style="width: 100%; border-collapse: collapse;"> <thead> <tr> <th style="width: 25%;">Data Contribution</th> <th style="width: 50%;">Relevant Data</th> <th colspan="2">Grading</th> </tr> </thead> <tbody> <tr> <td>Outcomes/Endpoints</td> <td>- Safety</td> <td style="text-align: center;"><b>Yes 1</b></td> <td style="text-align: center;">No 2</td> </tr> <tr> <td>Follow-up</td> <td>- Covered stent group: 57 months</td> <td style="text-align: center;"><b>Yes 1</b></td> <td style="text-align: center;">No 2</td> </tr> </tbody> </table>	Level of Evidence	Study Method/Design	Question Applied	Oxford LOE 2011						Retrospective study.	To study the early and late outcomes after stenting of native and recurrent CoA with uncovered and covered stents.	1	2	3	4	5	Suitability	Relevant Data	Grading			Device	<ul style="list-style-type: none"> <li>- CP Stent (Bare and Covered) – “D1” for subject devices</li> <li>- Unknown if pre-mounted on BIB</li> <li>- Other devices, including Advanta V12 stent (covered), Andra XL and XXL stents, Palmaz XL</li> </ul>	<b>D1</b>	D2	D3	Application	- CoA (native and recurrent)	<b>A1</b>	A2	A3	Patient	<ul style="list-style-type: none"> <li>- Patients with CoA (native and recurrent)</li> <li>- Sampling: n=45 (20 covered stents, 25 non-covered stents)                             <ul style="list-style-type: none"> <li>- Covered stents used were covered 7 CP Stent; 13 Advanta V12 Stent</li> <li>- Non-covered stents used were 17 CP Stent, 6 Andra XL and XXL stents, and 2 Palmaz XL.</li> </ul> </li> <li>- Mean age: 28±17.5 (range 8 to 65) years. Age per device group was not reported.</li> <li>- Sex: 32 M, 13 F. Sex per device group was not reported.</li> </ul>	P1	<b>P2</b>	P3	Report	- High quality with deficiencies	R1	<b>R2</b>	R3	Suitability Grade (Range 4-12)		6			Data Contribution	Relevant Data	Grading		Outcomes/Endpoints	- Safety	<b>Yes 1</b>	No 2	Follow-up	- Covered stent group: 57 months	<b>Yes 1</b>	No 2
Contribution																																																																	
S&P	X (S only)																																																																
SOA	x																																																																
Level of Evidence	Study Method/Design	Question Applied	Oxford LOE 2011																																																														
	Retrospective study.	To study the early and late outcomes after stenting of native and recurrent CoA with uncovered and covered stents.	1	2	3	4	5																																																										
Suitability	Relevant Data	Grading																																																															
Device	<ul style="list-style-type: none"> <li>- CP Stent (Bare and Covered) – “D1” for subject devices</li> <li>- Unknown if pre-mounted on BIB</li> <li>- Other devices, including Advanta V12 stent (covered), Andra XL and XXL stents, Palmaz XL</li> </ul>	<b>D1</b>	D2	D3																																																													
Application	- CoA (native and recurrent)	<b>A1</b>	A2	A3																																																													
Patient	<ul style="list-style-type: none"> <li>- Patients with CoA (native and recurrent)</li> <li>- Sampling: n=45 (20 covered stents, 25 non-covered stents)                             <ul style="list-style-type: none"> <li>- Covered stents used were covered 7 CP Stent; 13 Advanta V12 Stent</li> <li>- Non-covered stents used were 17 CP Stent, 6 Andra XL and XXL stents, and 2 Palmaz XL.</li> </ul> </li> <li>- Mean age: 28±17.5 (range 8 to 65) years. Age per device group was not reported.</li> <li>- Sex: 32 M, 13 F. Sex per device group was not reported.</li> </ul>	P1	<b>P2</b>	P3																																																													
Report	- High quality with deficiencies	R1	<b>R2</b>	R3																																																													
Suitability Grade (Range 4-12)		6																																																															
Data Contribution	Relevant Data	Grading																																																															
Outcomes/Endpoints	- Safety	<b>Yes 1</b>	No 2																																																														
Follow-up	- Covered stent group: 57 months	<b>Yes 1</b>	No 2																																																														



## NuMED

### Summary of Safety and Clinical Performance

#### SSCP – Stents – CoA & RVOT

	- Non-covered stent group: 35 months		
Statistical analysis	- Statistical analysis was done by the Statistical Package for Social Sciences (version 21.0). Quantitative data were presented as mean ± SD or as median and range and qualitative data were presented as frequency (percentages). The categorical parameters were compared by chi-square test, and the continuous variables were compared by Student t test for independent continuous data and Manne Whitney U test for nonparametric data.	Yes 1	No 2
Clinical significance	- Not reported specifically for subject devices.	Yes 1	No 2
Data Contribution Grade (Range 4-8)		5	

#### Overall S&P Appraisal, Disposition and Weighting

S&P Grade (Range 9-25)	LOE (3) + Suitability (6) + Data Contribution (5) = 14	Disposition and Weighting (select)	Accepted and Pivotal 9-12 <b>Accepted but not Pivotal, 13-21</b> Excluded, 22-25
---------------------------	---	------------------------------------	--

#### Relevant S&P Results

Safety data	Outcomes	Covered (n=18)	Uncovered (bare metal) (n=8)
	Late lumen loss (no or mild)	2 (Advanta 1, CP 1)	4 (CP 3, Palmaz 1)
	Late lumen loss (moderate)	12 (Advanta 7, CP 4, Andra 1)	4 (CP3, Palmaz 1)
	Late lumen loss (severe)	4 (Advanta 3, CP 1)	0
	Fracture	1 Advanta	0
Performance data	- Not reported specially for subject devices.		
Benefits/claims data	- Not reported		
Strengths	- Not reported.		
Weaknesses/ Potential bias	- Not reported.		

#### State of the Art

##### Appraisal

Medical condition		Alternatives		Risk/benefit		Side-effects		Equivalence		Surrogate endpoints	
Yes 1	No 2	Yes 1	No 2	Yes 1	No 2	Yes 1	No 2	Yes 1	No 2	Yes 1	No 2

#### Overall SOA Appraisal and Disposition

SOA Grade (Range 6-12)	8	Disposition (select)	<b>Accepted, &lt; 12</b> Excluded, 12
---------------------------	---	----------------------	--

#### Relevant SOA Results



## NuMED Summary of Safety and Clinical Performance SSCP – Stents – CoA & RVOT

	SOA data	<ul style="list-style-type: none"> <li>- Patients in the covered stent group were older and had greater basal pressure gradient. More patients in the covered stent group had residual gradient &gt;10 mm Hg after the procedure.</li> <li>- No mortality or aortic wall injury in either group.</li> <li>- Mean number of anti-hypertensive was 1.38 ± 0.74 in the covered group and 1+0.7 in the uncovered group</li> <li>- Greater incidence of severe late lumen loss (&gt;30% lumen loss) in the covered stent group on follow-up. According to the authors, this phenomenon was brand specific (Advanta V12 stent). Single strut fracture which was not causing any lumen obstruction was also noted in one Advanta V12 stent. The stents have an open cell stent geometry with consequent less radial strength.</li> <li>- A previous study on Advanta stent implantation in 25 patients did not show any complications related to the stent. However, the median period of follow-up in that study was only 4.9 months and longer follow-up is needed to look for re-coarctation or aneurysm formation.</li> <li>- Another study described 2 patients with Advanta stent implantation who developed in-folding of the proximal edge of the stent on follow-up and both the cases were managed by re-stenting. The authors had a similar proximal stent collapse in a patient who had Advanta stent implantation, which was managed by balloon angioplasty. Though the residual gradient was 5 mm Hg immediately after the balloon angioplasty, the gradient increased to 25mmHg on follow-up and he underwent a repeat balloon angioplasty with good result.</li> </ul>												
	Comments	<ul style="list-style-type: none"> <li>- Uncovered stents can be safely implanted with minimal risk of aortic wall injury in patients with low risk anatomic features. Covered stent implantation is associated with higher incidence of planned and unplanned re-intervention.</li> </ul>												
	<b>Safety &amp; Performance</b>													
	<b>Appraisal</b>													
	Level of Evidence	Study Method/Design	Question Applied	Oxford LOE 2011										
		Single center retrospective study.	The aim of this study was to investigate the impact and safety of covered stent placement for treatment of (re)CoA during a longer follow-up period.	1	2	3	4	5						
	Suitability	Relevant Data			Grading									
	Device	<ul style="list-style-type: none"> <li>- Only 8-zig covered Cheatham Platinum (CCP) stents (NuMED Inc. Hopkinton, NY, USA) were included in the study; 8z22 (1.1%), 8z28 (18.0%), 8z34 (25.8%), 8z39 (25.8%), 8z45 (28.1%) and 8z55 (1.1%).</li> <li>- Unknown if pre-mounted on BIB</li> </ul>			D1	D2	D3							
	Application	<ul style="list-style-type: none"> <li>- CoA (recurrent)</li> </ul>			A1	A2	A3							
	Patient	<ul style="list-style-type: none"> <li>- Patients with CoA who were treated with 102 covered stents from 2003 to 2017</li> <li>- All patients with a covered stent implantation for a native CoA or reCoA after surgical or transcatheter repair were included.</li> <li>- 89 patients with 102 covered stents in 93 procedures</li> <li>- Mean age 23.9±15.8 years (5.1-71.6)</li> <li>- 35 patients &lt;16 years and 54 patients ≥16 years</li> <li>- 60 (67.4%) male and 29 (32.6%) female</li> </ul>			P1	P2	P3							
	Report	<ul style="list-style-type: none"> <li>- High quality report</li> </ul>			R1	R2	R3							
19. Stassen et al. (2021) <table border="1" style="margin-top: 10px; width: 100px; border-collapse: collapse;"> <thead> <tr> <th colspan="2" style="background-color: #e0e0e0;">Contribution</th> </tr> </thead> <tbody> <tr> <td style="background-color: #e0e0e0;">S&amp;P</td> <td style="text-align: center;">x</td> </tr> <tr> <td style="background-color: #e0e0e0;">SOA</td> <td style="text-align: center;">x</td> </tr> </tbody> </table>	Contribution		S&P	x	SOA	x								
Contribution														
S&P	x													
SOA	x													



## NuMED Summary of Safety and Clinical Performance SSCP – Stents – CoA & RVOT

Suitability Grade (Range 4-12)      4

Data Contribution	Relevant Data	Grading	
Outcomes/Endpoints	- Short-term pre/post-implant hemodynamics and angiographic data were reported. Changes in blood pressure, the use of antihypertensive drugs and complications were recorded during follow-up.	Yes 1	No 2
Follow-up	- Mean follow-up time was 6.6±3.7 years (min max range 0.2-15.7 years).	Yes 1	No 2
Statistical analysis	- Continuous variables are presented as mean plus minus standard deviation (range minimum–maximum). In case of an asymmetric distribution of data, results are reported as median (interquartile range (IQR)). Proportions are noted as number and percentage. Comparison of individual parameters before and after stenting was performed using the two-tailed paired t test. Categorical data were compared with a McNemar. A p value of less than 0.05 was considered statistically significant. Statistical analysis was done using the SPSS software version 26 package (SPSS Inc., Chicago, IL USA).	Yes 1	No 2
Clinical significance	- The magnitude of the treatment effect observed was clinically significant.	Yes 1	No 2
Data Contribution Grade (Range 4-8)		4	

### Overall S&P Appraisal, Disposition and Weighting

S&P Grade (Range 9-25)	LOE (3) + Suitability (4) + Data Contribution (4) = 11	Disposition and Weighting (select)	<b>Accepted and Pivotal 9-12</b> Accepted but not Pivotal, 13-21 Excluded, 22-25
---------------------------	---	------------------------------------	--

### Relevant S&P Results

Safety data	- Long-term adverse events were found in 4.5% of patients (covered stent fracture (n=3), aneurysm formation (n=2)).
Performance data	- The procedural success rate was 100%. - The mean invasive ascending-to-descending aorta systolic gradient under general anaesthesia decreased from 25 ± 16mmHg to 4 ± 7mmHg (p<0.001). After a mean follow-up time of 6.6 ± 3.7 years, there was a persistent improvement of the mean systolic blood pressure gradient between right arm and leg (~7 ± 18 vs 38 ± 24mmHg; p<0.001). A larger proportion of patients required antihypertensive medication (33.7% vs 50.0%, p=0.017) and needed ≥ two drugs (20.2% vs 27.4%, p=0.066) to control blood pressure.
Benefits/claims data	- Covered stent implantation for CoA is highly successful, safe and results in a persistent hemodynamic improvement in the immediate and long-term outcome. Lifelong follow-up with additional antihypertensive drug treatment is mandatory to maintain favourable hemodynamic results after stenting.
Strengths	- Patients were followed for a mean period of 6.6±3.7 years (maximum follow-up time 15.7 years). To authors knowledge, this is the largest study with the longest follow-up of the use of covered stents in (re)CoA.
Weaknesses/ Potential bias	- Retrospective design - Single centre design - Incomplete follow-up achieved: 14 of the 89 patients had no follow-up data. Among the 75 remaining patients, 47 had 5 years follow-up (so with imaging).



## NuMED

### Summary of Safety and Clinical Performance

#### SSCP – Stents – CoA & RVOT

	<p><b>State of the Art Appraisal</b></p> <table border="1" style="width: 100%; border-collapse: collapse;"> <thead> <tr> <th colspan="2">Medical condition</th> <th colspan="2">Alternatives</th> <th colspan="2">Risk/benefit</th> <th colspan="2">Side-effects</th> <th colspan="2">Equivalence</th> <th colspan="2">Surrogate endpoints</th> </tr> </thead> <tbody> <tr> <td>Yes 1</td> <td>No 2</td> <td>Yes 1</td> <td>No 2</td> <td>Yes 1</td> <td>No 2</td> <td>Yes 1</td> <td>No 2</td> <td>Yes 1</td> <td>No 2</td> <td>Yes 1</td> <td>No 2</td> </tr> </tbody> </table> <p><b>Overall SOA Appraisal and Disposition</b></p> <table border="1" style="width: 100%; border-collapse: collapse;"> <tr> <td style="width: 25%;">SOA Grade (Range 6-12)</td> <td style="width: 30%;">7</td> <td style="width: 30%;">Disposition (select)</td> <td style="width: 15%;">Accepted, &lt; 12 Excluded, 12</td> </tr> </table> <p><b>Relevant SOA Results</b></p> <table border="1" style="width: 100%; border-collapse: collapse;"> <tr> <td style="width: 30%; vertical-align: top;">SOA data</td> <td> <p>CoA:</p> <ul style="list-style-type: none"> <li>- CoA is a congenital cardio-vascular malformation, characterised by a restriction of the lumen of the thoracic aorta. It occurs in approximately 4 of 10,000 live births and comprises 5% to 8% of CHD.</li> <li>- Mostly, CoA is detected in childhood and repaired surgically or by endovascular therapy. Occasionally it is diagnosed in adolescence or adulthood by investigations done for systemic hypertension.</li> <li>- The natural history of CoA carries a poor prognosis due to complications such as left ventricular failure, intracranial haemorrhage, aortic rupture or dissection, premature coronary artery disease and sudden death.</li> <li>- Smaller and younger infants are typically treated surgically but remain at risk for recurrent obstruction with up to 10% requiring further intervention during adulthood.</li> <li>- In older children and adults, the preferred treatment method depends on the individual anatomy and nature of the lesion, but endovascular therapy with either balloon angioplasty or stent implantation is commonly preferred over surgery. Although balloon angioplasty results in excellent acute hemodynamics, it is associated with a high rate of aortic wall injury and recurrent obstruction. Because of these concerns, stent implantation is usually favoured to avoid overdistension or the elastic recoil of the aorta. Bare stent implantation has become a worthy alternative to surgery and balloon angioplasty and seems to lead to better results and fewer complications. However, although interventions with bare stent implantation seem efficient and generally safe, major complications such as local aneurysm formation, aortic rupture, dissection and even death may occur. To prevent these aortic wall injuries (AWI) during the stent procedure, covered stents are increasingly used and their safety and efficacy for immediate and intermediate follow-up have been demonstrated. However, long-term results remain limited.</li> </ul> <p>Covered stents:</p> <ul style="list-style-type: none"> <li>- Covered stents are increasingly used in severe and complex coarctations of the aorta, mainly to avoid the risk of aortic wall injuries such as local aneurysm formation, dissection and aortic rupture. Nevertheless, the aorta can still rupture with a covered stent, but no unlimited bleeding will occur, unless there was insufficient sealing, the covering was torn or in case of vessel tear with retrograde bleeding from collaterals.</li> </ul> </td> </tr> <tr> <td style="vertical-align: top;">Comments</td> <td>- Not reported</td> </tr> </table>											Medical condition		Alternatives		Risk/benefit		Side-effects		Equivalence		Surrogate endpoints		Yes 1	No 2	Yes 1	No 2	Yes 1	No 2	Yes 1	No 2	Yes 1	No 2	Yes 1	No 2	SOA Grade (Range 6-12)	7	Disposition (select)	Accepted, < 12 Excluded, 12	SOA data	<p>CoA:</p> <ul style="list-style-type: none"> <li>- CoA is a congenital cardio-vascular malformation, characterised by a restriction of the lumen of the thoracic aorta. It occurs in approximately 4 of 10,000 live births and comprises 5% to 8% of CHD.</li> <li>- Mostly, CoA is detected in childhood and repaired surgically or by endovascular therapy. Occasionally it is diagnosed in adolescence or adulthood by investigations done for systemic hypertension.</li> <li>- The natural history of CoA carries a poor prognosis due to complications such as left ventricular failure, intracranial haemorrhage, aortic rupture or dissection, premature coronary artery disease and sudden death.</li> <li>- Smaller and younger infants are typically treated surgically but remain at risk for recurrent obstruction with up to 10% requiring further intervention during adulthood.</li> <li>- In older children and adults, the preferred treatment method depends on the individual anatomy and nature of the lesion, but endovascular therapy with either balloon angioplasty or stent implantation is commonly preferred over surgery. Although balloon angioplasty results in excellent acute hemodynamics, it is associated with a high rate of aortic wall injury and recurrent obstruction. Because of these concerns, stent implantation is usually favoured to avoid overdistension or the elastic recoil of the aorta. Bare stent implantation has become a worthy alternative to surgery and balloon angioplasty and seems to lead to better results and fewer complications. However, although interventions with bare stent implantation seem efficient and generally safe, major complications such as local aneurysm formation, aortic rupture, dissection and even death may occur. To prevent these aortic wall injuries (AWI) during the stent procedure, covered stents are increasingly used and their safety and efficacy for immediate and intermediate follow-up have been demonstrated. However, long-term results remain limited.</li> </ul> <p>Covered stents:</p> <ul style="list-style-type: none"> <li>- Covered stents are increasingly used in severe and complex coarctations of the aorta, mainly to avoid the risk of aortic wall injuries such as local aneurysm formation, dissection and aortic rupture. Nevertheless, the aorta can still rupture with a covered stent, but no unlimited bleeding will occur, unless there was insufficient sealing, the covering was torn or in case of vessel tear with retrograde bleeding from collaterals.</li> </ul>	Comments	- Not reported
Medical condition		Alternatives		Risk/benefit		Side-effects		Equivalence		Surrogate endpoints																																	
Yes 1	No 2	Yes 1	No 2	Yes 1	No 2	Yes 1	No 2	Yes 1	No 2	Yes 1	No 2																																
SOA Grade (Range 6-12)	7	Disposition (select)	Accepted, < 12 Excluded, 12																																								
SOA data	<p>CoA:</p> <ul style="list-style-type: none"> <li>- CoA is a congenital cardio-vascular malformation, characterised by a restriction of the lumen of the thoracic aorta. It occurs in approximately 4 of 10,000 live births and comprises 5% to 8% of CHD.</li> <li>- Mostly, CoA is detected in childhood and repaired surgically or by endovascular therapy. Occasionally it is diagnosed in adolescence or adulthood by investigations done for systemic hypertension.</li> <li>- The natural history of CoA carries a poor prognosis due to complications such as left ventricular failure, intracranial haemorrhage, aortic rupture or dissection, premature coronary artery disease and sudden death.</li> <li>- Smaller and younger infants are typically treated surgically but remain at risk for recurrent obstruction with up to 10% requiring further intervention during adulthood.</li> <li>- In older children and adults, the preferred treatment method depends on the individual anatomy and nature of the lesion, but endovascular therapy with either balloon angioplasty or stent implantation is commonly preferred over surgery. Although balloon angioplasty results in excellent acute hemodynamics, it is associated with a high rate of aortic wall injury and recurrent obstruction. Because of these concerns, stent implantation is usually favoured to avoid overdistension or the elastic recoil of the aorta. Bare stent implantation has become a worthy alternative to surgery and balloon angioplasty and seems to lead to better results and fewer complications. However, although interventions with bare stent implantation seem efficient and generally safe, major complications such as local aneurysm formation, aortic rupture, dissection and even death may occur. To prevent these aortic wall injuries (AWI) during the stent procedure, covered stents are increasingly used and their safety and efficacy for immediate and intermediate follow-up have been demonstrated. However, long-term results remain limited.</li> </ul> <p>Covered stents:</p> <ul style="list-style-type: none"> <li>- Covered stents are increasingly used in severe and complex coarctations of the aorta, mainly to avoid the risk of aortic wall injuries such as local aneurysm formation, dissection and aortic rupture. Nevertheless, the aorta can still rupture with a covered stent, but no unlimited bleeding will occur, unless there was insufficient sealing, the covering was torn or in case of vessel tear with retrograde bleeding from collaterals.</li> </ul>																																										
Comments	- Not reported																																										

| 20. Holzer et al. (2021) | **Safety & Performance (for safety only)**  **Appraisal**   | Contribution | Level of Evidence | Study Method/Design                    | Question Applied                              | Oxford LOE 2011 |   |   |   |   | |--------------|-------------------|--|---|-----------------|---|---|---|---| | S&P          | x (S)             | Included in this report are the 5-year | To report the late-term follow-up data and to | 1               | 2 | 3 | 4 | 5 | | | | | | | | | | | |





## NuMED Summary of Safety and Clinical Performance SSCP – Stents – CoA & RVOT

	only)						
SOA	x	<p>follow-up data from patients in these 2 trials and those treated through their Continued Access protocols. Data was prospectively collected during the 2 multi-center studies.</p>	<p>compare this to earlier follow-up data. For the purpose of this study, follow-up was defined as:</p> <ul style="list-style-type: none"> <li>- Immediate (1 month),</li> <li>- Early (12 months),</li> <li>- Late (48 or 60 months).</li> </ul> <p>To identify possible predictors of late-term outcome post-stent implantation.</p>				
		<b>Suitability</b>	<b>Relevant Data</b>	<b>Grading</b>			
		Device	<ul style="list-style-type: none"> <li>- CP Stent (Bare and Covered)</li> <li>- 52% received covered stents and 48% received bare stents.</li> <li>- No data if pre-mounted or not with BIB</li> <li>- The minimum stent diameter was 14.4mm (interquartile range (IQR), 12.6-16.0mm) with a minimum stent diameter to the aorta at diaphragm ratio of 0.87 (IQR, 0.77-1.0).</li> </ul>	<b>D1</b>	D2	D3	
		Application	<ul style="list-style-type: none"> <li>- CoA (native or recurrent)</li> <li>- Native coarctation was present in 49%, postsurgical in 24% and postcatheterization in 27%.</li> <li>- The minimum coarctation diameter was 8.0mm (IQR, 5.4-10.5mm), and median aortic diameter at the diaphragm was 16.0mm (IQR, 14.0-19.0mm).</li> </ul>	<b>A1</b>	A2	A3	
		Patient	<ul style="list-style-type: none"> <li>- All patients enrolled in the COAST or COAST II trials and their Continued Access extensions were included. Patients without late follow-up data were excluded from analysis, except for analyzing the estimated cumulative incidence of stent fractures, aortic wall injury, and reinterventions.</li> <li>- Cohort of 248 patients</li> <li>- COAST: 105 patients enrolled in COAST with 16 Continued Access patients (n=121)</li> <li>- COAST II: 82 participants from COAST II with an additional 45 Continued access patients (n=127).</li> <li>- From the 180 patient cohort, the median age at implant was 17 years (IQR, 13-28 years), the median weight (66.3kg, IQR, 53.8-78.1kg).</li> </ul>	<b>P1</b>	P2	P3	
		Report	<ul style="list-style-type: none"> <li>- High quality report</li> </ul>	<b>R1</b>	R2	R3	
				Suitability Grade (Range 4-12)		4	
		<b>Data Contribution</b>	<b>Relevant Data</b>	<b>Grading</b>			
		Outcomes/Endpoints	<ul style="list-style-type: none"> <li>- Parameters used to assess aortic stent outcomes:</li> <li>- Hemodynamic</li> <li>- Systemic systolic hypertension</li> <li>- Use of antihypertensive medication</li> <li>- Upper limb to lower limb blood pressure difference of <math>\geq 20</math>mm Hg</li> <li>- Reinterventions</li> <li>- Stent fractures</li> <li>- Aortic wall injury</li> </ul>	<b>Yes 1</b>		No 2	



## NuMED Summary of Safety and Clinical Performance SSCP – Stents – CoA & RVOT

		<ul style="list-style-type: none"> <li>- Predictor variables used to assess late-term results:</li> <li>-Demographics</li> <li>-Type of coarctation</li> <li>-Preimplantation clinical data</li> <li>-Baseline characterization data</li> <li>-Type of stent</li> <li>-Poststent catheterization data</li> <li>-Postcatheterization data</li> </ul>		
	Follow-up	<ul style="list-style-type: none"> <li>- Follow-up data was collected at 1, 6, 12, 24, 36, 48 and 60 months and included MRI at 12 and 24 months, and fluoroscopy at 12, 24, 48 and 60 months.</li> <li>- 96% of patients returned for 1-month follow-up, 86% for 12-month follow-up, and 63% for 60-month.</li> <li>- A total of 180 patients (73%) had either 48- or 60-month follow-up data.</li> <li>- Out of the 180 patients with late follow-up, 177 (98%) had also immediate and 180 (100%) early follow-up data available for analysis.</li> <li>- Aortic imaging (either MRI, computed tomography, or angiography) was available for 180/180 (100%) at immediate follow-up, 177/180 (98%) at intermediate follow-up, and 41:180 (23%) at late follow-up. Fluoroscopy was available for 180/180 (100%) at immediate follow-up, 178/180 (99%) at intermediate follow-up, and 136/180 (76%) at late follow-up.</li> </ul>	Yes 1	No 2
	Statistical analysis	<ul style="list-style-type: none"> <li>- Categorical variables are summarized as frequencies and percentages, and continuous variables as either means and SDs or medians with interquartile range (IQR) as noted. For the entire cohort, the cumulative incidence of stents fractures, aortic wall injury, and reinterventions at immediate, early and late follow-up was estimated using the Kaplan-Meier method. Patients who did not have an outcome event were censored at time. Changes in hemodynamic measures over time were evaluated using tests of trend. For patients with late follow-up, associations between patient and procedure characteristics and 4 binary outcome variables – suboptimal hemodynamic outcome, stent fractur, catheter reintervention, and aortic wall injury – were assessed using Fisher exact test. Characteristics significant at the 0.20 level were considered for inclusion in multivariable logistic regression models. Forward selection was used, and P &lt;0.05 was required for retention in the final model. To assess generalizability, characteristics of patients with and without late follow-up were compared using Fisher exact and Wilcoxon rank sum tests; no significant differences were found. All analytics were performed using SAS software version 9.4.</li> </ul>	Yes 1	No 2
	Clinical significance	<ul style="list-style-type: none"> <li>- Coarctation stenting is effective at maintaining obstruction relief up to 60 months postimplant with reduction in the number of patients requiring antihypertensive medication. However, an increase in-stent fractures and reinterventions were observed between medium and long-term follow-up. Covered stents appear to confer some protection from the development of stent fractures but do not provide complete protection from late aneurysm formation.</li> </ul>	Yes 1	No 2
Data Contribution Grade (Range 4-8)				4



## NuMED

### Summary of Safety and Clinical Performance

### SSCP – Stents – CoA & RVOT

Overall S&P Appraisal, Disposition and Weighting			
S&P Grade (Range 9-25)	LOE (3) + Suitability (4) + Data Contribution (4) = 11	Disposition and Weighting (select)	<b>Accepted and Pivotal 9-12</b> Accepted but not Pivotal, 13-21 Excluded, 22-25
Relevant S&P Results			
Safety data	<p><b>Aortic Wall Injury:</b></p> <ul style="list-style-type: none"> <li>- 13 patients were identified as having aneurysms or pseudo-aneurysms (COAST: 6/121 [5%], COAST II: 7/127 [5.5%].</li> <li>- No dissections were found.</li> <li>- The cumulative incidence was 1.2% by early and 6.3% by late follow-up.</li> <li>- In 3 patients, the aneurysm was proximal to the implanted stent, in one patient the location was not specified, and in the remainder, the aneurysm was within the borders of the implanted stent.</li> <li>- In 4 of 13 patients, aneurysms were identified on MRI or computed tomography before reintervention, while in 9 patients the aneurysms were diagnosed by angiography during catheterization performed for other reasons such as elective stent re-expansion.</li> <li>- 17 patients had covered stents implanted to treat the aneurysm; 2 did not.</li> <li>- By univariate analysis, coarctation minimum diameter &lt;6mmm was the only factor significantly associated with aortic wall injury (12% versus 2%, P=0.007).</li> <li>- There was a borderline relationship between minimum stent to aortic diameter at the diaphragm &lt;0.7 and aortic wall injury (19% versus 5%, M=0.059).</li> <li>- Aneurysms did not just occur in patients with bare metal stents, but equally in patients who had covered stents implanted. As such, the notion that covered stent implantation confers long-term protection from the development of aneurysm, may not be the case. Data are in contrast with Butera et al.<sup>1</sup> who did show a significant difference in the incidence of aneurysm formation when comparing patients bare versus covered stents, albeit in a much smaller cohort. Also, the median follow-up in that study was significantly longer for those with bare stents compared with covered stents (85 versus 35 months). This is important as the current study demonstrates that the majority of aneurysms were not identified until late follow-up.</li> <li>- Most aneurysms developed within the borders of the stent, including covered stents. One possible explanation is that pressure within the aorta distributes flow between the stent and the aortic wall, eventually leading to aneurysm formation. Another possibility is that the expanded polytetrafluoroethylene became damaged during initial implantation.</li> <li>- Current study did not investigate the benefit of a covered stent to reduce the risk of acute aortic wall injury during stent implantation because cases have not been randomly assigned and high-risk patients were excluded for bare stent implantation and received covered stents.</li> </ul> <p><b>Other Adverse Events:</b></p> <ul style="list-style-type: none"> <li>- Over the follow-up period, 2 patients had additional adverse events that were captured in the data set. One patient had a self-resolving neurological adverse event (possible transient ischemic attack) 2 weeks after the procedure without any clear relationship to the procedure itself. Another patient developed cardiogenic/septic chock 7 months after the procedure. No other serious adverse events were documented in any patients.</li> </ul>		

<sup>1</sup> Butera G, Manica JL, Marini D, Piazza L, Chessa M, Filho RI, Sarmento Leite RE, Carminati M. From bare to covered: 15-year single center experience and follow-up in trans-catheter stent implantation for aortic coarctation. Catheter Cardiovasc Interv. 2014 May 1;83(6):953-63. doi: 10.1002/ccd.25404. Epub 2014 Feb 4. PMID: 24459104.



## NuMED Summary of Safety and Clinical Performance SSCP – Stents – CoA & RVOT

		<ul style="list-style-type: none"> <li>- At late follow-up, freedom from surgical intervention was 100%, catheter reintervention 78.7%, stent fracture 75.6%, and freedom from aortic wall injury 93.7%.</li> <li>- 44% of patients had suboptimal long-term hemodynamic outcomes.</li> <li>- It has documented that hemodynamic results are generally maintained over the follow-up period. Stent fractures, catheterization reinterventions, and aortic wall injuries, all increase in frequency between medium and long-term follow-up. Overed stents appear to confer some protection from the development of stent fractures, but they do not provide complete protection from late aneurysm formation.</li> </ul>									
	Benefits/claims data										
	Strengths	<ul style="list-style-type: none"> <li>- The largest study to date with comprehensive follow-up data up to 60 months post-procedure.</li> </ul>									
	Weaknesses/ Potential bias	<ul style="list-style-type: none"> <li>- Small sample size</li> <li>- Did not have the statistical power to evaluate all parameters contributing to long-term morbidity in these patients, such as aortic wall injury.</li> <li>- Once the COAST studies were closed, it was not permissible to contact centers for additional data regarding stent fractures, indications for reintervention and other clinical data elements.</li> <li>- There were inherent differences between COAST and COAST II enrollment indications and the way some of the data was collected.</li> <li>- While this study defined 48 to 60 months follow-up as long-term, this is still a relatively short time period.</li> <li>- This study only analyzed the outcome of stent implantation for coarctation using CP stents. It did not compare the outcome of stent implantation to other treatment modalities, as was done in the Congenital Cardiovascular Interventional Study Consortium Report.<sup>2</sup></li> </ul>									
<b>State of the Art</b>											
<b>Appraisal</b>											
Medical condition		Alternatives		Risk/benefit		Side-effects		Equivalence		Surrogate endpoints	
<b>Yes 1</b>	No 2	<b>Yes 1</b>	No 2	<b>Yes 1</b>	No 2	<b>Yes 1</b>	No 2	Yes 1	<b>No 2</b>	<b>Yes 1</b>	No 2
<b>Overall SOA Appraisal and Disposition</b>											
SOA Grade (Range 6-12)		7				Disposition (select)				<b>Accepted, &lt; 12</b> Excluded, 12	
<b>Relevant SOA Results</b>											
SOA data		CoA: <ul style="list-style-type: none"> <li>- CoA is repaired during the neonatal period and infancy by surgery. Beyond infancy, percutaneous treatment using either balloon angioplasty or stent implantation are more frequently employed to treat native or recurrent coarctation.</li> <li>- The Cheatham-Platinum (CP) Stent was developed by NuMED (Hopkinton, NY) specifically designed to treat aortic</li> </ul>									

<sup>2</sup> Forbes TJ, Kim DW, Du W, Turner DR, Holzer R, Amin Z, Hijazi Z, Ghasemi A, Rome JJ, Nykanen D, Zahn E, Cowley C, Hoyer M, Waight D, Gruenstein D, Javois A, Foerster S, Kreutzer J, Sullivan N, Khan A, Owada C, Hagler D, Lim S, Canter J, Zellers T; CCISC Investigators. Comparison of surgical, stent, and balloon angioplasty treatment of native coarctation of the aorta: an observational study by the CCISC (Congenital Cardiovascular Interventional Study Consortium). J Am Coll Cardiol. 2011 Dec 13;58(25):2664-74. doi: 10.1016/j.jacc.2011.08.053. PMID: 22152954.



## NuMED Summary of Safety and Clinical Performance SSCP – Stents – CoA & RVOT

		<p>coarctation. It has rounded edges to reduce the incidence of aortic wall injury and can be expanded to 24mm diameter.</p> <ul style="list-style-type: none"> <li>- Stent implantation, balloon angioplasty, and surgery are all treatment options for coarctation in patients beyond infancy.</li> <li>- Treated coarctation is associated with long-term morbidity irrespective of treatment strategy.</li> </ul> <p>COAST Trials:</p> <ul style="list-style-type: none"> <li>- The FDA pivotal trials COAST (Coarctation of the Aorta Stent Trial; 2007-2016) and COAST II (Covered Cheatham-Platinum Stents for Prevention or Treatment of the Aorta; 2010-2016) demonstrated safety and efficacy of the bare and Covered CP Stents when used to treat aortic coarctation. Short- and medium-term results have been previously reported (Meadows et al. (48) and Taggart et al. (49)).</li> <li>- The Covered CP Stent is a CP stent covered by a 0.28” sleeve of 0.005” thick expanded polytetrafluoroethylene tubing and was available to centers participating in the COAST trial for compassionate and emergency use for aortic wall injury occurring during aortic interventions.</li> <li>- COAST II included patients who received a Covered CP stent as an emergency or compassionate use during the initial COAST trial (legacy arm) and prospectively enrolled patients between 2010 and 2011.</li> <li>- COAST II included higher-risk groups, such as patients with aortic wall injuries and those with nearly atretic descending aorta of 3mm or less diameter.</li> </ul>
	<p>Comments</p>	<p>Hemodynamic Outcome:</p> <ul style="list-style-type: none"> <li>- Study corroborates the results from the largest multi-center study of stenting for coarctation from the Congenital Cardiovascular Interventional Study Consortium, which reported 23% systolic hypertension at 12 to 60 months of follow-up, 9% arm-leg blood pressure gradient <math>\geq 20</math> mm Hg, 23% need for antihypertensive medication and the presence of any of these 3 in 37%.<sup>3</sup></li> </ul> <p>Stent Fractures:</p> <ul style="list-style-type: none"> <li>- Previous studies of the bare metal CP stent documented stent fractures of 2% at 12 months, and 12% at 24 months (Meadows et al. (48)). While the design and metallic composition of the CP stent may contribute, stents fractures are not limited to CP stents.<sup>4</sup> Boe et al.<sup>5</sup> reported a 21% fracture rate for Palmaz Genesis XD stents when used for coarctation therapy in children &lt; 20Kg at a mean follow-up of 75 months.</li> <li>- It is unclear whether somatic growth can add additional force and loading conditions to the implanted stent, or whether participation in contact sports might impact the incidence of stent fractures.</li> <li>- Bare metal stents have a significantly higher fracture rate than covered CP stent. Possible explanations could be that the struts of a bare stent become more solidly embedded into the aortic wall, and that the expanded polytetrafluoroethylene covering more equally distributes the radial force to multiple struts or that it reduces the transmission of aortic pulsability to the struts.</li> </ul>

<sup>3</sup> Holzer R, Qureshi S, Ghasemi A, Vincent J, Sievert H, Gruenstein D, Weber H, Alday L, Peirone A, Zellers T, Cheatham J, Slack M, Rome J. Stenting of aortic coarctation: acute, intermediate, and long-term results of a prospective multi-institutional registry--Congenital Cardiovascular Interventional Study Consortium (CCISC). *Catheter Cardiovasc Interv.* 2010 Oct 1;76(4):553-63. doi: 10.1002/ccd.22587. PMID: 20882661.

<sup>4</sup> McElhinney DB, Marshall AC, Schievano S. Fracture of cardiovascular stents in patients with congenital heart disease: theoretical and empirical considerations. *Circ Cardiovasc Interv.* 2013 Oct 1;6(5):575-85. doi: 10.1161/CIRCINTERVENTIONS.113.000148. PMID: 24129934.

<sup>5</sup> Boe BB, Loccoh E, Stockmaster K, Holzer RJ, Cheatham SL, Cheatham JP, Armstrong A, Berman DP. Median and long-term outcomes of stent implantation for coarctation of the aorta in small patients (<20 kg). [Abstract presented at PICS 2019]. *J Struct Heart Dis.* 2018;4:140.



## NuMED Summary of Safety and Clinical Performance SSCP – Stents – CoA & RVOT

		<p>Reinterventions:</p> <ul style="list-style-type: none"> <li>- Previously reported data documented transcatheter reinterventions of about 5% by 24 months follow-up (Meadows et al. (48)).</li> <li>- There is no expert consensus defining when a reintervention should be performed.</li> <li>- Reinterventions in this patient population are not unexpected and do not represent a poor outcome.</li> </ul> <p>Aortic Wall Injury:</p> <ul style="list-style-type: none"> <li>- Aneurysms did not just occur in patients with bare metal stents, but equally in patients who had covered stents implanted. As such, the notion that covered stent implantation confers long-term protection from the development of aneurysm, may not be the case. Data are in contrast with Butera et al.<sup>6</sup> who did show a significant difference in the incidence of aneurysm formation when comparing patients bare versus covered stents, albeit in a much smaller cohort. Also, the median follow-up in that study was significantly longer for those with bare stents compared with covered stents (85 versus 35 months). This is important as the current study demonstrates that the majority of aneurysms were not identified until late follow-up.</li> <li>- Most aneurysms developed within the borders of the stent, including covered stents. One possible explanation is that pressure within the aorta distributes flow between the stent and the aortic wall, eventually leading to aneurysm formation. Another possibility is that the expanded polytetrafluoroethylene became damaged during initial implantation.</li> <li>- Current study did not investigate the benefit of a covered stent to reduce the risk of acute aortic wall injury during stent implantation because cases have not been randomly assigned and high-risk patients were excluded for bare stent implantation and received covered stents.</li> </ul>																																											
<p>21. Kasar et al. (2022)</p> <table border="1" style="width: 100%; border-collapse: collapse; margin-top: 10px;"> <thead> <tr> <th colspan="2">Contribution</th> </tr> </thead> <tbody> <tr> <td style="width: 50%;">S&amp;P</td> <td style="width: 50%;">x (S only)</td> </tr> <tr> <td>SOA</td> <td>x</td> </tr> </tbody> </table>	Contribution		S&P	x (S only)	SOA	x	<p><b>Safety &amp; Performance Appraisal</b></p> <table border="1" style="width: 100%; border-collapse: collapse;"> <thead> <tr> <th style="width: 25%;">Level of Evidence</th> <th style="width: 30%;">Study Method/Design</th> <th style="width: 30%;">Question Applied</th> <th colspan="5">Oxford LOE 2011</th> </tr> </thead> <tbody> <tr> <td></td> <td>Non-randomized retrospective analysis.</td> <td>The aim of this study was to follow the procedural results of stenting for CoA in young children and adolescents and evaluate the safety and effectiveness of this procedure for younger patients.</td> <td style="text-align: center;">1</td> <td style="text-align: center;">2</td> <td style="text-align: center;">3</td> <td style="text-align: center;">4</td> <td style="text-align: center;">5</td> </tr> </tbody> </table> <table border="1" style="width: 100%; border-collapse: collapse;"> <thead> <tr> <th style="width: 25%;">Suitability</th> <th style="width: 45%;">Relevant Data</th> <th colspan="3">Grading</th> </tr> </thead> <tbody> <tr> <td>Device</td> <td> <ul style="list-style-type: none"> <li>- CP covered stent was implanted in 32 (73%) patients and CP uncovered stent in 12 (27%) patients</li> <li>- Unknown whether mounted or manual</li> </ul> </td> <td style="text-align: center;"><b>D1</b></td> <td style="text-align: center;">D2</td> <td style="text-align: center;">D3</td> </tr> <tr> <td>Application</td> <td> <ul style="list-style-type: none"> <li>- Native CoA</li> </ul> </td> <td style="text-align: center;"><b>A1</b></td> <td style="text-align: center;">A2</td> <td style="text-align: center;">A3</td> </tr> <tr> <td>Patient</td> <td> <ul style="list-style-type: none"> <li>- 44 total patients</li> <li>- Patients aged &lt; 18 years who underwent stenting for native CoA between Aug 2010 and Nov 2017</li> <li>- Patients with re-coarctation were excluded</li> </ul> </td> <td style="text-align: center;"><b>P1</b></td> <td style="text-align: center;">P2</td> <td style="text-align: center;">P3</td> </tr> </tbody> </table>			Level of Evidence	Study Method/Design	Question Applied	Oxford LOE 2011						Non-randomized retrospective analysis.	The aim of this study was to follow the procedural results of stenting for CoA in young children and adolescents and evaluate the safety and effectiveness of this procedure for younger patients.	1	2	3	4	5	Suitability	Relevant Data	Grading			Device	<ul style="list-style-type: none"> <li>- CP covered stent was implanted in 32 (73%) patients and CP uncovered stent in 12 (27%) patients</li> <li>- Unknown whether mounted or manual</li> </ul>	<b>D1</b>	D2	D3	Application	<ul style="list-style-type: none"> <li>- Native CoA</li> </ul>	<b>A1</b>	A2	A3	Patient	<ul style="list-style-type: none"> <li>- 44 total patients</li> <li>- Patients aged &lt; 18 years who underwent stenting for native CoA between Aug 2010 and Nov 2017</li> <li>- Patients with re-coarctation were excluded</li> </ul>	<b>P1</b>	P2	P3
Contribution																																													
S&P	x (S only)																																												
SOA	x																																												
Level of Evidence	Study Method/Design	Question Applied	Oxford LOE 2011																																										
	Non-randomized retrospective analysis.	The aim of this study was to follow the procedural results of stenting for CoA in young children and adolescents and evaluate the safety and effectiveness of this procedure for younger patients.	1	2	3	4	5																																						
Suitability	Relevant Data	Grading																																											
Device	<ul style="list-style-type: none"> <li>- CP covered stent was implanted in 32 (73%) patients and CP uncovered stent in 12 (27%) patients</li> <li>- Unknown whether mounted or manual</li> </ul>	<b>D1</b>	D2	D3																																									
Application	<ul style="list-style-type: none"> <li>- Native CoA</li> </ul>	<b>A1</b>	A2	A3																																									
Patient	<ul style="list-style-type: none"> <li>- 44 total patients</li> <li>- Patients aged &lt; 18 years who underwent stenting for native CoA between Aug 2010 and Nov 2017</li> <li>- Patients with re-coarctation were excluded</li> </ul>	<b>P1</b>	P2	P3																																									

<sup>6</sup> Butera G, Manica JL, Marini D, Piazza L, Chessa M, Filho RI, Sarmento Leite RE, Carminati M. From bare to covered: 15-year single center experience and follow-up in transcatheter stent implantation for aortic coarctation. Catheter Cardiovasc Interv. 2014 May 1;83(6):953-63. doi: 10.1002/ccd.25404. Epub 2014 Feb 4. PMID: 24459104.



## NuMED Summary of Safety and Clinical Performance SSCP – Stents – CoA & RVOT

	<ul style="list-style-type: none"> <li>- Mean age 10.4 years (4.3-18)</li> <li>- 77.2% male</li> <li>- Group I: 10-18 years, adolescent group and Group II: ≤ 9.9 years, pediatric group</li> </ul>				
Report	- Suitable for review	<b>R1</b>	R2	R3	
Suitability Grade (Range 4-12)		4			
<b>Data Contribution</b>	<b>Relevant Data</b>	<b>Grading</b>			
Outcomes/Endpoints	- Complications.	<b>Yes 1</b>	No 2		
Follow-up	- Mean follow-up was 23 months (range 2-84 months).	<b>Yes 1</b>	No 2		
Statistical analysis	- Categorical variables were stated as number (n) and percentage (%), and continuous variables as mean ± standard deviation, median, minimum and maximum values. Categorical data were compared using the Chi-square test or Fisher’s Exact test when expected frequencies were <5. In the comparisons of quantitative variables between the 2 groups, the Student’s <i>t</i> test, -test for paired data and the Mann-Whitney <i>U</i> -test were used. A 2-tailed <i>P</i> value <0.05 was considered statistically significant.	<b>Yes 1</b>	No 2		
Clinical significance	- The magnitude of the treatment effect observed was clinically significant.	<b>Yes 1</b>	No 2		
Data Contribution Grade (Range 4-8)		4			
<b>Overall S&amp;P Appraisal, Disposition and Weighting</b>					
S&P Grade (Range 9-25)	LOE (3) + Suitability (4) + Data Contribution (4) = 11	Disposition and Weighting (select)	<b>Accepted and Pivotal 9-12</b> Accepted but not Pivotal, 13-21 Excluded, 22-25		
<b>Relevant S&amp;P Results</b>					
Safety data	Complications: <ul style="list-style-type: none"> <li>- Complications developed in three patients, and all in Group I. Femoral hematoma developed in one patient, balloon rupture occurred during the procedure in one patient, and there was temporary loss of pulse in one patient. All the complications were treated successfully.</li> <li>- In the second case, a 16-year-old girl, the femoral artery pulse was lost. After 24 hours of heparin infusion, the pulse was viewed on color Doppler USG, the heparin treatment was ceased, and the patient was discharged.</li> </ul>				
Benefits/claims data	- Stent implantation for aortic coarctation in the pediatric age group may provide pleasing results, reducing the coarctation gradient, providing effective dilatation in the lesion area and eliminating hypertension.				
Strengths	- N/A				
Weaknesses/ Potential bias	- ... current stent technology, which requires a larger sheath, raises concerns in the very young age group.				
<b>State of the Art Appraisal</b>					
Medical condition	Alternatives	Risk/benefit	Side-effects	Equivalence	Surrogate endpoints
<b>Yes 1</b>	No 2	<b>Yes 1</b>	No 2	<b>Yes 1</b>	<b>No 2</b>
<b>Yes 1</b>	No 2	<b>Yes 1</b>	No 2	Yes 1	<b>No 2</b>



## NuMED Summary of Safety and Clinical Performance SSCP – Stents – CoA & RVOT

	<p><b>Overall SOA Appraisal and Disposition</b></p> <table border="1" style="width: 100%; border-collapse: collapse;"> <tr> <td style="width: 25%;">SOA Grade (Range 6-12)</td> <td style="width: 30%;">8</td> <td style="width: 30%;">Disposition (select)</td> <td colspan="2" style="width: 15%;">Accepted, &lt; 12 Excluded, 12</td> </tr> </table>					SOA Grade (Range 6-12)	8	Disposition (select)	Accepted, < 12 Excluded, 12																	
SOA Grade (Range 6-12)	8	Disposition (select)	Accepted, < 12 Excluded, 12																							
	<p><b>Relevant SOA Results</b></p> <table border="1" style="width: 100%; border-collapse: collapse;"> <tr> <td style="width: 25%;">SOA data</td> <td colspan="4"> <p>CoA:</p> <ul style="list-style-type: none"> <li>- CoA has been reported in 0.04% of all live births and in 5 to 8% of all patients with CHD.</li> <li>- Surgical repair is the gold standard treatment for CoA in infants and young children.</li> <li>- The aim of this study was to compare and evaluate the data of young pediatric and adolescent patients who underwent balloon-expandable stent implantation because of native aortic coarctation.</li> <li>- The patients were separated into two groups as the adolescent group (Group I: 10-18 years) and the pediatric group (Group II: ≤ 9.9 years). Group-I comprised of 18 patients and Group-II, 32 patients. Covered stent was implanted to 32 (73%) patients and uncovered stent to 12 (27%) patients.</li> <li>- The procedural success rate was 100%. Following stent implantation, peak systolic gradient decreased significantly in both groups (<math>P &lt; .0001</math>) (Group-I: from <math>35.9 \pm 16.6</math> mm Hg-<math>2.2 \pm 3.4</math> mm Hg, Group II: from <math>34 \pm 13.3</math> mm Hg-<math>3 \pm 4.09</math> mm Hg).</li> <li>- Complications developed in three patients, and all in Group I. Femoral hematoma developed in one patient, balloon rupture occurred during the procedure in one patient, and there was temporary loss of pulse in one patient. All the complications were treated successfully.</li> <li>- All the patients were taking anti-hypertensive drugs before intervention and during the mean 23-month follow-up period (range, 2-84 months), hypertension recovered in 35 (80%) patients and drugs were terminated.</li> <li>- Stent implantation for aortic coarctation in the pediatric age group may provide pleasing results, reducing the coarctation gradient, providing effective dilatation in the lesion area and eliminating hypertension.</li> </ul> </td> </tr> <tr> <td>Comments</td> <td colspan="4"> <ul style="list-style-type: none"> <li>- Limitations: Limitations include retrospective design and small population.</li> <li>- Authors have no funding and conflicts of interest to disclose.</li> </ul> </td> </tr> </table>					SOA data	<p>CoA:</p> <ul style="list-style-type: none"> <li>- CoA has been reported in 0.04% of all live births and in 5 to 8% of all patients with CHD.</li> <li>- Surgical repair is the gold standard treatment for CoA in infants and young children.</li> <li>- The aim of this study was to compare and evaluate the data of young pediatric and adolescent patients who underwent balloon-expandable stent implantation because of native aortic coarctation.</li> <li>- The patients were separated into two groups as the adolescent group (Group I: 10-18 years) and the pediatric group (Group II: ≤ 9.9 years). Group-I comprised of 18 patients and Group-II, 32 patients. Covered stent was implanted to 32 (73%) patients and uncovered stent to 12 (27%) patients.</li> <li>- The procedural success rate was 100%. Following stent implantation, peak systolic gradient decreased significantly in both groups (<math>P &lt; .0001</math>) (Group-I: from <math>35.9 \pm 16.6</math> mm Hg-<math>2.2 \pm 3.4</math> mm Hg, Group II: from <math>34 \pm 13.3</math> mm Hg-<math>3 \pm 4.09</math> mm Hg).</li> <li>- Complications developed in three patients, and all in Group I. Femoral hematoma developed in one patient, balloon rupture occurred during the procedure in one patient, and there was temporary loss of pulse in one patient. All the complications were treated successfully.</li> <li>- All the patients were taking anti-hypertensive drugs before intervention and during the mean 23-month follow-up period (range, 2-84 months), hypertension recovered in 35 (80%) patients and drugs were terminated.</li> <li>- Stent implantation for aortic coarctation in the pediatric age group may provide pleasing results, reducing the coarctation gradient, providing effective dilatation in the lesion area and eliminating hypertension.</li> </ul>				Comments	<ul style="list-style-type: none"> <li>- Limitations: Limitations include retrospective design and small population.</li> <li>- Authors have no funding and conflicts of interest to disclose.</li> </ul>														
SOA data	<p>CoA:</p> <ul style="list-style-type: none"> <li>- CoA has been reported in 0.04% of all live births and in 5 to 8% of all patients with CHD.</li> <li>- Surgical repair is the gold standard treatment for CoA in infants and young children.</li> <li>- The aim of this study was to compare and evaluate the data of young pediatric and adolescent patients who underwent balloon-expandable stent implantation because of native aortic coarctation.</li> <li>- The patients were separated into two groups as the adolescent group (Group I: 10-18 years) and the pediatric group (Group II: ≤ 9.9 years). Group-I comprised of 18 patients and Group-II, 32 patients. Covered stent was implanted to 32 (73%) patients and uncovered stent to 12 (27%) patients.</li> <li>- The procedural success rate was 100%. Following stent implantation, peak systolic gradient decreased significantly in both groups (<math>P &lt; .0001</math>) (Group-I: from <math>35.9 \pm 16.6</math> mm Hg-<math>2.2 \pm 3.4</math> mm Hg, Group II: from <math>34 \pm 13.3</math> mm Hg-<math>3 \pm 4.09</math> mm Hg).</li> <li>- Complications developed in three patients, and all in Group I. Femoral hematoma developed in one patient, balloon rupture occurred during the procedure in one patient, and there was temporary loss of pulse in one patient. All the complications were treated successfully.</li> <li>- All the patients were taking anti-hypertensive drugs before intervention and during the mean 23-month follow-up period (range, 2-84 months), hypertension recovered in 35 (80%) patients and drugs were terminated.</li> <li>- Stent implantation for aortic coarctation in the pediatric age group may provide pleasing results, reducing the coarctation gradient, providing effective dilatation in the lesion area and eliminating hypertension.</li> </ul>																									
Comments	<ul style="list-style-type: none"> <li>- Limitations: Limitations include retrospective design and small population.</li> <li>- Authors have no funding and conflicts of interest to disclose.</li> </ul>																									
	<p><b>Safety &amp; Performance</b></p> <p><b>Appraisal</b></p> <table border="1" style="width: 100%; border-collapse: collapse;"> <tr> <th style="width: 25%;">Level of Evidence</th> <th style="width: 25%;">Study Method/Design</th> <th style="width: 30%;">Question Applied</th> <th colspan="5" style="width: 15%;">Oxford LOE 2011</th> </tr> <tr> <td></td> <td>Single center prospective study.</td> <td>The aim of this study was to evaluate the immediate and intermediate results of CP stenting for native CoA.</td> <td style="text-align: center;">1</td> <td style="text-align: center;">2</td> <td style="text-align: center;">3</td> <td style="text-align: center;">4</td> <td style="text-align: center;">5</td> </tr> </table>					Level of Evidence	Study Method/Design	Question Applied	Oxford LOE 2011						Single center prospective study.	The aim of this study was to evaluate the immediate and intermediate results of CP stenting for native CoA.	1	2	3	4	5					
Level of Evidence	Study Method/Design	Question Applied	Oxford LOE 2011																							
	Single center prospective study.	The aim of this study was to evaluate the immediate and intermediate results of CP stenting for native CoA.	1	2	3	4	5																			
<p>22. Bairam et al. (2021)</p> <table border="1" style="width: 100%; border-collapse: collapse;"> <tr> <th colspan="2">Contribution</th> </tr> <tr> <td style="width: 50%;">S&amp;P</td> <td style="text-align: center;">x</td> </tr> <tr> <td>SOA</td> <td style="text-align: center;">x</td> </tr> </table>	Contribution		S&P	x	SOA	x	<table border="1" style="width: 100%; border-collapse: collapse;"> <tr> <th style="width: 25%;">Suitability</th> <th style="width: 50%;">Relevant Data</th> <th colspan="3" style="width: 25%;">Grading</th> </tr> <tr> <td>Device</td> <td> <ul style="list-style-type: none"> <li>- Covered CP stent was hand crimped on a balloon dilation catheter (either Z-med balloon dilation catheter or BIB balloon according to availability in the stock)</li> <li>- A total of 39 stents implanted; the CP stent used in 30 patients, and uncovered stent was used in one patient to avoid occlusion of the left subclavian artery</li> </ul> </td> <td style="text-align: center;"><b>D1</b></td> <td style="text-align: center;">D2</td> <td style="text-align: center;">D3</td> </tr> <tr> <td>Application</td> <td>- Native CoA</td> <td style="text-align: center;"><b>A1</b></td> <td style="text-align: center;">A2</td> <td style="text-align: center;">A3</td> </tr> </table>					Suitability	Relevant Data	Grading			Device	<ul style="list-style-type: none"> <li>- Covered CP stent was hand crimped on a balloon dilation catheter (either Z-med balloon dilation catheter or BIB balloon according to availability in the stock)</li> <li>- A total of 39 stents implanted; the CP stent used in 30 patients, and uncovered stent was used in one patient to avoid occlusion of the left subclavian artery</li> </ul>	<b>D1</b>	D2	D3	Application	- Native CoA	<b>A1</b>	A2	A3
Contribution																										
S&P	x																									
SOA	x																									
Suitability	Relevant Data	Grading																								
Device	<ul style="list-style-type: none"> <li>- Covered CP stent was hand crimped on a balloon dilation catheter (either Z-med balloon dilation catheter or BIB balloon according to availability in the stock)</li> <li>- A total of 39 stents implanted; the CP stent used in 30 patients, and uncovered stent was used in one patient to avoid occlusion of the left subclavian artery</li> </ul>	<b>D1</b>	D2	D3																						
Application	- Native CoA	<b>A1</b>	A2	A3																						





## NuMED Summary of Safety and Clinical Performance SSCP – Stents – CoA & RVOT

<b>Patient</b>	<ul style="list-style-type: none"> <li>- 32 consecutive adult patients with native CoA who underwent aortic stent replacement</li> <li>- Referred with clinical evidence for CoA (hypertension, arm-leg blood pressure difference <math>\geq 20</math> mmHg), for whom the diagnosis of CoA was confirmed by both transthoracic echocardiography and by CT angiography</li> <li>- Mean age <math>30.83 \pm 11.179</math> years (16-56)</li> <li>- 15 (46.9%) male and 17 (53.1%) female</li> </ul>	<b>P1</b>	P2	P3
<b>Report</b>	<ul style="list-style-type: none"> <li>- Publication reports the CP stent was implanted in 30 patients and uncovered stent in one patient but does not report the total of each stent design implanted.</li> </ul>	R1	<b>R2</b>	R3
Suitability Grade (Range 4-12)		5		
<b>Data Contribution</b>	<b>Relevant Data</b>	<b>Grading</b>		
<b>Outcomes/Endpoints</b>	<ul style="list-style-type: none"> <li>- Short-term pre/post-implant hemodynamics and angiographic data were reported. Clinical, echocardiographic and CT exam for restenosis, aneurysm formation, stent migration, changes in blood pressure, and the use of antihypertensive drugs were recorded during follow-up.</li> </ul>	<b>Yes 1</b>	No 2	
<b>Follow-up</b>	<ul style="list-style-type: none"> <li>- Ranged from 6-32 months (mean 14.9 months) included clinical, echocardiographic and CT exam assessing for restenosis, aneurysm formation, stent migration, blood pressure control and the intensity of antihypertensive medications</li> </ul>	<b>Yes 1</b>	No 2	
<b>Statistical analysis</b>	<ul style="list-style-type: none"> <li>- Mean, standard deviation and P-values reported</li> </ul>	<b>Yes 1</b>	No 2	
<b>Clinical significance</b>	<ul style="list-style-type: none"> <li>- The magnitude of the treatment effect observed was clinically significant.</li> </ul>	<b>Yes 1</b>	No 2	
Data Contribution Grade (Range 4-8)		4		
<b>Overall S&amp;P Appraisal, Disposition and Weighting</b>				
<b>S&amp;P Grade (Range 9-25)</b>	LOE (3) + Suitability (5) + Data Contribution (4) = 12	<b>Disposition and Weighting (select)</b>	<b>Accepted and Pivotal 9-12</b> Accepted but not Pivotal, 13-21 Excluded, 22-25	
<b>Relevant S&amp;P Results</b>				
<b>Safety data</b>	<ul style="list-style-type: none"> <li>- There was no major complications, with no deaths.</li> <li>- There was no recorded complication at the implantation site except for one case of ruptured balloon used to deliver a second CP stent, the balloon retrieved to the iliac artery then successfully removed surgically.</li> <li>- No significant complications were seen during procedure and at six months follow up.</li> </ul>			
<b>Performance data</b>	<ul style="list-style-type: none"> <li>- Procedure success rate was 93.4% (n=31). The technique was considered effective if the invasive grade was decrease to <math>&lt; 20</math> mmHg and increased the angiographic diameter <math>&gt; 50\%</math>.</li> <li>- Peak gradient across the coarctation fell from <math>60.0 \pm 21.960</math> to <math>10.0 \pm 19.821</math> mm Hg post procedure (P = 0.0001).</li> <li>- Systolic blood pressures fell from <math>164.6 \pm 25.889</math> mm Hg to <math>138.1 \pm 17.006</math> mm Hg immediately after stenting and <math>134.3 \pm 12</math> mm Hg at six months.</li> </ul>			
<b>Benefits/claims data</b>	<ul style="list-style-type: none"> <li>- Treating native CoA with stent implantation is a less invasive, safe and highly successful technique with excellent angiographic and clinical response that is sustained to the midterm follow up of this study.</li> </ul>			
<b>Strengths</b>	<ul style="list-style-type: none"> <li>- Prospective design</li> <li>- Consecutive patients</li> </ul>			
<b>Weaknesses/</b>	<ul style="list-style-type: none"> <li>- Single center design</li> </ul>			



## NuMED Summary of Safety and Clinical Performance SSCP – Stents – CoA & RVOT

	Potential bias	- Follow up ranged from 6-32 months (mean 14.9 Months) although study intended as immediate and intermediate results																	
<b>State of the Art</b>																			
<b>Appraisal</b>																			
Medical condition		Alternatives		Risk/benefit		Side-effects		Equivalence		Surrogate endpoints									
Yes 1	No 2	Yes 1	No 2	Yes 1	No 2	Yes 1	No 2	Yes 1	No 2	Yes 1	No 2								
<b>Overall SOA Appraisal and Disposition</b>																			
SOA Grade (Range 6-12)		8				Disposition (select)				Accepted, < 12 Excluded, 12									
<b>Relevant SOA Results</b>																			
SOA data		<p>Balloon angioplasty:</p> <ul style="list-style-type: none"> <li>- Balloon angioplasty suggested as the favored management for adults and children they have coarctation or recoarctation afterward operation. The main problem of angioplasty only is shrinking of wall of vessel with stenosis reappearance, separation of aortic wall occur in 1–4% of total patients, and creation of aneurysm in 4–11%. Next balloon enlargement, about 21-37% continue hypertensive.</li> </ul> <p>CoA:</p> <ul style="list-style-type: none"> <li>- CoA is well-defined as a separate stenosis of the proximal thoracic aorta frequently at the insertion point of the ductus arteriosus, it accounts for about 6-8% of patients with congenital heart disease with a male/ female ratio in a range from 1.4:1 to 3:1. Eighty five percent of patient with COA have an associated bicuspid aortic valve.</li> <li>- The chief goals with any management for CoA is continued hemodynamic advantage as supported by control of hypertension.</li> </ul>																	
Comments		- No author-identified limitations. Funding and conflicts of interest not addressed.																	
<b>Safety &amp; Performance</b>																			
<b>Appraisal</b>																			
23. Alsamarra et al. (2023)		Level of Evidence		Study Method/Design			Question Applied			Oxford LOE 2011									
				Single center prospective study.			The aim of the study is to evaluate the immediate results & the complications of balloon angioplasty with covered CP-stent for coarctation of descending aorta.			1	2	3	4	5					
<table border="1" style="width: 100%; border-collapse: collapse;"> <tr> <th colspan="2">Contribution</th> </tr> <tr> <td>S&amp;P</td> <td style="text-align: center;">x</td> </tr> <tr> <td>CoA</td> <td style="text-align: center;">x</td> </tr> </table>		Contribution		S&P	x	CoA	x	Suitability		Relevant Data						Grading			
		Contribution																	
S&P	x																		
CoA	x																		
		Device		<ul style="list-style-type: none"> <li>- 33 covered CP stents, 24 used in Group 1 and 9 in Group 2 (P=0.7)</li> <li>- Not specified if stent was pre-mounted or not</li> <li>- Mean stent length was 32.4 ± 7.8 mm (22-52) with no significant difference between groups (P=0.8)</li> <li>- Balloon size varied significantly between groups, being larger in Group 1 (P=0.002)</li> </ul>						D1	D2	D3							



## NuMED Summary of Safety and Clinical Performance SSCP – Stents – CoA & RVOT

Application	- Native CoA (n=20) and post-angioplasty recoarctation (n=7)	<b>A1</b>	A2	A3
Patient	- 27 patients treated using 33 covered CP stents - Patients divided into two groups: Group 1 (n=20) with native CoA and Group 2 (n=7) with recoarctation after prior balloon angioplasty - Mean age 25.22 ± 12 years (11-56), Group 1: mean age 28.4 ± 12.3 years (13-56), Group 2: mean age 16.14 ± 3.4 years (11-21), P = 0.001 - 15 (55.6%) male and 12 (44.4%) female, Group 1: 10 male (50%) and 10 female (50%), Group 2: 5 male (71.42%) and 2 female (28.57%)	<b>P1</b>	P2	P3
Report	- Publication presents as a summary.	R1	<b>R2</b>	R3
Suitability Grade (Range 4-12)		5		

Data Contribution	Relevant Data	Grading	
Outcomes/Endpoints	- Immediate pressure gradient, aortic diameter and complication data were reported.	<b>Yes 1</b>	No 2
Follow-up	- Immediate results, no follow-up period	Yes 1	<b>No 2</b>
Statistical analysis	- Mean, standard deviation and P-values reported.	<b>Yes 1</b>	No 2
Clinical significance	- The magnitude of the treatment effect observed was clinically significant.	<b>Yes 1</b>	No 2
Data Contribution Grade (Range 4-8)		5	

### Overall S&P Appraisal, Disposition and Weighting

S&P Grade (Range 9-25)	LOE (3) + Suitability (5) + Data Contribution (5) = 13	Disposition and Weighting (select)	Accepted and Pivotal 9-12 <b>Accepted but not Pivotal, 13-21</b> Excluded, 22-25
---------------------------	---	------------------------------------	--

### Relevant S&P Results

Safety data	- No deaths related to the procedure were reported. - Four minor complications (14.8%) were noted, predominantly in Group 2. Two involved stent migration, and two involved femoral pulse loss, which resolved with heparin treatment. Group 2 had a significantly higher complication rate. - Two instances of stent migration were observed with the use of Z-MED balloon catheters, and two other cases noted a loss of femoral pulse.
Performance data	- The overall success rate was 96.3%, with 100% in Group 1 and 85.7% in Group 2. - Invasive gradient decreased from 62.2 ± 21.5 mm Hg (20-100) to 9.6 ± 14.4 mm Hg (0-60) (P=0.001). - Aortic diameter increased from 6.5 ± 1.6 mm (4-10) to 13.7 ± 3.6 mm (6-20) (P = 0.001). - Echocardiographic gradient fell from 60.8 ± 19.3 mm Hg (25-100) to 29.5 ± 11.6 mm Hg (18-70) (P=0.001).
Benefits/claims data	- Effective gradient alleviation and rising lesion sizes in native and recurrent coarctation.
Strengths	- Prospective design
Weaknesses/ Potential bias	- Single center design - Immediate results - Two instances of stent migration with use of Z-MED balloon catheters in Group 2 (recoarctation after primary balloon angioplasty)



## NuMED Summary of Safety and Clinical Performance SSCP – Stents – CoA & RVOT

	<p><b>State of the Art Appraisal</b></p> <table border="1" style="width: 100%; border-collapse: collapse;"> <thead> <tr> <th colspan="2">Medical condition</th> <th colspan="2">Alternatives</th> <th colspan="2">Risk/benefit</th> <th colspan="2">Side-effects</th> <th colspan="2">Equivalence</th> <th colspan="2">Surrogate endpoints</th> </tr> <tr> <th>Yes 1</th> <th>No 2</th> <th>Yes 1</th> <th>No 2</th> <th>Yes 1</th> <th>No 2</th> <th>Yes 1</th> <th>No 2</th> <th>Yes 1</th> <th>No 2</th> <th>Yes 1</th> <th>No 2</th> </tr> </thead> <tbody> <tr> <td></td><td></td><td></td><td></td><td></td><td></td><td></td><td></td><td></td><td></td><td></td><td></td> </tr> </tbody> </table> <p><b>Overall SOA Appraisal and Disposition</b></p> <table border="1" style="width: 100%; border-collapse: collapse;"> <tr> <td style="width: 25%;">SOA Grade (Range 6-12)</td> <td style="width: 30%;">9</td> <td style="width: 30%;">Disposition (select)</td> <td style="width: 15%;"><b>Accepted, &lt; 12</b> Excluded, 12</td> </tr> </table> <p><b>Relevant SOA Results</b></p> <table border="1" style="width: 100%; border-collapse: collapse;"> <tr> <td style="width: 25%;">SOA data</td> <td> <p>Balloon angioplasty:</p> <ul style="list-style-type: none"> <li>- In 1982, balloon dilation emerged as a non-surgical remedy. Balloon angioplasty became a widely accepted solution for both initial and recurrent coarctations in patients aged over six months. However, the technique isn't without issues. There's potential for vessel recoil and lesion persistence, with over-extension risking aortic wall damage.</li> </ul> <p>CoA:</p> <ul style="list-style-type: none"> <li>- O' Laughlin et al introduced the endovascular stent's application for aortic coarctation in 1991. Initially reserved for cases resistant to surgery and balloon angioplasty, stents have, with time, been recognized as the preferred treatment in certain complex anatomies, coarctation with patent ductus arteriosus, and CoA associated with aneurysms.</li> </ul> </td> </tr> <tr> <td>Comments</td> <td> <ul style="list-style-type: none"> <li>- No author-identified limitations.</li> <li>- Funding and conflicts of interest not addressed.</li> </ul> </td> </tr> </table>											Medical condition		Alternatives		Risk/benefit		Side-effects		Equivalence		Surrogate endpoints		Yes 1	No 2	Yes 1	No 2	Yes 1	No 2	Yes 1	No 2	Yes 1	No 2	Yes 1	No 2													SOA Grade (Range 6-12)	9	Disposition (select)	<b>Accepted, &lt; 12</b> Excluded, 12	SOA data	<p>Balloon angioplasty:</p> <ul style="list-style-type: none"> <li>- In 1982, balloon dilation emerged as a non-surgical remedy. Balloon angioplasty became a widely accepted solution for both initial and recurrent coarctations in patients aged over six months. However, the technique isn't without issues. There's potential for vessel recoil and lesion persistence, with over-extension risking aortic wall damage.</li> </ul> <p>CoA:</p> <ul style="list-style-type: none"> <li>- O' Laughlin et al introduced the endovascular stent's application for aortic coarctation in 1991. Initially reserved for cases resistant to surgery and balloon angioplasty, stents have, with time, been recognized as the preferred treatment in certain complex anatomies, coarctation with patent ductus arteriosus, and CoA associated with aneurysms.</li> </ul>	Comments	<ul style="list-style-type: none"> <li>- No author-identified limitations.</li> <li>- Funding and conflicts of interest not addressed.</li> </ul>			
Medical condition		Alternatives		Risk/benefit		Side-effects		Equivalence		Surrogate endpoints																																																
Yes 1	No 2	Yes 1	No 2	Yes 1	No 2	Yes 1	No 2	Yes 1	No 2	Yes 1	No 2																																															
SOA Grade (Range 6-12)	9	Disposition (select)	<b>Accepted, &lt; 12</b> Excluded, 12																																																							
SOA data	<p>Balloon angioplasty:</p> <ul style="list-style-type: none"> <li>- In 1982, balloon dilation emerged as a non-surgical remedy. Balloon angioplasty became a widely accepted solution for both initial and recurrent coarctations in patients aged over six months. However, the technique isn't without issues. There's potential for vessel recoil and lesion persistence, with over-extension risking aortic wall damage.</li> </ul> <p>CoA:</p> <ul style="list-style-type: none"> <li>- O' Laughlin et al introduced the endovascular stent's application for aortic coarctation in 1991. Initially reserved for cases resistant to surgery and balloon angioplasty, stents have, with time, been recognized as the preferred treatment in certain complex anatomies, coarctation with patent ductus arteriosus, and CoA associated with aneurysms.</li> </ul>																																																									
Comments	<ul style="list-style-type: none"> <li>- No author-identified limitations.</li> <li>- Funding and conflicts of interest not addressed.</li> </ul>																																																									
<p>24. Schleiger et al. (2023)</p> <table border="1" style="width: 100%; border-collapse: collapse; margin-top: 10px;"> <thead> <tr> <th colspan="2">Contribution</th> </tr> </thead> <tbody> <tr> <td>S&amp;P</td> <td style="text-align: center;">x</td> </tr> <tr> <td>SOA</td> <td style="text-align: center;">x</td> </tr> </tbody> </table>	Contribution		S&P	x	SOA	x	<p><b>Safety &amp; Performance Appraisal</b></p> <table border="1" style="width: 100%; border-collapse: collapse;"> <thead> <tr> <th rowspan="2">Level of Evidence</th> <th rowspan="2">Study Method/Design</th> <th rowspan="2">Question Applied</th> <th colspan="5">Oxford LOE 2011</th> </tr> <tr> <th>1</th> <th>2</th> <th>3</th> <th>4</th> <th>5</th> </tr> </thead> <tbody> <tr> <td></td> <td>Single center retrospective study.</td> <td>The aim of this study was to analyze long-term results after CoA treatment with bare and covered CP stents in our institution and to derive recommendations for the differential use of these stent types.</td> <td></td> <td></td> <td></td> <td></td> <td></td> </tr> </tbody> </table> <table border="1" style="width: 100%; border-collapse: collapse; margin-top: 10px;"> <thead> <tr> <th>Suitability</th> <th>Relevant Data</th> <th colspan="3">Grading</th> </tr> </thead> <tbody> <tr> <td>Device</td> <td> <ul style="list-style-type: none"> <li>- 212 patients received treatment with bare (n=71) and covered (n=141) CP stents between September 1999 and July 2021</li> <li>- Stents were mounted on BIB catheters, not specified as pre-mounted</li> </ul> </td> <td style="text-align: center;"><b>D1</b></td> <td style="text-align: center;">D2</td> <td style="text-align: center;">D3</td> </tr> <tr> <td>Application</td> <td> <ul style="list-style-type: none"> <li>- Native CoA (n=110/212, 51.9%) and recoarctation after primary surgical or interventional treatment (n=102/212, 48.1%)</li> </ul> </td> <td style="text-align: center;"><b>A1</b></td> <td style="text-align: center;">A2</td> <td style="text-align: center;">A3</td> </tr> <tr> <td>Patient</td> <td> <ul style="list-style-type: none"> <li>- Median patient age was 18.8 years (IQR 11.9; 35.8)</li> <li>- Median patient weight 61.3 kg (IQR 43.3; 74.7)</li> <li>- 146/212 (68.9%) male</li> </ul> </td> <td style="text-align: center;"><b>P1</b></td> <td style="text-align: center;">P2</td> <td style="text-align: center;">P3</td> </tr> </tbody> </table>											Level of Evidence	Study Method/Design	Question Applied	Oxford LOE 2011					1	2	3	4	5		Single center retrospective study.	The aim of this study was to analyze long-term results after CoA treatment with bare and covered CP stents in our institution and to derive recommendations for the differential use of these stent types.						Suitability	Relevant Data	Grading			Device	<ul style="list-style-type: none"> <li>- 212 patients received treatment with bare (n=71) and covered (n=141) CP stents between September 1999 and July 2021</li> <li>- Stents were mounted on BIB catheters, not specified as pre-mounted</li> </ul>	<b>D1</b>	D2	D3	Application	<ul style="list-style-type: none"> <li>- Native CoA (n=110/212, 51.9%) and recoarctation after primary surgical or interventional treatment (n=102/212, 48.1%)</li> </ul>	<b>A1</b>	A2	A3	Patient	<ul style="list-style-type: none"> <li>- Median patient age was 18.8 years (IQR 11.9; 35.8)</li> <li>- Median patient weight 61.3 kg (IQR 43.3; 74.7)</li> <li>- 146/212 (68.9%) male</li> </ul>	<b>P1</b>	P2	P3
Contribution																																																										
S&P	x																																																									
SOA	x																																																									
Level of Evidence	Study Method/Design	Question Applied	Oxford LOE 2011																																																							
			1	2	3	4	5																																																			
	Single center retrospective study.	The aim of this study was to analyze long-term results after CoA treatment with bare and covered CP stents in our institution and to derive recommendations for the differential use of these stent types.																																																								
Suitability	Relevant Data	Grading																																																								
Device	<ul style="list-style-type: none"> <li>- 212 patients received treatment with bare (n=71) and covered (n=141) CP stents between September 1999 and July 2021</li> <li>- Stents were mounted on BIB catheters, not specified as pre-mounted</li> </ul>	<b>D1</b>	D2	D3																																																						
Application	<ul style="list-style-type: none"> <li>- Native CoA (n=110/212, 51.9%) and recoarctation after primary surgical or interventional treatment (n=102/212, 48.1%)</li> </ul>	<b>A1</b>	A2	A3																																																						
Patient	<ul style="list-style-type: none"> <li>- Median patient age was 18.8 years (IQR 11.9; 35.8)</li> <li>- Median patient weight 61.3 kg (IQR 43.3; 74.7)</li> <li>- 146/212 (68.9%) male</li> </ul>	<b>P1</b>	P2	P3																																																						



## NuMED

### Summary of Safety and Clinical Performance

#### SSCP – Stents – CoA & RVOT

<b>Report</b>	- High quality.	<b>R1</b>	<b>R2</b>	<b>R3</b>
Suitability Grade (Range 4-12)			4	
<b>Data Contribution</b>	<b>Relevant Data</b>	<b>Grading</b>		
<b>Outcomes/Endpoints</b>	- Procedural success, survival rate, freedom from re-intervention, peri-procedural and long-term complications were reported.	<b>Yes 1</b>	No 2	
<b>Follow-up</b>	- Median follow-up of 7.3 years (IQR: 4.3-12.6)	<b>Yes 1</b>	No 2	
<b>Statistical analysis</b>	- Patient characteristics expressed as median and IQR. Survival and freedom from re-intervention were assessed using Kaplan-Meier survival analysis. Survival and reintervention rates between groups were compared using the log rank test. Differences between groups were analyzed using the $\chi^2$ test for categorical variables and Wilcoxon rank sum test for continuous variables. Potential risk factors for re-intervention were evaluated with univariate logistic and Cox regression analysis. Time-independent variables were included in a multivariable model using HR. A $p < 0.05$ was considered statistically significant.	<b>Yes 1</b>	No 2	
<b>Clinical significance</b>	- The magnitude of the treatment effect observed was clinically significant.	<b>Yes 1</b>	No 2	
Data Contribution Grade (Range 4-8)			5	
<b>Overall S&amp;P Appraisal, Disposition and Weighting</b>				
<b>S&amp;P Grade</b> (Range 9-25)	LOE (3) + Suitability (4) + Data Contribution (4) = 11	Disposition and Weighting (select)	<b>Accepted and Pivotal 9-12</b> Accepted but not Pivotal, 13-21 Excluded, 22-25	
<b>Relevant S&amp;P Results</b>				
<b>Safety data</b>	<ul style="list-style-type: none"> <li>- Survival rate: Survival rate was 98.1% after five, and 95.6% after 10 and 15 years, respectively, and did not differ between patients who received bare or covered CP stents (Log Rank <math>p = 0.263</math>). In-hospital mortality occurred in 1/212 patients (0.5%) and late mortality in 8/158 patients (5.1%). Late mortality was not attributable to previous CoA treatment.. There was no difference in late mortality according to stent type (<math>p = 0.261</math>).</li> <li>- Complications rate:               <ul style="list-style-type: none"> <li>o Peri-procedural complications – Covered CP stent (n=141)                   <ul style="list-style-type: none"> <li>▪ Injury/thrombosis of casvular access vessel: 6/141 (4.3%)</li> <li>▪ Bleeding of vascular access vessel: 1/141 (0.7%)</li> <li>▪ Aortic dissection/aortic wall rupture: 1/141 (0.7%)</li> </ul> </li> <li>o Long-term complications – Covered CP stent                   <ul style="list-style-type: none"> <li>▪ Aneurysm formation: 9/83 (10.8%)</li> <li>▪ Stent fracture: 9/69 (13.0%)</li> <li>▪ Endoleak: 1/98 (1.0%)</li> </ul> </li> </ul> </li> </ul>			
<b>Performance data</b>	<ul style="list-style-type: none"> <li>- Procedural success was achieved in 187/212 (88.2%) patients.</li> <li>- After stent implantation a significant reduction of systolic blood pressure was achieved from a preinterventional median pressure of 145 mmHg (IQR 134; 157) to a postinterventional median pressure of 123 mmHg (IQR 112; 135) (<math>p &lt; 0.001</math>).</li> <li>- In 25 patients the interventional procedure was not considered successful: In 22 of these patients a reduction of the</li> </ul>			




## NuMED

### Summary of Safety and Clinical Performance

#### SSCP – Stents – CoA & RVOT

		<p>peak systolic pressure <math>\leq 10</math> mmHg was not achieved.</p> <ul style="list-style-type: none"> <li>○ Fifteen of these patients were additionally diagnosed with a hypoplastic aortic arch with a remaining systolic ascending to descending aortic pressure difference <math>&gt;10</math> mmHg after successful implantation of the CP stent in the CoA region.</li> <li>○ In two patients with an unsuccessful procedure the stent migrated into the descending aorta immediately after placement. Both patients underwent subsequent surgical repair after fixation of the stent in the descending aorta by balloon dilatation.</li> <li>○ Another patient developed an aortic wall rupture immediately after stent implantation and died during extracorporeal cardiopulmonary resuscitation.</li> </ul> <p>- Re-intervention rate: Planned re-interventions were performed in 33/158 patients (20.9%)... In 44/158 patients (27.8%), unplanned re-interventions were performed to treat re-stenosis or aortic wall injuries... The probability of freedom from re-intervention in the entire cohort was 81.0% after five, 64.0% after 10 and 62.0% after 15 years, respectively. Re-intervention rate did not differ between patients who received endovascular CoA treatment with bare or covered CP stents (<math>p = 0.50</math>)... Multivariable risk factor analysis revealed previous CoA surgery (HR: 2.0, 95% confidence interval [CI]: 1.1-3.9, <math>p = 0.029</math>), postdilatation (HR: 2.9, 95% CI: 1.1-6.3, <math>p = 0.028</math>) and age at intervention (HR: 0.96, 95% CI: 0.94-0.99, <math>p = 0.002</math>) as independent risk factors for re-intervention.</p> <p>- Antihypertensive medications: Residual arterial hypertension was present in 53/158 patients (33.5%); 23 of these patients were adult and 30 patients pediatric (<math>p = 0.173</math>). Before endovascular treatment with bare or covered CP stent 114 patients received no medical antihypertensive therapy, 44 patients received monotherapy, 30 patients dual therapy, and 24 patients triple or quadruple therapy. In 76/158 patients (48.1%) the number of antihypertensive medications did not change during follow-up, whereas in 57/158 patients (36.1%) the number of antihypertensive medication was increased and in 25/158 patients (15.8%) decreased.</p>							
	Benefits/claims data	- In conclusion, our study documents excellent long-term results after CoA treatment with bare and covered CP stents. In our cohort, mortality, re-intervention and complication rate did not significantly differ between both stent types.							
	Strengths	- Long-term follow-up (median of 7.3 years, IQR: 4.3-12.6)							
	Weaknesses/ Potential bias	- Single center design - Retrospective design							
<b>State of the Art</b>									
<b>Appraisal</b>									
Medical condition		Alternatives	Risk/benefit	Side-effects	Equivalence	Surrogate endpoints			
Yes 1	No 2	Yes 1	No 2	Yes 1	No 2	Yes 1	No 2	Yes 1	No 2
<b>Overall SOA Appraisal and Disposition</b>									
SOA Grade (Range 6-12)	8			Disposition (select)		Accepted, < 12 Excluded, 12			
<b>Relevant SOA Results</b>									



## NuMED

### Summary of Safety and Clinical Performance

#### SSCP – Stents – CoA & RVOT

	SOA data	<p>CoA:</p> <ul style="list-style-type: none"> <li>- CoA comprises 5% to 8% of all congenital heart disease.</li> <li>- The introduction of covered Chatham-Platinum (CP) stents in the late 1990s has substantially expanded the spectrum of endovascular CoA treatment including complex CoA anatomy with atretic or subatretic coarctation, severe aortic arch tortuosity, treatment of patients with CoA aneurysms or stent related complications such as fracture, endoleak or aortic wall injury.</li> </ul> <p>Complications rate – Bare CP Stent:</p> <ul style="list-style-type: none"> <li>- Peri-procedural complications – Bare CP stent (n=71)             <ul style="list-style-type: none"> <li>o Injury/thrombosis of vascular access vessel: 3/71 (4.2%)</li> <li>o Stent dislocation: 2/71 (2.8%)</li> <li>o Aortic dissection/aortic wall rupture: 2/71 (2.8%)</li> </ul> </li> <li>- Long-term complications – Bare CP stent             <ul style="list-style-type: none"> <li>o Aneurysm formation: 5/50 (10.0%)</li> <li>o Stent fracture: 10/39 (25.6%)</li> </ul> </li> </ul>
	Comments	<ul style="list-style-type: none"> <li>- Author-identified limitations included unequal distribution between subgroups with more patients receiving Covered CP stents due to institutional preference, difference in follow-up duration between sub-groups due to differences in time points of availability and product approval, 54 patients lost to follow-up could not be considered in long-term analysis, the low event rate of major complications (e.g., aortic dissection, stent fracture, aneurysm formation) may limit statistical comparison between groups, incidence of long-term aortic wall complications may be underestimated as re-catheterization or cross-sectional imaging was not available for all patients, non-invasive blood measurement during exercise or 24-hours blood pressure measurements were not available to identify unmasked arterial hypertension and the number of antihypertensive medications may be affected by cofounders.</li> <li>- Authors declare no conflict of interest. Open access funding enabled and organized by Projekt DEAL.</li> </ul>



# NuMED

## Summary of Safety and Clinical Performance

### SSCP – Stents – CoA & RVOT

#### **An overall summary of the clinical performance and safety:**

A comprehensive, systematic, and critical evaluation of the pertinent clinical data and pre-clinical study data in relation to the Covered Stents has been carried out and documented in this report. Based on the results of this evaluation, it is considered that:

- a) Conformity with relevant general safety and performance requirements set out in MDR Annex I under the normal conditions of the intended use of the device has been confirmed.
- b) Undesirable side-effects and acceptability of the benefit-risk ratio have been evaluated and are acceptable according to the current knowledge/the state of the art in the medical fields concerned and according to available medical alternatives.
- c) The information materials supplied by NuMED, and the risk reduction measures are adequate taking into account the intended purpose of the device.
- d) Usability aspects have been adequately considered and the Covered Stents including the IFUs, are suitable for the intended users.
- e) The claims foreseen in the information materials provided with the Clinical Evaluation Report (CER) are adequate taking into account the intended purpose of the device.
- f) The information materials supplied and the RM documentation for the device under evaluation are consistent with the clinical data and pre-clinical study data presented in this CER and with the current knowledge/state of the art.

Overall, it is concluded that the risks associated with the use of the Covered Stents are acceptable when weighed against the benefits to the patient and are compatible with a high level of protection of health and safety, taking into account the generally acknowledged state of the art; that the intended clinical performances are achieved by the device; and that known and foreseeable risks and undesirable side-effects are considered acceptable when weighed against the benefits from performance achieved by the device.

#### **Ongoing planned post-market clinical follow-up:**

The Stent Device Family has been on the market since 2004 in the EU and 1999 in other markets. Over time variants of the Stent Device Family have been introduced to these markets. Since then, the devices are likely to have been used in a variety of patients and populations. The Stents have been subjected to several clinical investigations where efficacy and safety has been demonstrated.

For the original Stent Device Family, a PMCF study is not warranted at this time due to the fact that the long-term safety and clinical performance has been established via device use and ample clinical experience. This experience would likely have identified any rare complications or problems that would become apparent only after widespread device use. Continued PMS activities will provide sufficient data to adequately address clinical risks, and detect emerging risks on the basis of evidence. Additional clinical studies were conducted in the US under the COAST, COAST II, and PARCS clinical trials.

A PMCF study was initiated in 2018 for the additional sizes that were added to the product line, to determine if there were any new complications which were previously not addressed through actual clinical use, or if any new risks are introduced. The study had a target size of 59 patients, based on a confidence level of 95%. The study was conducted by issuing a form to the treating physician and collecting data. The results of the 2018 study are included in the clinical data that is used for the clinical evaluation.

#### **6. Possible diagnostic or therapeutic alternatives**

Alternative treatments for CoA include surgery or balloon angioplasty.

Alternative treatments for RVOT include surgery, transcatheter pulmonary valve replacement, or balloon valvuloplasty / angioplasty (to delay the need for replacement only).

#### **7. Suggested profile and training for users**

The RVOT/COA Stent Device Family is intended for use by trained cardiology and surgical professionals undertaking stent implantation.

#### **8. Reference to any harmonised standards and CS applied**

There are no Common Specifications for this type of device.

The following harmonised standards are followed for this device:

- EN ISO 10993-10: 2023 – Biological Evaluation of Medical Devices – Part 10: Tests for Skin Sensitization
- EN ISO 10993-18: 2020 – Biological Evaluation of Medical Devices – Part 18: Chemical characterization of medical device materials within a risk management process





## NuMED

### Summary of Safety and Clinical Performance

### SSCP – Stents – CoA & RVOT

- EN ISO 10993-23: 2021 – Biological Evaluation of Medical Devices – Part 23: Tests for Irritation
- EN ISO 11135: 2014 / A1:2019 – Sterilization of health-care products – Ethylene oxide – Requirements for the development, validation and routine control of a sterilization process for medical devices.
- BS EN ISO 11607-1: 2020 +A1: 2023 – Packaging for Terminally Sterilized Medical Devices – Part 1: Requirements for materials, sterile barriers systems and packaging systems
- BS EN ISO 11607-2: 2020 +A1: 2023 – Packaging for Terminally Sterilized Medical Devices – Part 2: Validation requirements for forming, sealing and assembly processes
- EN ISO 11737-1: 2018 / A1:2021 – Sterilization of medical devices – Microbiological methods – Part 1: Determination of a population of microorganisms on products
- EN ISO 13485: 2016 / A11:2021 – Medical devices – Quality management systems – Requirements for regulatory purposes
- EN ISO 14971: 2019 / A11:2021 – Medical Devices – Application of Risk Management to Medical Devices
- EN ISO 15223-1: 2021 – Medical devices – Symbols to be used with medical device labels, labelling and information to be supplied – Part 1: General requirements

#### 9. References

1. Delaney JW, Goldstein BH, Bishnoi RN, Bisselou KSM, McEnaney K, Minahan M, Ringel RE; PARCS Investigators. Covered CP stent for treatment of right ventricular conduit injury during melody transcatheter pulmonary valve replacement, *Circ Cardiovasc Interv* **11**(10), 1-10 (2018).
2. Baykan A, Demiral AG, Tasci O, Pamukcu O, Sunkak S, Uzum K, Sezer S, Narin N, Is hypertension the fate of aortic coarctation patients treated with Cheatham Platinum (CP) stent? *Journal of Interventional Cardiology* **31**, 244-250 (2018).
3. Morgan GJ, Kenny D, Duke C, Walsh KP, Qureshi SA, Initial assessment of a novel delivery system (NuDEL™) for the covered Cheatham-Platinum stent. *Cardiology in the Young* **27**, 1465-1469 (2017).
4. Bishnoi RN, Jones TK, Kreutzer J, Ringel RE, NuMED Covered Cheatham-Platinum Stent for the treatment or prevention of right ventricular outflow tract conduit disruption during transcatheter pulmonary valve replacement. *Catheterization and cardiovascular interventions : official journal of the Society for Cardiac Angiography & Interventions* **85**, 421-427 (2015).
5. Sohrabi B, Jamshidi P, Yaghoubi A, Habibzadeh A, Hashemi-Aghdam Y, Moin A, Kazemi B, Ghaffari S, Abdolazadeh BM, Mahmoodi K, Comparison between covered and bare Cheatham-Platinum stents for endovascular treatment of patients with native post-ductal aortic coarctation: immediate and intermediate-term results. *JACC. Cardiovascular interventions* **7**(4), 416-423 (2014).
6. Vanagt WY, Cools B, Boshoff DE, Frerich S, Heying R, Troost E, Louw J, Eyskens B, Budts W, Gewillig M, Use of covered Cheatham-Platinum stents in congenital heart disease. *International Journal of Cardiology* **175**, 102-107 (2014).
7. Alcibar J, Blanco R, Fernandez L, Arriola J, Garcia K, Pena N, Inguanzo R, Voces R, Castellanos E, Montes PM, Elective implantation of covered stents for coarctation and recoarctation in adolescents and adults. *Revista espanola de cardiologia (English ed.)* **66**, 443-449 (2013).
8. Chang ZP, Jiang SL, Xu ZY, Zhang GJ, Huang LJ, Zhao SH, Ling J, Zheng H, Jin JL, Wu WH, Hu HB, Li SG, Yu JH, Yan CW, Use of covered Cheatham-Platinum stent as the primary modality in the treatment for native coarctation of the aorta. *Chinese medical journal* **125**, 1005-1009 (2012).
9. Erdem A, Akdeniz C, Sarıtaş T, Erol N, Demir F, Karaci AR, Yalçın Y, Celebi A, Cheatham-Platinum stent for native and recurrent aortic coarctation in children and adults: immediate and early follow-up results. *Anadolu Kardiyol Derg.* **Aug;11**(5), 441-449 (2011).
10. Butera G, Heles M, MacDonald ST, Carminati M, Aortic coarctation complicated by wall aneurysm: the role of covered stents. *Catheterization and cardiovascular interventions : official journal of the Society for Cardiac Angiography & Interventions* **78**, 926-932 (2011).
11. Tanous D, Collins N, Dehghani P, Benson LN, Horlick EM, Covered stents in the management of coarctation of the aorta in the adult: initial results and 1-year angiographic and hemodynamic follow-up. *Int J Cardiol* **140**, 287-295 (2010).
12. Moltzer E, Roos-Hesselink JW, Yap SC, Cuypers JA, Bogers AJ, de Jaegere PP, Witsenburg M, Endovascular stenting for aortic (re)coarctation in adults. *Netherlands heart journal : monthly journal of the Netherlands Society of Cardiology and the Netherlands Heart Foundation* **18**, 430-436 (2010).
13. Agnoletti G, Marini D, Ou P, Vandrell MC, Boudjemline Y, Bonnet D, Cheatham platinum (CP) and Palmaz stents for cardiac and vascular lesions treatment in patients with congenital heart disease. *EuroIntervention* **4**, 620-625 (2009).
14. Bruckheimer E, Dagan T, Amir G, Birk E, Covered Cheatham-Platinum stents for serial dilation of severe native aortic coarctation. *Catheterization and cardiovascular interventions : official journal of the Society for Cardiac Angiography & Interventions* **74**, 117-123 (2009).
15. Tzifa A, Ewert P, Brzezinska-Rajszy G, Peters B, Zubrzycka M, Rosenthal E, Berger F, Qureshi SA, Covered Cheatham-platinum stents for aortic coarctation: early and intermediate-term results. *J Am Coll Cardiol* **47**, 1457-1463 (2006).
16. Meadows J, Minahan M, McElhinney DB, McEnaney K, Ringel R, Intermediate Outcomes in the Prospective, Multicenter Coarctation of the Aorta Stent Trial (COAST). *Circulation* **131**, 1656-1664 (2015).



**NuMED**  
**Summary of Safety and Clinical Performance**  
**SSCP – Stents – CoA & RVOT**

17. Taggart NW, Minahan M, Cabalka AK, Cetta F, Usmani K, Ringel RE, Immediate Outcomes of Covered Stent Placement for Treatment or Prevention of Aortic Wall Injury Associated With Coarctation of the Aorta (COAST II). *JACC Cardiovasc Interv* **9**, 484-493 (2016).
18. Sasikumar D, Sasidharan B, Rashid A, Ayyappan A, Goplakrishnan A, Krishnamoorthy K, Sivasubramonian S. Early and late outcome of covered and non-covered stents in the treatment of coarctation of aorta- A single centre experience, *Indian Heart Journal* **72**, 278-282 (2020).
19. Stassen J., De Meester P., Troost E., Roggen L., Moons P., Gewillig M., Van De Bruaene A., Budts W. Covered stent placement for treatment of coarctation of the aorta: immediate and long-term results. *Acta Cardiologica* 2020.
20. Holzer R.J., Gauvreau K., McEnaney K., Watanabe H., Ringel R. Long-Term Outcomes of the Coarctation of the Aorta Stent Trials. *Circulation: Cardiovascular Interventions* 2021 (582-589) Article Number e010308.
21. Kasar T, Erkut O, Tanidir İC, Şahin M, Topkarci MA, Guzeltas A. Balloon-expandable stents for native coarctation of the aorta in children and adolescents. *Medicine (United States)*. 2022;101(51). doi: 10.1097/MD.00000000000032332.
22. Bairam AR, Ali AH, Al Ghizzi HJ. Immediate and intermediate term results of Cheatham platinum Stenting for native coarctation of aorta in adults. *Revista Latinoamericana de Hipertension*. 2021;15(3):192-9. doi: 10.5281/zenodo.5651255.
23. Alsamarra H, Shareif MM, Ibrahim SA, Alhasani HA. Immediate results of balloon angioplasty with stenting in patients with coarctation of descending aorta. *Revista Latinoamericana de Hipertension*. 2023;18(7):310-3. doi: 10.5281/zenodo.10091668.
24. Schleiger A, Al Darwish N, Meyer M, Kramer P, Berger F, Nordmeyer J. Long-term follow-up after endovascular treatment of aortic coarctation with bare and covered Cheatham platinum stents. *Catheterization and Cardiovascular Interventions*. 2023;102(4):672-82. doi: 10.1002/ccd.30793.

<b>10. Revision History</b>			
<b>SSCP revision number</b>	<b>Date Issued</b>	<b>Change Description</b>	<b>Revision validated by Notified Body</b>
00	21 June 2022	Initial implementation	<input type="checkbox"/> Yes Validation Language: English <input checked="" type="checkbox"/> No
01	06 July 2023	Updated sections 4, 5, 7, 8, and 9 for CER Update.	<input type="checkbox"/> Yes Validation Language: English <input checked="" type="checkbox"/> No
02	28 July 2023	Updated Section 2.	<input type="checkbox"/> Yes Validation Language: English <input checked="" type="checkbox"/> No
03	18 February 2025	Revised Section 1 to remove reference to G-Armor Covered Stents and updated Basic UDI-DI. Revised Section 2 to update intended purpose. Revised Section 3 to add model variants. Revised Section 5 to remove reference to G-Armor Covered Stents, and update clinical literature. Updated Section 8 for harmonized standards and Section 9 for references.	<input type="checkbox"/> Yes Validation Language: English <input checked="" type="checkbox"/> No



# NuMED

## Summary of Safety and Clinical Performance

### SSCP – Stents – CoA & RVOT

Document Revision: 03  
Date issued: 18 February 2025

*This Summary of Safety and Clinical Performance (SSCP) is intended to provide public access to an updated summary of the main aspects of the safety and clinical performance of the device. The information presented below is intended for patients or lay person. A more extensive summary of its safety and clinical performance prepared for healthcare professionals is found in the first part of this document.*

*The SSCP is not intended to give general advice on the treatment of a medical condition. Please contact your healthcare professional in case you have questions about your medical condition or about the use of the device in your situation. This SSCP is not intended to replace an Implant card or the Instructions for Use to provide information on the safe use of the device.*

<b>1. Device identification and general information</b>	
Device trade name(s)	Covered CP Stent Covered Mounted CP Stent
Manufacturer's name and address	NuMED, Inc. 2880 Main Street Hopkinton, NY 12965 USA
Year when first certificate (CE) was issued	2004 (Covered CP Stent) 2009 (Covered Mounted CP Stent)
Basic UDI-DI	Covered CP Stent – 08877141650TH Covered Mounted CP Stent – 08877141660TL

<b>2. Intended use of the device</b>	
Intended purpose	<p>The Stents are intended to dilate aortic coarctations.</p> <p>An aortic coarctation is a partial blockage or narrowing in the aorta, the body's main blood vessel distributing blood to all parts of the body. This blockage of the aorta makes the heart work harder to pump blood to your body and can weaken the heart muscle. Furthermore, this blockage can cause severe upper body hypertension (high blood pressure), increasing the risk of stroke. This blockage is present from birth.</p> <p>The Stents are also intended to contain right ventricle to pulmonary artery (right ventricular outflow tract) conduit disruptions that are identified during conduit pre-dilatation procedures performed in preparation for transcatheter pulmonary valve replacement.</p> <p>A Right Ventricular Outflow Tract (RVOT) is also known as a pulmonary conduit, and it is a tube that connects the heart to the lungs. Placement of an RVOT is typically associated in patients that have one of the following conditions: Pulmonary Atresia, Tetralogy of Fallot, or Double Outlet Right Ventricle. These three conditions can lead to pulmonary conduit failure.</p>
Indications and intended patient groups	The device is used to treat any patients that have an aortic coarctation or RVOT conduit disruptions as long as none of the below listed contraindications and/or limitations are applicable.
Contraindications and/or limitations	<p>The following patients should NOT receive the Stent:</p> <ul style="list-style-type: none"> <li>• Patients who are too small to allow the stent to pass through their arteries without damaging the artery;</li> <li>• Patients with a stiff aorta that does not get larger with balloon dilation. (CoA only)</li> </ul>



## NuMED

### Summary of Safety and Clinical Performance

#### SSCP – Stents – CoA & RVOT

- Patients with blocked leg arteries making it difficult or unsafe to move the catheter and stent to the narrowed aorta; (CoA only)
- Patients with any signs of infection;
- Patients with active infection in the heart or blood vessels (endocarditis);
- Patients with a known allergy to aspirin, other antiplatelet agents, or heparin; (CoA only)
- Pregnancy.

<b>3. Device description</b>	
Description of the device	<p>The Stents are balloon expandable and intended to permanently stay in your body. The Stents are used for coarctation of the aorta or treatment of right ventricle to pulmonary artery (right ventricular outflow tract) conduit disruptions that are identified during conduit pre-dilatation procedures performed in preparation for transcatheter pulmonary valve replacement (TPVR). The covering acts as a fluid barrier creating a fluid tight conduit through the stent length. Blood cannot flow across the covering.</p> <p>The BIB Stent Placement Catheter is triaxial in construction with two lumens being used to inflate the balloon while one lumen is being used for tracking over a guidewire. The inner balloon is ½ of the outer balloon diameter and 1 cm shorter. The purpose of the double balloon catheter is to apply an incremental inflation for the purpose of dilating a stent. The inner balloon provides initial expansion of the stent and also acts as a tool to hold the stent on the catheter prior to the outer balloon being inflated. The outer balloon is then inflated, providing the remainder of the expansion. There are radiopaque platinum marker bands under the balloon shoulders, to aid during placement. The balloons are designed to inflate to the diameter and length listed on the label at a specific pressure. Thus, it is recommended that the device be used in conjunction with a mechanism to monitor pressure, an inflation device with pressure gauge.</p> <p>The Stents are composed of heat treated 90% platinum / 10% iridium wire that is arranged in a "zig" pattern, laser welded at each joint, and over brazed with 24K gold. The number of rows determines the unexpanded length of the stent. The Covered versions have an ePTFE covering that is attached to the metal wire frame.</p>
Medicinal Substances	The Stents do not contain any medicinal substances.
Mode of Action	The Stents are implanted using a thin hollow tube (catheter) with a balloon on the end. Your physician will place the stent on the balloon at the start of your procedure. The catheter with the stent is then placed through the skin, typically into the artery in your upper leg. The balloon and stent are moved to the appropriate position at the narrowed part of your aorta or in the RVOT. Once in place, the balloons are inflated to expand the stent. The catheter is then removed from the body and the stent stays in place.
Description of Accessories	<p>All Stents are packaged and shipped to the physician with hemostasis valve tools. These tools are hollow tubes that are placed in the valve of the introducer to help the Stent move through that valve without any issues. The valve of the introducer is very tight to prevent blood loss during the procedure, so the tools help the Stent move through the valve without causing damage to the stent or moving the stent on the catheter.</p> <p>The stent is also used with other accessories not provided by NuMED, including, a delivery catheter to expand the stent (unmounted version only), guidewire, introducer, balloon inflation medium to inflate the balloon, inflation device with pressure gauge, and a stopcock.</p>

<b>4. Risks and Warning</b>	
<i>Contact your healthcare professional if you believe that you are experiencing side effects related to the device or its use or if you are concerned about risks. This document is not intended to replace a consultation with your healthcare professional if needed.</i>	
How potential risks have been controlled	The Stent Device Family has been developed in accordance with documented processes to ensure that it is designed, manufactured, packaged, and labelled in accordance with the current state of the art



## NuMED

### Summary of Safety and Clinical Performance

#### SSCP – Stents – CoA & RVOT

or managed	and meets all requirements of the appropriate regulations. Design verification activities were performed and include pre-clinical testing and clinical investigations. A clinical literature review has also been performed on the Stent Device Family. All risks identified during these activities were mitigated as far as possible and are considered acceptable in regards to the clinical benefit of the device. Continued review of all Post Market Surveillance and Post Market Clinical Follow-up Data is performed to identify any additional risks that may be identified after the device was placed on the market.
Remaining risks and undesirable effects	<p>Cardiac catheterization and stent insertion carry certain risks. Potential complications &amp; adverse effects associated with device use and indication include:</p> <ul style="list-style-type: none"> <li>• Femoral Artery Injury</li> <li>• Stent Migration – movement of the stent away from original implant site</li> <li>• Stent Stenosis – growth of tissue within the stent, leading to return of the blockage</li> <li>• Stent Fracture – break in the frame of the stent</li> <li>• Aneurysm/Pseudoaneurysm – weakening or injury of the aorta wall</li> <li>• Aortic Rupture/Tear – perforation or tearing of the aorta, causing internal bleeding</li> <li>• Stent Malposition – poor position of stent, requiring a 2nd stent</li> <li>• Hematoma – bruising at the site where the device is introduced into the body</li> <li>• Sepsis/infection – Infection</li> <li>• Thrombosis – formation or presence of a blood clot</li> <li>• Embolization – passage and lodging of an embolus within the bloodstream</li> <li>• Transitory arrhythmia – Irregular heartbeat</li> <li>• Endocarditis – infection within the stent</li> <li>• Bleeding – at the site of where the device is introduced into the body</li> <li>• Cerebrovascular Incident – stroke</li> <li>• Death</li> </ul>
Warning and Precautions	<p>The majority of warnings and precautions listed for the Stents pertain to the placement and use of the device in the cath. lab by the physician.</p> <p>MRI Conditional information is applicable to the Stents after they are implanted. This information should be used by any MRI technician that is performing an MRI procedure on any patient with a NuMED Stent implanted. All patients will be provided with an Implant Card after their procedure. This Implant Card will give the location of where to find the most up to date MRI parameters to be used for patients that have a NuMED Stent implanted.</p>
Summary of any field safety corrective actions (FSCA including FSN) if applicable	There have not been any Field Safety Corrective Actions or Field Safety Notices on any versions of the Stents listed in this SSCP.

<b>5. Summary of clinical evaluation and post-market clinical follow-up</b>	
Clinical background of the device	<p>The NuMED Stent Device Family has been sold globally since 1999.</p> <p>The NuMED Stent was tested and found to be safe and effective to repair aortic wall injuries and to widen the narrow part of the aorta related to coarctation of the aorta. A study was conducted with 82 patients weighing more than 31 lbs at the time of implant. Most of the patients (89%) were treated with one Covered CP stent, 11% needed more than one to complete the repair.</p> <p>On average arm systolic blood pressure was 25 mmHg higher than the leg pressure before the procedure. A reduction of a gradient to 15mmHg or less following the procedure suggests that the blockage is reduced effectively. By one month after covered stent placement the average arm pressure was only 1 mmHg higher than the leg pressure. Two years after implant, 85% of patients</p>




## NuMED

### Summary of Safety and Clinical Performance

### SSCP – Stents – CoA & RVOT

	<p>had arm blood pressures less than 15 mmHg above their leg pressure, which suggests that most of the treated aortas did not re-narrow. Repair of aortic wall injury was successful in all of the 49 patients who received their Covered CP Stent to repair their weakened aortic wall. An overview of complications and additional treatments provided after the stenting procedure is shown below:</p> <ul style="list-style-type: none"> <li>• Serious complications related to the Covered CP Stent or implant procedure, such as: causing injury to the aortic wall or damage to the leg artery used for Stent insertion, were identified in 6 out of 100 (6%) of patients within the first month of implant.</li> <li>• No patients needed surgery to repair the aorta or to remove the stent.</li> <li>• One patient required stent repair of the leg artery damaged during insertion of the implant catheter.</li> <li>• One patient required surgical repair of the leg artery damaged during insertion of the implant catheter.</li> <li>• 1 out of 20 (5%) patients developed small aneurysms (weakened areas of the aorta) in the area of stent placement in the years following stent therapy, making CT or MRI imaging an important part of follow up care. However, none of the patients who developed aneurysms demonstrated symptoms or required surgery. All were successfully treated with additional covered stent placement.</li> <li>• Overall, 16% of patients required repeat cardiac catheterization for a second dilation of the stent, mostly to keep up with the size of the patient as he/she grew and for some to repair aortic wall injuries as noted above.</li> </ul> <p>The NuMED Covered Stent was tested and found to be safe and effective to use as a Treatment of right ventricle to pulmonary artery (right ventricular outflow tract) conduit disruptions that are identified during conduit pre-dilatation procedures performed in preparation for transcatheter pulmonary valve replacement (TPVR). A study was conducted with 50 patients weighing an average of 58 kg, at the time of implant. Most patients (80%) were treated with one Covered CP stent.</p> <p>Out of 49 patients treated with the Covered CP Stent (CCPS), 81.6% of them had device and lesion success with no adverse events attributed to the CCPS. Out of 49 patients treated with the CCPS, 93.9% of the patients had successful coverage of conduit disruption followed by successful implantation of an artificial valve. An overview of complications and additional treatments provided after the stenting procedure is shown below:</p> <ul style="list-style-type: none"> <li>• Serious complications related to the CCPS or stent implant procedure, such as: stent embolization was identified in 1 out of 50 (2%) patients.</li> <li>• 7 (14%) of the patients required a second CCPS, and (3) 6% of the patients required a third CCPS during the procedure. Of these 10 patients, 4 (40%) of them planned on having the second CCPS implanted before the procedure.</li> </ul>
<p>The clinical evidence for the CE marking</p>	<p>The CE marking was based on data from three clinical studies, a review of published literature, and a review of post market surveillance data provided by NuMED. Additional pre-clinical testing was performed as part of the development and design of the device. In vitro (on the bench) testing was performed on the devices as part of the Design History File. Biocompatibility testing was also performed on the materials used to manufacture this device to determine if it met the requirements for an implant in the human body. The device passed all tests.</p>
<p>Safety</p>	<p>The clinical data and pre-clinical study data demonstrated that the device performed as intended by NuMED in the clinical setting; the device does not pose unacceptable safety concerns in the clinical setting; and any risks associated with clinical use of the device are acceptable when weighed against the benefits to the patient.</p>



## NuMED

### Summary of Safety and Clinical Performance

### SSCP – Stents – CoA & RVOT

#### 6. Possible diagnostic or therapeutic alternatives

*When considering alternative treatments, it is recommended to contact your healthcare professional who can take into account your individual situation.*

##### **Coarctation of the Aorta**

Your cardiologist believes that relief of the blockage is important for your health and safety. There are three ways to relieve the blockage: by surgery, by stent implantation without surgery, or by balloon angioplasty.

##### **Surgical Therapy**

Surgical treatment of the blockage is usually performed through an incision on the side of the chest, approaching the aorta by spreading the ribs. The narrowed portion of the aorta is removed and then the aorta is sewn back together. For more complicated coarctation, surgery might be performed from the front of the chest, opening the breast bone and using heart lung bypass. For some patients a benefit of a surgical approach is that the repair can be performed without the use of man-made materials. However, for other (especially adult) patients a man-made tube graft or patch may be needed. Please consult with your surgeon regarding his or her approach. For younger patients, surgery results in a lower need for a second procedure to keep up with growth when compared to balloon or stent therapy.

Risks of surgery include: pain from the surgical incision, prolonged fluid drainage from the chest after surgery, chest or wound infection, longer recovery time compared to stent therapy, prolonged postoperative rib discomfort and increased risk of very high blood pressure occurring after immediately after surgery, requiring intravenous therapy in an ICU, compared to stent repair. There is a low risk, probably less than 5%, of developing an aneurysm (weakened areas of the aorta) in the area of surgery in the years following stent therapy, making CT or MRI imaging an important part of follow up care.

##### **Stent Therapy (without surgery)**

A stent is an expandable metal tube that is implanted into your aorta to keep it open. Surgery is not required for this procedure. The stent is implanted using a thin hollow tube (catheter) with a balloon on the end. The catheter with stent is inserted through the artery in the upper leg. The balloon and stent are then moved to the appropriate position to the narrowed part of your aorta. Once in place, the balloons are inflated to expand the stent against the aortic wall. The catheter is then removed from the body and the stent remains in place.

##### **Balloon Angioplasty**

A specially designed catheter with a tiny balloon is carefully guided through the artery to the blockage, then inflated to widen the opening and increase blood flow to the heart.

#### **RVOT**

There are three ways to treat pulmonary conduit failure. One is a surgical conduit replacement, one is Transcatheter Pulmonary Valve Replacement, and the last is Balloon Valvuloplasty / Angioplasty.

##### **Surgical Replacement:**

Surgical replacement of a pulmonary valve conduit involves a physician removing the narrow or leaking conduit and replacing it with an artificial valve.

##### **Transcatheter Pulmonary Valve Replacement:**

An artificial valve is mounted on a thin hollow tube (catheter) with a balloon on the end, and is inserted into the artery in your upper leg. It is then advanced to the pulmonary conduit and the balloon inflated to place the new artificial pulmonary valve. The catheter is then removed from the body.

##### **Balloon Valvuloplasty / Angioplasty:**

A thin hollow tube (catheter) with a balloon on the end is inserted into the artery in your upper leg and advanced to the pulmonary conduit. The balloon is then inflated to a specified pressure to open your conduit so that the blood will flow better. The catheter is then removed from the body.

#### 7. Suggested profile and training for users

The Stent Device Family is intended for use by trained cardiology and surgical professionals undertaking stent implantation.