

G-Armor Stent®

COARCTATION OF THE AORTA

INSTRUCTIONS FOR USE


General Instructions.....	Page 2
Stent Sizing Charts and Information.....	Page 22

**CAUTION: FEDERAL (USA) LAW RESTRICTS THIS DEVICE TO
SALE BY OR ON THE ORDER OF A PHYSICIAN.**

Read all instructions prior to use.

Interventional Systems
B | BRAUN

Distributed by:
B. Braun Interventional Systems Inc.
824 Twelfth Avenue
Bethlehem, PA 18018
Customer Service:
TEL: (877) 836-2228
FAX: (610) 849-1334
Technical Support
TEL: (800) 443-8362
Made in U.S.A.



Manufactured By:
NuMED, Inc.
2880 Main Street
Hopkinton, New York
USA 12965
Telephone: (315) 328-4491
Facsimile: (315) 328-4941
Email: info@numedusa.com
Internet: www.numedforchildren.com

INDICATION

The G-Armor Stent® is indicated for use in the treatment of native and/or recurrent coarctation of the aorta involving a compliant aortic isthmus or first segment of the descending aorta where there is adequate size and patency of at least one femoral artery and balloon angioplasty is contraindicated or predicted to be ineffective.

DESCRIPTION

The G-Armor Stent® is balloon expandable and intended for permanent implant. The stent is composed of heat treated 90% platinum and 10% iridium wire that is arranged in laser welded rows with a “zig” pattern. The number of zigs in a row can be varied and will impact the strength of the stent as well as the eventual expanded diameter and percentage of stent shortening, while the number of rows will determine the unexpanded length of the stent.

CONTRAINDICATIONS

- Patients too small to allow safe delivery of the stent without compromise to the systemic artery used for delivery;
- Unfavorable aortic anatomy that does not dilate with high pressure balloon angioplasty;
- Curved vasculature;
- Occlusion or obstruction of systemic artery precluding delivery of the stent;
- Clinical or biological signs of infection;
- Active endocarditis;
- Known allergy to aspirin, other antiplatelet agents, or heparin;
- Pregnancy.

HOW SUPPLIED

Supplied sterilized by ethylene oxide gas. Sterile and non-pyrogenic if package is unopened or undamaged. Do not use the product if there is doubt as to whether the product is sterile. Avoid extended exposure to light. Upon removal from package, inspect the product to ensure no damage has occurred.

WARNINGS

- Radiofrequency heating during MRI scans on overlapped, 10 zig stents has not been evaluated.
- Coarctation of the aorta involving the aortic isthmus or first segment of the descending aorta should be confirmed by diagnostic imaging.
- The stent has not been evaluated in patients weighing less than 20kg.
- As with any type of implant, infection secondary to contamination of the stent may lead to aortitis, or abscess.
- Over-stretching of the artery may result in rupture or aneurysm formation.
- When the stent is crimped onto a balloon delivery catheter, the maximum balloon inflation pressure must not exceed the recommended inflation pressure specified in the manufacturer’s instructions.
- Excessive force while crimping may weaken welds of the stent.
- Crimping the 8 zig stent on a balloon catheter smaller than 12mm, and the 10 zig on a balloon catheter smaller than 26mm, may cause damage to the stent.

- This device is intended for single use only. Do not resterilize and/or reuse it, as this can potentially result in compromised device performance and increased risk of cross-contamination.

PRECAUTIONS

- Use of an inflation device with pressure gauge is highly recommended during this procedure.
- The stent is rigid and may make negotiation through vessels difficult.
- Dilatation procedures should be conducted under fluoroscopic guidance with appropriate x-ray equipment.
- The inflation diameter of the balloon used during stent delivery should approximate the diameter of the obstructive vessel and the intended implant site.
- Careful attention must be paid to the maintenance of tight catheter connections and by aspiration before proceeding to avoid air introduction into the system.
- Under no circumstances should any portion of the catheter system be advanced against resistance. The cause of the resistance should be identified with fluoroscopy and action taken to remedy the problem.

POTENTIAL COMPLICATIONS / ADVERSE EFFECTS

NOTE: Circumferential tear of the delivery balloon catheter prior to complete expansion of the stent may cause the balloon to become tethered to the stent, requiring surgical removal. In case of rupture of an adequately sized balloon after stent expansion, it can be withdrawn and a new balloon catheter exchanged over a guidewire to complete expansion of the stent.

Cardiac catheterization carries certain risks. In addition, potential complications, and related adverse effects associated with implants include, but are not limited to:

- Femoral artery injury, thrombosis or pseudoaneurysm
- Stent Migration
- Stent Fracture
- Aortic Rupture/Tear
- Hematoma
- Thrombosis/Thromboembolism
- Death
- Endocarditis
- Cell necrosis at the site of implant
- Cerebrovascular Incident
- Stent Stenosis
- Aortic Aneurysm/Pseudoaneurysm
- Stent Malposition
- Sepsis/infection
- AV fistula formation
- Transitory arrhythmia
- Bleeding



MRI SAFETY INFORMATION

Nonclinical testing and modeling has demonstrated that the stent is MR Conditional. A patient with this device can be safely scanned in an MR system meeting the following conditions:

- Static magnetic field of 1.5 T and 3 T
- Maximum spatial gradient magnetic field of 2500 gauss/cm (25 T/m)
- Maximum MR system reported, whole body averaged specific absorption rate (SAR) of 2.0 W/kg for 15 minutes of scanning (Normal Operating Mode)

Based on nonclinical testing and modeling, under the scan conditions defined above, the stent is expected to produce a maximum in vivo temperature rise of less than 2°C after 15 minutes of continuous scanning.

MR image quality may be compromised if the area of interest is in the same area, or relatively close to the position of the device. In nonclinical testing, the image artifact caused by the device extends approximately 3 mm from the stent when imaged with a spin echo pulse sequence and 6 mm when imaged with a gradient echo pulse sequence and a 3 T MRI System. The lumen of the device was obscured.

The presence of other implants or medical circumstances of the patient may require lower limits on some or all of the above parameters.

Warning: Radiofrequency heating during MRI scans on overlapped, 10 zig stents has not been evaluated.

INSTRUCTIONS FOR USE:

Select Stent Size

1. Measure the length of the target stricture to determine the length of stent required. Size the stent length to extend slightly proximal and distal to the stricture.
2. The appropriate stent length should be selected based on covering the entire obstructed segment with a single stent.
Note: Should more than one stent be required, place the stent most distal from the puncture site first, followed by placement of the proximal stent in tandem.
3. Measure the diameter of the reference stricture and vessel proximal and distal to the target lesion to determine the appropriate size stent and delivery system.

Preparation of Stent

- Remove the stent from the packaging vial.

Preparation of Stent Delivery Catheter

Note: The BIB Stent Placement Catheter is the only device used in the clinical study. Results with other delivery systems is unknown.

- See Instruction for Use provided with the recommended balloon catheter.

Preparation of Stent System

1. Always place an appropriately sized guidewire or mandrel through the balloon catheter guidewire lumen while prepping the balloon and crimping the stent.
2. Slide the stent over the distal end of the balloon, maintaining the balloon fold, until the radiopaque markers are equal distance from the ends of the stent.
3. Gently crimp (see insert IFU-CPCE) the stent on the balloon by using finger pressure and a “rolling action” to exert equal pressure on all sides of the stent. Only crimp until no movement is felt on the catheter. Platinum is very malleable and should be easily crimped. Place a small amount of undiluted contrast to “coat” the stent and improve adherence to the balloon. (AVOID BENDING OR TWISTING THE STENT)

4. Visually inspect the balloon/stent assembly to assure proper placement of the stent and to assure that the stent is evenly crimped. Uneven crimping could cause the stent to deploy in a non-symmetrical manner.

Stent Deployment

1. **Use of the tools supplied with the stent is necessary to defeat the hemostasis valve without damaging the stent. Refer to insert IFU-CPCE. Once the stent is past the hemostasis valve, the tool must be pulled out of the valve.**
2. The system is advanced through the long delivery sheath and over the stiff guidewire into the desired location for implant.
3. After correct positioning of the stent, pull back on the sheath to expose the stent. Confirm proper stent position by a small injection of contrast through the sidearm of sheath or through a second catheter. Expand the stent by inflating the balloon catheter to the recommended inflation pressure. Do not exceed the balloon manufacturer's rated burst pressure.

Delivery System Withdrawal

1. Once the stent is expanded, deflate the balloon and rotate to ensure the stent is free and properly deployed.
2. Remove the balloon catheter and confirm the result with angiography.

NOTE: Diameter of the stent may be increased after placement by expanding with a larger diameter balloon. Do not exceed the maximum recommended expanded stent diameter of 24mm for the 8 zig stents, and 30mm for the 10 zig stents.

CLINICAL STUDY INFORMATION

The COAST clinical study was performed to establish a reasonable assurance of safety and effectiveness of implantation of the Bare CP Stent, another version of the G-Armor Stent®, in the native and/or recurrent coarctation of the aorta in the US. The study was a prospective, multi-center, single-arm clinical study comparing stent treatment of native or recurrent aortic coarctation to a performance goal (PG) derived from surgical treatment. Surgical PGs are derived from retrospective data collection at selected participating centers and from the literature.

1. Clinical Inclusion and Exclusion Criteria

Enrollment in the COAST study was limited to patients who met the following inclusion criteria:

Pre-catheterization Inclusion Criteria:

- a. Native or recurrent aortic coarctation
- b. Weight 35 kg
- c. Noninvasive, arm-leg cuff systolic blood pressure* difference or catheter measured systolic coarctation gradient 20 mmHg

*Patients receiving antihypertensive therapy can be included in the study. The type and dose of the medication will be recorded and used for comparisons with follow up evaluations.

Catheterization Inclusion Criteria:

- a. Coarctation of the aorta, either native or recurrent that is demonstrated, angiographically to involve the aortic isthmus or first segment of the descending aorta
- b. Coarctation of the aorta found to be compliant on pre-stent balloon dilation
- c. Patency of at least one femoral artery

Patients were not permitted to enroll in the COAST study if they met any of the following exclusion criteria:

Pre-catheterization Exclusion Criteria:

- a. Age > 60 years
- b. Connective tissue disorders, including Marfan syndrome and other genetic syndromes such as Turner syndrome and Noonan syndrome
- c. Inflammatory aortitis
- d. Bloodstream infection, including endocarditis
- e. Pregnancy
- f. Aortic aneurysm
- g. Prior stent placement
- h. Adults lacking capacity to consent
- i. Foster children and/or wards of the court

Catheterization Exclusion Criteria

- a. Angiography that demonstrates aortic coarctation involving a “curved” region of the aorta, the transverse aortic arch, carotid arterial branches or obstruction extending into or beyond the mid-thoracic descending aorta
- b. Complete aortic atresia demonstrated angiographically
- c. Anatomic location of coarctation judged by operator to preclude safe placement of a stent
- d. Coarctation of the aorta found to be non-compliant on pre-stent balloon dilation

2. Follow-up Schedule

All patients were scheduled to return for follow-up examinations at 1 month, 6 months, 12 months, 24 months, and annually to 5 years. Adverse events and complications were recorded at all visits.

3. Clinical Endpoints

With regards to safety, the following criteria were evaluated:

Primary Safety Endpoint #1: Occurrence of any serious or somewhat serious adverse event attributed to the stent or implantation procedure within 30 days of the catheterization procedure.

The following hypothesis was tested using a one-sample, one-sided test of proportions conducted at the 0.05 level of significance:

$$H_0: p \geq 0.18 \text{ vs. } H_A: p < 0.18$$

Primary Safety Endpoint #2: Occurrence of post-procedure paradoxical hypertension.

The following hypothesis was tested using a two-sided, one-sample test of proportions conducted at the 0.05 level of significance:

$$H_0: p = 0.84 \text{ vs. } H_A: p \neq 0.84$$

With regards to effectiveness, the following criteria were evaluated.

Primary Effectiveness Endpoint #1: Reduction in arm-leg systolic blood pressure gradient from pre-dilation to the 12 month post-dilation follow-up.

Assuming that μ represents the true mean gradient reduction among stent patients, the following hypothesis was tested using a one-sample, one-sided t test conducted at the 0.05 level of significance:

$$H_0: \mu \leq 31 \text{ vs } H_A: \mu > 31$$

Primary Effectiveness Endpoint #2: Length of stay in the hospital, measured in days.

The following hypothesis was evaluated using a two-sided, one-sample t test conducted at the 0.05 level of significance:

$$H_0: \mu = 3.5 \text{ vs. } H_A: \mu \neq 3.5$$

Of the 167 enrolled patients, 102 met the study eligibility criteria and were treated with the Bare CP Stent. 112 patients were evaluated for safety, of which 5 patients crossed-over to the Covered CP Stent therapy. Approximately 107 stents were implanted in 105 patients and these patients were included in the evaluation of effectiveness. Study accountability is detailed in Table 1.

Table 1. COAST Accountability

	Possible N (100%)	1 Month Visit n (%)	12 Month Visit n (%)	24 Month Visit n (%)	3 years n (%)	4 years n (%)	5 years n (%)
Safety Cohort	112	102 (91%)	94 (84%)	90 (80%)	80 (71%)	76 (68%)	73 (65%)
Effectiveness Cohort	105*	100 (95%)	92 (88%)	87 (83%)	77 (73%)	69 (66%)	56 (53%)

*112 patients underwent catheterization and pre-stenting balloon angioplasty, five then received Covered CP Stents and were entered into the COAST II trial where they have been followed since. Two patients did not receive study stents and are followed for safety outcomes only.

Subject Demographics

Table 2 presents subject demographics and baseline characteristics analyzed for the enrolled subjects.

Table 2. COAST Patient Characteristics

Assessment	Number (Percent) or Median (Range)	
	Safety Cohort (n=112)	Efficacy Cohort (n=105)
Gender		
Male	77 (69%)	73 (70%)
Female	35 (31%)	32 (30%)
Age, years	16 (8 to 52)	16 (8 to 52)
NYHA Classification		
I	88 (79%)	82 (78%)
II	22 (20%)	21 (20%)
III	1 (1%)	1 (1%)
IV	1 (1%)	1 (1%)
Primary Indication		
Native Coarctation	65 (58%)	60 (57%)
Recurrent Coarctation	47 (42%)	45 (43%)

The analysis of safety was based on the implanted cohort of 112 COAST patients. The primary safety outcomes are presented in Table 3.

Table 3. Summary of COAST Outcomes and Pre-Specified Safety Endpoints

COAST	Safety Endpoint	Event Rate	P Value (CI)
Primary	Serious or Somewhat Serious Adverse event attributed to the Stent or Implantation procedure within 30 days of the procedure	8.9%	0.006 (4.9%, 14.7%)*
	Post-procedure paradoxical hypertension	7.5%	<0.001 (3.3%, 14.2%)+

*90% Confidence interval

+ 95% Confidence Interval

confidence interval provided to illustrate the variability only and should not be used to draw any statistical conclusion.

The COAST primary safety endpoints were met with the occurrence of any serious or somewhat serious adverse events within 30 days post procedure being less than the predefined 18%. Post procedural paradoxical hypertension was observed in 7.5% of patients in the COAST Trial. The COAST primary safety endpoint for the incidence of paradoxical hypertension (< 84%) was met.

The overall incidence and types of adverse events were within expected ranges. Aortic wall injuries were rare and treated appropriately without the need for emergency surgery. There were no late complications related to device fracture noted, though incidence of stent fracture increased with time for COAST patients. Table 4 provides a summary of the adverse events reported under COAST.

Table 4. Summary of Adverse Events (AEs) for COAST

	Stent Related Events¹ (Rates)	Stent, Implantation, or Catheterization Related Events² (Rates)	All Events (Rates)
Patients with adverse events at 30 days	1 (0.9%)	37 (33.0%)	50 (44.6%)
Serious or somewhat serious events at 30 days	1 (0.9%)	10 (8.9%)	11 (9.8%)
Serious or somewhat serious events at 30 days, excluding stent fracture	1 (0.9%)	10 (8.9%)	11 (9.8%)
Serious event at 30 days	0 (0.0%)	0 (0.0%)	0 (0.0%)
Patients with adverse events at 12 Months	4 (3.6%)	40 (35.7%)	69 (61.6%)
Serious or somewhat serious events at 12 months	4 (3.6%)	13 (11.6%)	16 (14.3%)
Serious or somewhat serious events at 12 months, excluding stent fracture	2 (1.8%)	11 (9.8%)	14 (12.5%)
Serious event at 12 months	0 (0.0%)	0 (0.0%)	0 (0.0%)
Patients with adverse event at 24 Months	13 (11.6%)	47 (42.0%)	73 (65.2%)

¹Includes events that are due to or possible due to stent, and stent fractures.

²Includes events that are due to or possible due to stent, implantation, or catheterization, and stent fractures

Table 5 through 7 document the stent related, implantation related and catheterization procedure related adverse events in the COAST.

Table 5: Stent Related Adverse Events for COAST

	Event	n (Event Rate)
COAST (n=112)	Aortic Aneurysm	1 (0.9 %)
	Increased cardiac output, tachycardia/light-headedness	1 (0.9 %)

Table 6: Implantation Related Adverse Events for COAST

	Event	n (Event Rate)
COAST (n=112)	Aortic Aneurysm	2 (1.8 %)
	Back Pain	1 (0.9 %)
	Chest Pain	5 (4.5 %)
	Confined Vascular Tear	3 (2.7 %)
	Groin Pain	4 (3.6 %)
	Intimal tear with vessel irregularity	1 (0.9 %)
	Jailed left subclavian	1 (0.9 %)
	Local hematoma groin	1 (0.9 %)
	Pain	1 (0.9 %)
	Right leg pain	1 (0.9 %)
	Stent malposition	1 (0.9 %)

Table 7: Catheterization Related Adverse Events for COAST

	Event	n (Event Rate)
COAST (n=112)	AV fistula	1 (0.9 %)
	Bleeding	1 (0.9 %)
	Corneal abrasion	1 (0.9 %)
	Decreased pulse	1 (0.9 %)
	Fever	1 (0.9 %)
	Groin pain	1 (0.9 %)
	Local hematoma groin	4 (3.6 %)
	Right inguinal hematoma	1 (0.9 %)

In the COAST trial five patients experienced aortic wall injury, four that occurred prior to hospital discharge and one that occurred by 24 month follow-up. These five events are detailed in Table 8, below.

Table 8: COAST Aortic Wall Injuries by 24 Month Follow-up

Patient	Bare Metal CP Stent Implanted	Type of Injury	Outcome of Event
1	No	Small aneurysm after dilation	Cross-over to Covered CP Stent
2	Yes	Therapeutic and localized tear	Resolved after implantation of Bare Metal CP Stent; tear no longer visible and not noted on further imaging
3	Yes	Contained rupture	Possible minimal aneurysm with no progression during admission; not noted on further imaging
4	Yes	Aneurysm	Implantation of Covered CP Stent
5	No	Acute, rapidly expanding aneurysm after dilation	Cross-over to Covered CP Stent

In COAST, four patients with events underwent catheter reinterventions, representing 3.7% of patients with an event. No surgical interventions were performed. Table 9 provides a summary of these four interventions.

Table 9: COAST Coarctation-Related Reintervention by 24 Months

Approximate Time to Intervention Post-procedure (Months)	Indication for Reintervention	Procedure Performed
11	Planned reintervention to fully expand stent	Redilation of stent, implantation of non-study stent
20	Planned re-expansion of stent	Redilation of stent
14	Aneurysm detected at 12m visit	Therapy for new aortic wall injury - implantation of Covered CP Stent
14	Planned re-expansion of stent, signs of restenosis at coarctation site	Redilation of stent

In COAST, two patients experienced non-coarctation related reinterventions that were documented by the 24 month follow-up. These patients underwent surgical interventions. The time to intervention for one patient was 20 months, at which time the patient had an aortic valve replacement to address progressive and severe left ventricular enlargement and progression of exercise intolerance. The time to intervention for the second patient was approximately two months, when the patient received a mitral valve replacement to address mitral regurgitation.

Stent fracture was increasingly common during follow-up ranging from 12% at 24 months to 36% at 60 months in patients treated with bare metal stents (COAST trial). However, no loss of structural integrity and no complications resulting in patient injury were observed. Table 10 shows the stent fractures in the COAST pivotal cohort.

Table 10: COAST Pivotal Cohort – Safety Data – Adverse Events – Stent Fracture

	Completed 12 Month Fluoroscopy (n=91) ¹	Completed 24 Month Fluoroscopy (n=87) ³
Percentage of Eligible Subjects Undergoing Fluoroscopy	91/104 ² (88%)	87/103 ⁴ (84%)
Stent Fracture	2 (2.2%)	11 (12.6%)
No loss of structural integrity	2	11
Loss of structural integrity	0	0

¹ Among 104 eligible subjects, excludes: 1 patient lost to follow-up at 1 month, 2 patients lost to follow-up at 6 months, and 3 patients lost to follow-up at 12 months. An additional 7 patients did not undergo fluoroscopy at 12 months.

² Cross-over patients treated with Covered CP Stent (5) were followed only through implantation of Bare Metal CP Stent; intent to treat patients (2) were followed only through hospital discharge. One patient withdrew consent prior to the 12 month visit.

³ Among 103 eligible subjects, excludes: 6 patients previously lost to follow-up, and 5 patients lost to follow-up at 24 months. An additional 5 patients did not undergo fluoroscopy at 24 months.

⁴ In addition to cross-over patients treated with Covered CP Stent, intent to treat patients, and 1 patient who withdrew consent prior to the 12 month follow-up, 1 patient was noted to have an aneurysm at the 12 month visit and was treated using a Covered CP Stent; this patient is no longer followed for COAST and is currently enrolled in COAST II.

The analysis of effectiveness was based on the 105 COAST patients receiving bare metal CP stents. The key effectiveness outcomes are presented in Tables 11 through 12. Although the COAST primary effectiveness for blood pressure gradient reduction was not met, the observed blood pressure gradients were clinically meaningful. Failure to meet the endpoint was a function of the endpoint chosen rather than a failure to achieve reduction of the underlying gradient. From a clinical perspective, relief of systolic blood pressure gradient was complete and sustained.

Table 11. Summary of Late Outcomes and Major Pre-Specified Effectiveness Study Endpoints

COAST	Effectiveness Endpoint	Event Rate	P Value (CI)
Primary	Mean Reduction in Systolic Blood Pressure Difference, Pre-Dilation to the 12 Month Post-Dilation Follow-Up	30 ± 22mmHg	0.64 (26mmHg, 34mmHg)*
	Length of Stay in Hospital	1.1 ± 0.3 days	<0.001 (1.0 days, 1.1 days) [†]

*90% Confidence interval

[†] 95% Confidence Interval

confidence interval provided to illustrate the variability only and should not be used to draw any statistical conclusion.

Table 12. COAST Pivotal Cohort - Systolic Blood Pressure

	Number (Percent) or Median (Range) And Mean ± Standard Deviation		
	Completed 1 Month Follow-up (n=100) ¹	Completed 12 Month Follow-up (n=92) ²	Completed 24 Month Follow-up (n=87) ³
Upper Extremity Systolic Blood Pressure (mmHg)			
Median (range)	118 (83 to 148)	122 (82 to 148)	121 (87 to 175)
Mean ± standard deviation	120 ± 12	123 ± 12	122 ± 14
Lower Extremity Systolic Blood Pressure (mmHg)			
Median (range)	119 (91 to 163)	122 (89 to 180)	123 (84 to 172)
Mean ± standard deviation	122 ± 15	123 ± 15	125 ± 16
Systolic Blood Pressure Difference (mmHg)			
Median (range)	-1 (-45 to 32)	-1 (-37 to 40)	-4 (-46 to 43)
Mean ± standard deviation	-2 ± 13	-1 ± 15	-3 ± 15
Reduction in Systolic Blood Pressure Difference Pre-Dilation (n=99 ⁴ , 91 ⁵)	31 ± 18	30 ± 22 ⁶	33 ± 20

¹ Among 104 eligible subjects, excludes: 1 patient lost to follow-up at 1 month. An additional 3 patients missed the 1 month visit.

² Among 104 eligible subjects, excludes: 1 patient previously lost to follow-up, 2 patients lost to follow-up at 6 months, and 3 patients lost to follow-up at 12 months. An additional 6 patients missed the 12 month visit.

³ Among 101 eligible subjects, excludes: 6 patients previously lost to follow-up, and 5 patients lost to follow-up at 24 months. An additional 3 patients missed the 24 month visit.

⁴ One patient does not have 1 month systolic blood pressure difference due to missing lower extremity pressure.

⁵ One patient does not have 12 month systolic blood pressure difference due to missing lower extremity pressure.

⁶ Primary effectiveness outcome.

Procedural Data

A summary of procedural data of those enrolled patients who underwent cardiac catheterization for the purpose of Coarctation of the Aorta is provided in Table 13.


Table 13: COAST Pivotal Cohort – Procedural Data

	Number (Percent)
	Treated with Bare Metal CP Stent, or Meeting Study Eligibility Criteria but Not Treated with Bare Metal CP Stent (n=107) ¹
Bare Metal CP Stent Implanted	105 (98%)
Second Bare Metal CP Stent Implanted (n=105)	2 (2%)

¹ Includes: Patients treated with Bare Metal CP Stent meeting study eligibility criteria (102), patients treated with Bare Metal CP Stent not meeting study eligibility criteria (3), patients meeting study eligibility but not treated with Bare Metal CP Stent (2).


COAST/COAST II Follow-up Data

A.1. Percent of Patients with New Stent Fracture




% Stent Fracture (new at each interval)		
Interval	COAST I	COAST II
12 months	4/121 (3.3%)	1/127 (0.8%)
24 months	12/121 (9.9%)	5/127 (3.9%)
36 months	2/121 (1.6%)	1/127 (0.8%)
48 months	10/121 (8.2%)	4/127 (3.1%)
60 months	9/121 (7.4%)	2/127 (1.6%)

A.2. Percent of Patients with Stent Fracture (Cumulative)



% Stent Fracture (cumulative)		
Interval	COAST I	COAST II
12 months	4/121 (3%)	1/127 (0.8%)
24 months	16/121 (13%)	6/127 (4.7%)
36 months	18/121 (15%)	7/127 (5.5%)
48 months	28/121 (23%)	11/127 (8.6%)
60 months	37/121 (31%)	13/127 (10%)


B.1. Percent of Patients with Aortic Wall Injury (New or Progressive)*



% patients with AWI		
Interval	COAST I	COAST II
12 months	1/121 (0.8%)	1/127 (0.8%)
24 months	0/121	0/127
36 months	3/121 (2.5%)	0/127
48 months	1/121 (0.8%)	0/127
60 months	1/121 (0.8%)	2/127 (1.6%)

*All recorded injuries were “new”, none designated as progressive


B.2. Cumulative Percent of Patients with Aortic Wall Injury (New or Progressive)*



% patients with AWI (cumulative)		
Interval	COAST I	COAST II
12 months	1/121 (0.8%)	1/127 (0.8%)
24 months	1/121 (0.8%)	1/127 (0.8%)
36 months	4/121 (3.3%)	1/127 (0.8%)
48 months	5/121 (4.1%)	1/127 (0.8%)
60 months	6/121 (5%)	3/127 (2.4%)

*All recorded injuries were “new”, none designated as progressive


C.1. Percent of Patients with Coarctation Related Re-intervention or Surgery*



% patients with reintervention or surgery		
Interval	COAST I	COAST II
12 months	4/121 (3.3%)	9/128 (7%)
24 months	3/121 (2.5%)	3/128 (2.3%)
36 months	6/121 (5%)	3/128 (2.3%)
48 months	6/121 (5%)	3/128 (2.3%)
60 months	2/121 (1.7%)	3/128 (2.3%)


*All patients listed had catheter reinterventions, none had surgery

C.2. Percent of Patients with Coarctation Related Re-intervention or Surgery (Cumulative)




% patients with catheter-based re-intervention or surgery (cumulative)		
Interval	Bare Metal	Covered Stent
12 months	3/121 (2.5%)	9/128 (7%)
24 months	6/121 (5%)	12/128 (9.3%)
36 months	12/121 (9.9%)	15/128 (11.8%)
48 months	18/121 (14.9%)	18/128 (14%)
60 months	20/121 (16.5%)	21/128 (16.4%)

C.3. Percent of Bare Metal Patients with Catheter-Based or Surgical Coarctation-Related Re-intervention




% Bare Metal patients with catheter-based re-intervention or surgery		
Interval	Catheterization	Surgery
12 months	3/121 (2.5%)	0 (0%)
24 months	3/121 (2.5%)	0 (0%)
36 months	6/121 (5%)	0 (0%)
48 months	6/121 (5%)	0 (0%)
60 months	2/121 (1.7%)	0 (0%)

C.4. Percent of Bare Metal Patients with Catheter-Based or Surgical Coarctation-Related Re-intervention (Cumulative)




% Bare Metal patients with catheter-based re-intervention or surgery (cumulative)		
Interval	Catheterization	Surgery
12 months	3/121 (2.5%)	0 (0%)
24 months	6/121 (5%)	0 (0%)
36 months	12/121 (9.9%)	0 (0%)
48 months	18/121 (14.9%)	0 (0%)
60 months	20/121 (16.5%)	0 (0%)

C.5. Percent of Covered Stent Patients with Catheter-Based or Surgical Coarctation-Related Re-intervention




% Covered Stent patients with re-intervention or surgery		
Interval	Catheterization	Surgery
12 months	9/128 (7%)	0 (0%)
24 months	3/128 (2.3%)	0 (0%)
36 months	3/128 (2.3%)	0 (0%)
48 months	3/128 (2.3%)	0 (0%)
60 months	3/128 (2.3%)	0 (0%)

C.6. Percent of Covered Stent Patients with Catheter-Based or Surgical Coarctation-Related Re-intervention (Cumulative)



% Covered Stent patients with re-intervention or surgery (cumulative)		
Interval	Catheterization	Surgery
12 months	9/128 (7%)	0 (0%)
24 months	12/128 (9.3%)	0 (0%)
36 months	15/128 (11.8%)	0 (0%)
48 months	18/128 (14%)	0 (0%)
60 months	21/128 (16.4%)	0 (0%)


D.1 Percent of Patients with Upper to Lower Extremity Pressure Gradients > 15mmHg



% patients with Residual UE to LE pressure gradient > 15mmHg		
Interval	COAST I	COAST II
12 months	16/104 (15.4%)	16/107 (15%)
24 months	8/95 (8.4%)	13/81 (16%)
36 months	6/81 (7%)	12/72 (14%)
48 months	6/67 (9%)	6/53 (10%)
60 months	7/65 (11%)	14/47 (30%)

****Note, n is different at each time interval and based on the number of patients with pressures measured at each interval****

E.1. Percent of Patients with Proximal SBP >140



% patients with UESBP > 140mmHg		
Interval	COAST I	COAST II
12 months	18/104 (17%)	26/104 (25%)
24 months	16/95 (17%)	20/81 (25%)
36 months	6/86 (7%)	12/88 (14%)
48 months	8/74 (11%)	10/64 (16%)
60 months	11/74 (15%)	12/52 (23%)

Note, n is different at each time interval and based on the number of patients with pressures measured at each interval

F.1 Baseline Characteristics

Baseline Characteristics – COAST/COAST II		
Median (Range)		
	Bare Metal Patients (n=128)	Bare Metal Re-intervention Patients (n= 20)
Age (yrs)	16 (8 to 52)	13 (9 to 20)
Weight (kg)	65 (35 to 157)	51.7 (35 to 104)
	Covered Stent Patients (n=127)	Covered Stent Re-intervention Patients (n=21)
Age (yrs)	19 (6 to 67)	15 (6 to 35)
Weight (kg)	68 (17 to 122)	50 (17 to 95)

G.1 Systolic Blood Pressure Differences after Initial CoA Stenting

Systolic Blood Pressure Differences after Initial CoA Stenting		
Median (Range)		
	Bare Metal <u>Non</u> -Re-intervention Patients (n=108)	Bare Metal Re-intervention Patients (n=20)
SBP difference (AAo-Dao) (mmHg)	0 (-12 to 15)	1.5 (-5 to 12)
	Covered Stent <u>Non</u> -Re-intervention Patients (n=106)	Covered Stent Re-intervention Patients (n=21)
SBP difference (AAo-Dao) (mmHg)	2 (-12 to 38)	7 (-3 to 20)

H.1 Stent Diameter Measurements after Initial CoA Stenting

Stent Measurements after Initial CoA Stenting Median (Range)		
	Bare Metal No Re-intervention or only AWI-related Re- intervention Patients (n=111)	Bare Stent Non-AWI Related Re-intervention Patients (n=17)
Minimum Stent Diameter (mm)	14.4 (7.8 to 20.8)	11.7 (8.6 to 16)
Maximum Stent Diameter (mm)	16.5 (9.2 to 22.5)	13.5 (9.2 to 19.4)
	Covered Stent No Re- intervention or only AWI- related Re-intervention Patients (n=110)	Covered Stent Non-AWI Related Re-intervention Patients (n=17)
Minimum Stent Diameter (mm)	14.8 (9 to 23)	11.5 (7.8 to 16.5)
Maximum Stent Diameter (mm)	17 (11 to 25.7)	13.2 (9.2 to 18)

I.1 Upper Extremity Systolic Blood Pressures

	Number (Percent) or Median (Range) and Mean $\pm$ Standard Deviation	
	Upper Extremity Systolic Blood Pressures at 4 years	
Bare Metal Patients	Upper and Lower Extremity SBP measured	Only Upper Extremity SBP measured
Upper Extremity Systolic Blood Pressure (mmHg) (n=67,7) Median (range) Mean $\pm$ standard deviation	123 (79 to 231) 125 $\pm$ 19	120 (70 to 134) 113 $\pm$ 21
Covered Stent Patients		
Upper Extremity Systolic Blood Pressure (mmHg) (n=59,5) Median (range) Mean $\pm$ standard deviation	126 (93 to 174) 126 $\pm$ 15	124 (110 to 140) 123 $\pm$ 11

	Upper Extremity Systolic Blood Pressures at 5 years	
Bare Metal Patients	Upper and Lower Extremity SBP measured	Only Upper Extremity SBP measured
Upper Extremity Systolic Blood Pressure (mmHg) (n=65,9) Median (range) Mean $\pm$ standard deviation	128 (106 to 165) 129 $\pm$ 12	117 (100 to 160) 121 $\pm$ 19
Covered Stent Patients		
Upper Extremity Systolic Blood Pressure (mmHg) (n=47,5) Median (range) Mean $\pm$ standard deviation	131 (101 to 159) 129 $\pm$ 14	122 (121 to 134) 126 $\pm$ 6

RETURN OF EXPLANTED DEVICE:

NuMED, Inc. is interested in obtaining recovered Stents. Place the explanted device in a container or vial immediately after excision. For further instructions on the return of an explanted device, contact the RA Manager, NuMED, Inc. 2880 Main Street, Hopkinton, New York, 12965. Phone number: 315-328-4491.

WARNING:

NuMED stents are placed in the extremely hostile environment of the human body. Stents may fail to function for a variety of causes including, but not limited to, medical complications or failure of stent by fracture and embolization. In addition, despite the exercise of all due care in design, component selection, manufacture, and testing prior to sale, stents may be easily damaged before, during, or after insertion by improper handling, crimping or other intervening acts. Metal stents placed where there are extrinsic forces of compression, i.e. right ventricular outflow tract, are especially prone to fatigue fracture and embolization and should be avoided.

WARRANTY AND LIMITATIONS

Stents and accessories are sold in an 'as is' condition. The entire risk as to the quality and performance of the stent is with the buyer. NuMED disclaims all warranties, expressed or implied, with respect to catheters and accessories, including but not limited to, any implied warranty of merchantability or fitness for a particular purpose. NuMED shall not be liable to any person for any medical expenses or any direct or consequential damages resulting from the use of any catheter or accessory or caused by any defect, failure, or malfunction of any catheter or accessory, whether a claim for such damages is based upon warranty, contract, tort, or otherwise. No person has any authority to bind NuMED to any representation or warranty with respect to catheters and accessories.

G-Armor Stent® Foreshortening Chart

Inflated Balloon Diameter	G8Z46 (Stent length after expansion) Percentage Shortening	G8Z52 (Stent length after expansion) Percentage Shortening	G8Z57 (Stent length after expansion) Percentage Shortening	G8Z63 (Stent length after expansion) Percentage Shortening
12mm	(4.43) cm 3.61%	(4.98) cm 3.43%	(5.49) cm 3.62%	(6.02) cm 4.08%
14mm	(4.37) cm 4.92%	(4.91) cm 4.83%	(5.39) cm 5.32%	(5.92) cm 5.66%
15mm	(4.32) cm 6.05%	(4.85) cm 6.00%	(5.33) cm 6.39%	(5.85) cm 6.71%
16mm	(4.27) cm 7.07%	(4.80) cm 6.87%	(5.26) cm 7.54%	(5.77) cm 7.96%
18mm	(4.16) cm 9.41%	(4.66) cm 9.66%	(5.13) cm 9.95%	(5.64) cm 10.15%
20mm	(4.03) cm 12.30%	(4.54) cm 12.00%	(4.95) cm 13.01%	(5.44) cm 13.36%
22mm	(3.91) cm 14.99%	(4.37) cm 15.25%	(4.77) cm 16.27%	(5.23) cm 16.58%
24mm	(3.75) cm 18.37%	(4.15) cm 19.60%	(4.53) cm 20.38%	(4.98) cm 20.54%

Inflated Balloon Diameter	G10Z46 (Stent length after expansion) Percentage Shortening	G10Z52 (Stent length after expansion) Percentage Shortening	G10Z57 (Stent length after expansion) Percentage Shortening	G10Z63 (Stent length after expansion) Percentage Shortening
26mm	(3.97) cm 14.14%	(4.43) cm 14.70%	(4.89) cm 14.79%	(5.33) cm 15.23%
28mm	(3.85) cm 16.76%	(4.31) cm 17.04%	(4.72) cm 17.77%	(5.14) cm 18.25%
30mm	(3.74) cm 19.15%	(4.16) cm 19.90%	(4.56) cm 20.51%	(4.94) cm 21.44%

G-Armor Stent® 8 Zig Balloon Sizing Chart

Inner Balloon Pressure (atm)	Stent ID (mm)							
	12mm Diameter RBP = 7.0	14mm Diameter RBP = 6.0	15mm Diameter RBP = 5.0	16mm Diameter RBP = 5.0	18mm Diameter RBP = 4.0	20mm Diameter RBP = 4.0	22mm Diameter RBP = 3.0	24mm Diameter RBP = 3.0
1.0	4.50	5.73	5.59	6.47	7.37	7.97	8.72	9.57
2.0	4.84	5.99	5.78	6.85	7.58	8.31	9.22	10.15
3.0	5.03	6.18	6.06	7.17	7.91	8.77	9.83	10.67
4.0	5.30	6.36	6.31	7.50	8.33	9.22	10.37	11.17
4.5							10.79	11.71
5.0	5.54	6.56	6.59	7.94	8.77	9.81		
Outer Balloon Pressure (atm)								
1.0	10.09	12.13	13.54	14.23	16.48	18.24	19.87	21.66
2.0	10.64	12.61	14.03	14.77	17.07	19.09	20.77	22.78
3.0	11.08	13.13	14.70	15.34	17.76	19.89	22.22	24.05
4.0	11.64	13.60	15.02	15.87	18.34	20.59		
5.0	12.09	14.05	15.53	16.46				
6.0	12.34	14.37						
7.0	12.62							

*This data is based on testing performed using the NuMED BIB® Stent Placement Catheter.

The figures in bold face represent the stent ID @ Rated Burst Pressure.

G-Armor Stent® 10 Zig Balloon Sizing Chart

Inner Balloon Pressure (atm)	Stent ID (mm)		
	26mm Diameter RBP = 3.0	28mm Diameter RBP = 2.0	30mm Diameter RBP = 2.0
1.0	11.35	12.29	12.79
2.0	11.91	12.99	13.37
3.0	12.40	13.50	14.07
4.0	12.98	14.27	14.61
Outer Balloon Pressure (atm)			
1.0	23.78	26.57	28.62
2.0	25.08	28.27	30.12
3.0	26.33		

*This data is based on testing performed using the NuMED BIB® Stent Placement Catheter.

The figures in bold face represent the stent ID @ Rated Burst Pressure.

FOR ALL NuMED CATHETERS AN INFLATION DEVICE WITH PRESSURE GAUGE SHOULD BE USED.


G-Armor Stent® 8 Zig

BIB DELIVERY CATHETER BALLOON DIAMETER AND INTRODUCER SIZE	REQUIRED INTRODUCER WITH BARE G-ARMOR STENT	REQUIRED INTRODUCER WITH COVERED G-ARMOR STENT
12MM (8F)	13F	14F
14MM (8F)	13F	14F
15MM (9F)	13F	14F
16MM (9F)	14F	14F
18MM (10F)	14F	14F
20MM (10F)	14F	16F
22MM (11F)	14F	16F
24MM (11F)	14F	16F

G-Armor Stent® 10 Zig

BIB DELIVERY CATHETER BALLOON DIAMETER AND INTRODUCER SIZE	REQUIRED INTRODUCER WITH BARE G-ARMOR STENT	REQUIRED INTRODUCER WITH COVERED G-ARMOR STENT
26MM (16F)	16F	16F
28MM (16F)	16F	18F
30MM (16F)	16F	18F

Description of Graphical Symbols

	Use By		Consult Instructions For Use	REF	Catalog Number		Manufacturer
	MR Conditional		Model Number		Batch Code		Sterilized Using Ethylene Oxide
	Keep Away From Sunlight		Temperature Limitation		Do Not Reuse		